

# **ENTECH ENGINEERING INC.**

**P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118**

January 13, 2010


Mr. Sean O'Brien  
Environmental Department  
ConocoPhillips Company  
Sweeny Refinery  
Old Ocean, Texas 77463  
P.O. Box 866  
Sweeny, Texas 77480

**SUBJECT: TRANSMITTAL OF ENTECH ENGINEERING REPORT NO. ER2009-12-409 ENTITLED "CONOCOPHILLIPS COMPANY. SWEENY REFINERY UNIT 39.1 TAIL GAS INCINERATOR (EPN 39.1-95-118), INITIAL DEMONSTRATION OF COMPLIANCE TEST (TCEQ REGULATED ENTITY NUMBER: RN101619179; CUSTOMER REFERENCE NUMBER: CN601674351; ACCOUNT ID NO. BL-0042-G; PERMIT NO. 5920A AND PSD-TX-103M3)"**

Entech Engineering conducted an initial demonstration of compliance test on the Unit 39.1 Tail Gas Incinerator at the ConocoPhillips Company, Sweeny Refinery in Old Ocean, Texas. The test program was conducted on November 17 and 18, 2009 under ConocoPhillips Company Service Order No. SO4511980040.

Two copies and a CD of the Entech final report, which documents the findings and results of this program, are enclosed. Please note that the results presented in this report are only related to the items tested or the samples as received by Entech's lab; further, this report will not be reproduced, without the written approval of the client. Please contact us at our League City, Texas office if you have any questions or comments concerning the findings of this program.

Sincerely,



Joseph Muir  
Senior Environmental Scientist

Reviewed by:



Jared Vawter  
Environmental Scientist II

Approved by:



Edward J. Pasternak  
Technical Manager

# ***ENTECH ENGINEERING INC.***

---

**P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118**

**CONOCOPHILLIPS COMPANY**

**SWEENY REFINERY**

**UNIT 39.1 TAIL GAS INCINERATOR (EPN 39.1-95-118)**

**INITIAL DEMONSTRATION OF COMPLIANCE TEST**

**(TCEQ REGULATED ENTITY NUMBER: RN101619179; CUSTOMER REFERENCE  
NUMBER: CN601674351; ACCOUNT ID. BL-0042-G; PERMIT 5920A AND PSD-TX-103M3)**

**(ENTECH REPORT NO. ER2009-12-409)**

**PREPARED BY**

**ENTECH ENGINEERING INC.**

**LEAGUE CITY, TEXAS**

**NOVEMBER 2009**

**PREPARED FOR**

**CONOCOPHILLIPS COMPANY**

**OLD OCEAN, BRAZORIA COUNTY, TEXAS**

**SAMPLING LOCATION**

**CONOCOPHILLIPS COMPANY**

**SWEENY REFINERY, UNIT 39.1 TAIL GAS INCINERATOR**

**OLD OCEAN, BRAZORIA COUNTY, TEXAS**

# ***ENTECH ENGINEERING INC.***

---

***P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118***

## **TABLE OF CONTENTS**

<u>Section</u>	<u>Page</u>
1.0 Summary	1
2.0 Test Method and Equipment Description	3
3.0 Results and Discussion	9

## **APPENDICES**

### Appendix

- A. Field “Raw” Data
- B. Laboratory Analyses
- C. Example Calculations and QA/QC Data
- D. Instrument Specifications
- E. Equipment Calibrations
- F. Calibration Gas Certifications
- G. Process Data
- H. Resumes
- I. Chain of Custody
- J. Personnel Information

# ***ENTECH ENGINEERING INC.***

---

***P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118***

## **SECTION 1.0 SUMMARY**

Entech Engineering Inc. was retained by ConocoPhillips Company to conduct an initial demonstration of compliance test on the Unit 39.1 Tail Gas Incinerator (TGI) at the Sweeny Refinery in Old Ocean, Brazoria County, Texas. The objective of this program was to demonstrate initial compliance of Unit 39.1 TGI Unit according to the Texas Commission on Environmental Quality (TCEQ) Permit No. 5920A and PSD-TX-103M3. The emission compliance test program was performed on November 17 and 18, 2009, and was coordinated by Mr. Sean O'Brien of ConocoPhillips Company. TCEQ was given notification of testing but no investigator was present to observe the compliance testing.

The Sweeny Refinery's Unit 39.1 TGI, which is designated in the permit as Emission Point Number (EPN) 39.1-95-118, is designed to control reduced sulfur in the tail gas of the sulfur recovery unit (SRU). The TGI fires either fuel gas or sweet natural gas as its primary fuel. According to the permit, the SRU TGI has to conduct an initial stack sampling to establish the actual pattern and quantities of air contaminants being emitted into the atmosphere.[Special Condition 15]. Air contaminants emitted from TGI to be tested include volatile organic compound (VOC), nitrogen oxides (NO<sub>x</sub>), carbon monoxide (CO), sulfur dioxide (SO<sub>2</sub>), hydrogen sulfide (H<sub>2</sub>S), and particulate matter less than or equal to ten microns (PM<sub>10</sub>) emissions.

For the compliance test, Entech Engineering conducted three one-hour runs on particulate matter (PM) instead of PM<sub>10</sub>, and NO<sub>x</sub>, CO, VOC, SO<sub>2</sub> emissions, and three four-hour runs on H<sub>2</sub>S emissions, with the unit operating at or near the maximum achievable capacity. Velocity, moisture, and flow were measured at the TGI stack to determine mass emission rates. Additionally, ConocoPhillips Company personal collected and analyzed fuel samples during the compliance test period for fuel composition.

A summary of the emission compliance test results is presented in Table 1. The results showed that the incinerator met all requirements under the TCEQ permit. Test methods and equipment descriptions are presented in Section 2.0 and results and discussion are presented in Section 3.0.

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**Table 1**  
**ConocoPhillips Company**  
**Sweeny Refinery**  
**Old Ocean, Brazoria County, Texas**  
**Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)**  
**Initial Demonstration of Compliance Test Summary**  
**(TCEQ Regulated Entity Number: RN101619179; Customer Reference Number CN601674351;**  
**Account ID No. BL-0042-G; Permit 5920A and PSD-TX-103M3)**

Pollutant	Pollutant Conc.			Allowable Permit	Allowable Exceeded (Yes/No)
	ppmv, dry	grains/DSCF	lb/hr	lb/hr	
CO	64.74	N/A	5.41	8.95	No
SO <sub>2</sub> *	0.9	N/A	0.18	51.17	No
NO <sub>x</sub>	13.85	N/A	1.90	2.37	No
VOC	1.08	N/A	0.14	0.24	No
PM	N/A	0.007	1.16	0.29	Yes
H <sub>2</sub> S*	0.27	N/A	0.03	1.09	No

\* - SO<sub>2</sub> and H<sub>2</sub>S stack concentration were determined by the method detection limits.

## **SECTION 2.0 TEST METHOD AND EQUIPMENT DESCRIPTION**

The emission test equipment, procedures, and analytical methods used in this program were those described in the Environmental Protection Agency (EPA) Code of Federal Regulations (CFR), Title 40 (Protection of the Environment), Part 60 (New Source Performance Standard), Appendix A. Sampling equipment was prepared and calibrated at the Entech Engineering facility prior to transportation. At the site, sampling equipment was set up, checked out, and employed according to the following Reference Method procedures.

### **EPA Reference Method 1- “Sample and Velocity Traverses for Stationary Sources”.**

The TGI outlet is a circular stack with an internal diameter (ID) measuring 78 inches with two three-inch ports installed at 90° apart that conform to the requirements specified in the EPA Reference Method 1. Based on the sample port locations, Entech Engineering determined that a minimum of twelve traverse points was required to obtain representative emission samples. Each traverse point was sampled for five-minutes to obtain representative samples. Reference Method 1 procedures were also used to verify the absence of cyclonic flow; a schematic of the outlet sampling location is shown in Figure 1.

**EPA Reference Method 2- “Determination of Stack Gas Velocity and Volumetric Flow Rate (Type S Pitot Tube)”.** Volumetric flow rate was determined by measuring the velocity using a Type S (Staustscheibe or reverse type) pitot tube with an assigned baseline coefficient of 0.84 and differential pressure gauge (inclined manometer).

**EPA Reference Method 3A - “Determination of Oxygen and Carbon Dioxide Concentrations in Emissions From Stationary Sources (Instrumental Analyzer Procedure)”.** Excess oxygen (O<sub>2</sub>) was measured by instrumental methods.

**EPA Reference Method 3B - “Gas Analysis for the Determination of Dry Molecular Weight”.** Excess carbon dioxide (CO<sub>2</sub>) concentrations were measured per Reference Method 3. Tedlar bag sample was collected using an evacuation tank and analyzed for CO<sub>2</sub> concentrations using an Orsat Gas Analyzer. The results were used to determine the stack gas molecular weight.

**EPA Reference Method 4 - “Determination of Moisture Content in Stack Gases”.** Moisture content of the stack gas was measured by condensing flue gas moisture in Modified and Greenburg-Smith impingers immersed in an ice water bath and determining the moisture gain by gravimetric analysis. A schematic of the RM 4 sampling train is shown in Figure 3.

**EPA Reference Method 5 - “Determination of Particulate Emissions in Stationary Sources”.** The Test includes both the front-half catch and the back-half catch as in required by the TCEQ.

## **ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**EPA Reference Method 6C-** *“Determination of Sulfur Dioxide Emissions From Stationary Sources (Instrumental Analyzer Procedure)”*.

**EPA Reference Method 7E -** *“Determination of Nitrogen Oxides Emissions From Stationary Sources (Instrumental Analyzer Procedure)”*.

**EPA Reference Method 10 -** *“Determination of Carbon Monoxide Emissions From Stationary Sources”*.

**EPA Reference Method 15-** *“Determination of Hydrogen Sulfide, Carbonyl Sulfide, and Carbon Disulfide Emissions From Stationary Sources”*. Sulfide concentrations were determined using a Shimadzu gas chromatograph equipped with a flame photometric detector (GC/FPD). Sulfide calibrations were performed with a VICI Metronics Dynacalibrator using sulfide permeation tubes. Off-gas samples were conditioned with a particulate filter and SO<sub>2</sub> (citrate buffer) scrubber.

**EPA Reference Method 18-** *“Measurement of Gaseous Organic Compound Emissions by Gas Chromatography”*. Integrated VOC samples were collected in Tedlar bags and analyzed for methane and ethane using gas chromatograph equipped with flame ionization detector (GC/FID). Methane and ethane measured in the samples were excluded from the VOCs measured by the THC.

**Reference Method 25A-** *“Determination of Total Gaseous Organic Concentrations Using a Flame Ionization Analyzer”*. VOCs will be measured as propane equivalent.

In this test program, Continuous Emission Monitoring Systems (CEMS) were used to determine the gaseous concentrations of Nitrogen Oxides (NO<sub>x</sub>), Sulfur Dioxide (SO<sub>2</sub>), Carbon Monoxide (CO) and excess Oxygen (O<sub>2</sub>). An Entech test trailer equipped with a sampling systems conforming to the Reference Method specifications was used for the test. A description of the Entech Engineering CEM systems is presented as follows:

Thermo Environmental Instruments, Inc. NO<sub>x</sub> Analyzer-

Model No.	:	42C
Serial No.	:	42CHL-77922-387
Principal	:	Chemiluminescence
Calibration	:	High Span - 53.7 ppmv NO <sub>x</sub> in N <sub>2</sub> Mid Span - 24.6 ppmv NO <sub>x</sub> in N <sub>2</sub> Low Span - N <sub>2</sub> Zero Gas
Full Span	:	0 - 53.7 ppmv NO <sub>x</sub>

**ENTECH ENGINEERING INC.***P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

Thermo Environmental Instruments, Inc. SO<sub>2</sub> Analyzer -

Model No. : 43C High Level  
Serial No. : 410005567  
Principal : Non-dispersive Ultraviolet  
Calibration : High Span - 47.3 ppmv SO<sub>2</sub> in N<sub>2</sub>  
Mid Span - 22.2 ppmv SO<sub>2</sub> in N<sub>2</sub>  
Low Span - N<sub>2</sub> Zero Gas  
Full Span : 0 - 47.3.0 ppmv SO<sub>2</sub>

California Analytical Instruments, Inc. CO/CO<sub>2</sub>/O<sub>2</sub> Analyzer-

Model No. : 300  
Serial No. : 1M08015  
Principal : NDIR  
Calibration : High Span - 297.0 ppmv CO in N<sub>2</sub>  
Mid Span - 132.0 ppmv CO in N<sub>2</sub>  
Low Span - N<sub>2</sub> Zero Gas  
Full Span : 0 - 297 ppmv CO  
  
Calibration : High Span - 9.49 % O<sub>2</sub> in N<sub>2</sub>  
Mid Span - 4.94 % O<sub>2</sub> in N<sub>2</sub>  
Low Span - N<sub>2</sub> Zero Gas  
Full Span : 0 - 9.49 % O<sub>2</sub>

## J.U.M. Engineering, Inc. THC Analyzer (Stack) -

Model No. : VE 7  
Serial No. : Entech #4 (9560390)  
Principal : Flame ionization detector  
Calibration : High Span - 88.4 ppmv propane in N<sub>2</sub>  
Mid Span - 52.6ppmv propane in N<sub>2</sub>  
Low Span - 31.0 ppmv propane in N<sub>2</sub>  
Zero Span - N<sub>2</sub> Zero Gas  
Full Span : 0 - 100 ppmv THC

Sampling System (NO<sub>x</sub>/SO<sub>2</sub>/CO/O<sub>2</sub> CEMS) - A diaphragm pump was used to draw samples continuously from the sampling points through a heated probe and a heated Teflon® sample line. A condenser and a series of filters were used to condition the samples by removing moisture and particulate matter from the samples. Control valves and rotameters were used to regulate the conditioned gas flow to the instruments. All emission concentrations were measured on a dry-basis.

Instrument Calibration - Calibration gases for NO<sub>x</sub>, SO<sub>2</sub>, CO and O<sub>2</sub> were first sent directly to the instruments to check linearity of the instruments (Calibration Error Check). The calibration gases were then sent via a sample line to a three-way




valve located behind the sample probe and back to the instrument to verify system bias (System Bias Check). Procedures used in the system bias check were repeated after each test run to measure the instrument drift (Calibration Drift). The calibration gases for the NO<sub>x</sub>, SO<sub>2</sub>, CO and O<sub>2</sub> calibrations were EPA Protocol 1 gases.


Data Acquisition - A Yokogawa HR1300 hybrid recorder was used to record the CEMS output on a continuous basis. Final NO<sub>x</sub>, SO<sub>2</sub>, CO and O<sub>2</sub> readings were corrected for the system bias and instrument drift.

The Reference Methods 2, 3A, and 4 results were used to calculate the stack flow rates, which were then used to determine the pollutant mass emission rates in pounds-per-hour (lb/hr). Entech Engineering gaseous emission sampling configurations and schematic of the stack sampling location is presented in Figure 1. Reference Method 5 sampling equipment is presented in Figure 2.

Figure 1.  
ConocoPhillips Company  
Old Ocean, Texas  
Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)  
(TCEQ Account No. BL-0042-G; Permit No. 5920A)  
Gaseous Emission Sampling Configuration  
(Drawing Not to Scale)



**Figure 2**  
**ConocoPhillips Company**  
**Old Ocean, Texas**  
**Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)**  
**(TCEQ Account No. BL-0042-G; Permit No. 5920A)**  
**Reference Method 5 - Particulate Matter Sampling Train**  
**(Drawing Not to Scale)**



### SECTION 3.0 RESULTS AND DISCUSSIONS

Entech Engineering conducted an Initial Demonstration of compliance test on the Unit 39.1 Tail Gas Incinerator at the ConocoPhillips Company, Sweeny Refinery in Old Ocean, Brazoria County, Texas. The objective of the program was to determine initial demonstration of compliance of the incinerator according to TCEQ Permit 5920A and PSD-TX-103M3. The compliance test was conducted on November 17 and 18, 2009, with the unit operating at maximum achievable conditions.

During the compliance test, the NO<sub>x</sub> emissions averaged 13.85 parts-per-million-by-volume-dry-basis (ppmvd) and 1.90 pounds-per-hour (lb/hr). SO<sub>2</sub> emissions averaged 0.9 ppmvd and 0.18 lb/hr. The CO emissions averaged 64.74 part-per-million-by-volume-dry-basis (ppmvd) and 5.41 pounds-per-hour (lb/hr). The VOC emissions averaged 1.08 part-per-million-by-volume-dry-basis (ppmvd) and 0.14 pounds-per-hour (lb/hr). The PM emissions averaged 0.007 grains-per-dry-standard-cubic-feet (grains/DSCF) and 1.16 pounds-per-hour (lb/hr). The H<sub>2</sub>S emissions averaged 0.27 part-per-million-by-volume-dry-basis (ppmvd) and 0.03 pounds-per-hour (lb/hr).

Test results and selected test parameters for the incinerator are presented in Table 2. Stack VOC emission and particulate matter data summaries are presented in Table 3 and 4 respectively. All other pertinent data of the test program is contained in the appendices. The field raw data and laboratory analysis are contained in Appendices A and B. Example calculations and QA/QC data and instrument specifications are presented in Appendices C and D. Equipment calibrations, calibration gas certifications, process data, resumes, chain of custody, and personnel information are presented in Appendices E through J.

**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Table 2**  
**ConocoPhillips Company, Sweeny Refinery**  
**Old Ocean, Brazoria County, Texas**  
**Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)**  
**Initial Demonstration of Compliance Test**  
**(TCEQ Regulated Entity Number: RN101619179;**  
**Customer Reference Number: CN601674351; Account ID No. BL-0042-G;**  
**Permit No. 5920A and PSD-TX-103M3)**

Test ID		Test 1	Test 2	Test 4	Average
Sampling Date		11/17/09	11/18/09	11/18/09	-
Sampling Time		15:10 - 16:10	11:15 - 12:15	13:15 - 14:15	-
Sampling Duration	Minutes	60	60	60	60

**Operation Data**

Production Rate	Long Tons/day	NA	NA	NA	NA
-----------------	---------------	----	----	----	----

**Stack Data**

Oxygen (O <sub>2</sub> )	vol%, dry	2.71	3.00	1.85	2.52
Carbon Dioxide (CO <sub>2</sub> )	vol%, dry	4.30	4.43	4.47	4.40
Moisture	%	8.19	10.15	9.71	9.35
Stack Temperature	°F	547.5	548.7	546.3	547.5
Stack Gas Velocity	ft/sec	20.17	20.43	20.28	20.29
Volumetric Flow Rate	DSCFM	19326.95	19118.60	19116.66	19187.40

**Emission Data**

Nitrogen Oxides (NO <sub>x</sub> )	ppmv, dry	13.98	13.67	13.91	13.85
	lb/hr	1.94	1.87	1.90	1.90
Sulfur Dioxide (SO <sub>2</sub> )	ppmv, dry	0.9	0.9	0.9	0.9
	lb/hr	0.18	0.18	0.18	0.18
Carbon Monoxide (CO)	ppmv, dry	55.49	60.14	78.60	64.74
	lb/hr	4.68	5.01	6.55	5.41
Volatile Organic Compound (VOC)	ppmv, dry	1.37	1.37	0.49	1.08
	lb/hr	0.18	0.18	0.06	0.14
Hydrogen Sulfide (H <sub>2</sub> S)	ppmv, dry	0.27	0.27	0.27	0.27
	lb/hr	0.03	0.03	0.03	0.03
Particulate Matter	mg/DSCF	0.42	0.38	0.49	0.46
	grains/DSCF	0.006	0.006	0.008	0.007
	lb/hr	1.07	0.96	1.24	1.16

Note: Shaded areas represent minimum detectable limits.  
NA - Not Available

**Table 3. ConocoPhillips Company, Sweeny Refinery  
Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)  
Stack VOC Emission Data Summary (Based on THC)**

**Test Parameters**

Test ID	1	2	3	Average
Test Date	11/17/09	11/18/09	11/18/09	NA
Testing Period	15:10 - 16:10	11:15 - 12:15	13:15 - 14:15	NA

**Duct or Stack Conditions**

Oxygen (O2)	(vol%)	2.71	3.00	1.85	2.52
Carbon Dioxide, (CO2)	(vol%)	4.30	4.43	4.47	4.40
Flue Gas Temperature	(F)	547.50	548.70	546.30	547.50
Moisture Content	(vol%)	8.19	10.15	9.71	9.35
Velocity	(ft/sec)	20.17	20.43	20.28	20.29
Volumetric Flow Rate	(DSCF/min)	19326.95	19118.60	19116.66	19187.40

**Emission Data**

Methane (C1)	ppmv, dry	1.981	2.114	4.728	2.941
Methane as Propane	ppmv, dry	0.660	0.705	1.576	0.980
Ethane (C2)	ppmv, dry	0.220	0.220	0.220	0.220
Ethane as Propane	ppmv, dry	0.147	0.147	0.147	0.147
C1 & C2 as Propane	ppmv, dry	0.807	0.851	1.723	1.127
THC (including C1 & C2)	ppmv, wet	2.000	2.000	2.000	2.000
THC (including C1 & C2)	ppmv, dry	2.178	2.226	2.215	2.206
C1 & C2 as Propane	ppmv, dry	0.807	0.851	1.723	1.127
THC (excluding C1 & C2)*	ppmv, dry	1.371	1.375	0.492	1.079
THC (excluding C1 & C2)*	lb/hr	0.182	0.180	0.064	0.142

Note: Shaded areas represent minimum detectable limits.

\* Note: Total Hydrocarbons reported as propane equivalent, excluding methane and ethane.

**Table 4.**  
**ConocoPhillips Company**  
**Old Ocean, Texas**  
**Unit 39.1 Tail Gas Incinerator (EPN 39.1-95-118)**  
**Initial Demonstration of Compliance Test**  
**Reference Method 5 Particulate Results**

Sampling Location	Test ID	Front Half			Back Half		Corrected Dry Gas Meter Volume	Particulate Matter Concentration	Particulate Matter Concentration
		Filter	Front Half Wash	Total	Back Half Rinse + Impinger 1	Total PM			
		(mg)	(mg)	(mg)	(mg)	(mg)	dscf	mg/dscf	grain/dscf
Stack	1	4.9	2.1	7.0	10.1	17.1	40.895	0.42	0.0065
	2	5.7	4.1	9.8	5.0	14.8	38.603	0.38	0.0059
	3	4.1	8.8	12.9	7.2	20.1	40.982	0.49	0.0076

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**APPENDICES**

Appendix

- A. FIELD "RAW" DATA
- B. LABORATORY ANALYSIS
- C. EXAMPLE CALCULATIONS AND QA/QC DATA
- D. INSTRUMENT SPECIFICATIONS
- E. EQUIPMENT CALIBRATIONS
- F. CALIBRATION GAS CERTIFICATIONS
- G. PROCESS DATA
- H. RESUMES
- I. CHAIN OF CUSTODY
- J. PERSONNEL INFORMATION



***ENTECH ENGINEERING INC.***

*P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

**APPENDIX A.**

**FIELD “RAW” DATA**

Determination of Minimum Number of Traverse Points  
(40CFR60 Appendix A, Reference Method 1 or 1A)

By using the appropriate minimum number of traverse points figure in RM1 determine the number of traverse points recommended by both distance A and distance B, which have units of distance "stack duct diameters". This figure must be read vertically from either the top down, for distance A, or from bottom up, for distance B. The greater number of traverse points, as determined by each of the two distances will be the number of traverse points recommended for use. *Sketch a drawing of the stack below with both distances A and B and duct diameter shown. Show both elevation and plane views. Place an "X" at location of sampling console (if applicable) and include its height with respect to grade.*

How distances are determined: measured / schematic (circle one) other \_\_\_\_\_

Distance A 20 feet / inches (circle one) Duct diameters 3.08

Recommended number of traverse points as determined by Distance A 12

Distance B 64 feet / inches (circle one) Duct diameters 9.85

Recommended number of traverse points as determined by Distance B 12

Stack diameter at port 78 feet / inches (circle one)

Plane View	Elevation View

Date: 11/17/09

Signature (Personnel): \_\_\_\_\_

Signature (Team Leader): [Signature]

TO MAKE CORRECTION, USE SINGLE LINE, MARK THROUGH THE INCORRECT DATA, INITIAL, AND DATA.  
DO NOT USE PEN OR LIQUID PAPER TO COMPLETE COVER INCORRECT DATA.

[\\Enteng\\F\\Templates\\DataSheets\\Ver.2\\Traversepointworksheets.wpd]

November 2004

A1

**ENTECH ENGINEERING INC.**


P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

**Traverse Point Layout for Circular Stacks**Stack ID at Port 74 feet / inches (circle one)Port Extension 14 feet / inches (circle one)

Traverse Point No.	Column A Length Factor $K_L$	Column B Traverse Points $K_L \times \text{Stack ID}$	Column C Traverse Pt. + Reference $(K_L \times \text{Stack ID}) + \text{Port Ext.}$
1	0.044	3.432	21.432
2	0.146	11.388	29.388
3	0.296	23.088	41.088
4	0.704	54.912	72.912
5	0.854	66.612	84.612
6	0.956	74.568	92.568
7			
8			
9			
10			
11			
12			
13			
14			
15			
16			
17			
18			
19			
20			
21			
22			
23			
24			


Date: 11/17/09

AZ

Signature: 

$$\begin{array}{r} \cancel{17.2} \\ 19.2 \\ 20.1 \end{array}$$
[illegible]

Source Measurements	
Emission Point Height (ft):	200
Sampling Point Height (ft):	150
Stack Diameter (@ Exit):	578"
Stack Diameter (@ Port):	78"
Port Length (in.):	18"
Port Size (in.):	4"



Team Leader (Signature/Date) \_\_\_\_\_

Plant: CONOCO Phillips	Barometer ID: I3	Probe ID: G7.SC
Location: Sweeny	Barometric Pressure (in. Hg): 30.04	Thermocouple ID: G7.SC
Unit: 39.1 SRU	Static Pressure (in. H <sub>2</sub> O): 2.10	Thermometer / Thermocouple: ✓ / ✓
Source/ EPN: INCINERATOR	Ambient Temperature (°F): 70°F	Pitot Tube ID: G7.SC
Date: 11-15-09	Meter Pre Leak Check: .001 @ 15" VAC	PTCF: 0.84
Test ID: 2	Meter Post Leak Check: .001 @ 15" VAC	Pump ID: 1
Test Operator: G.S. VARGAS	Pitot (+) Tube Leak Check: 3" H <sub>2</sub> O @ 15 SEC	Meter ID: 1
Team Leader: JOE MUIR	Pitot (-) Tube Leak Check: 3" H <sub>2</sub> O @ 15 SEC	DGMCF: 1.0016
	Nozzle ID: #80 mm .470	ΔH Std: 1.89
	Filter ID: 1110	K factor: 22.9

[illegible]

Elapsed Time	Gas Volume	Average	Average	Average	Average	Average
(min)	(ft <sup>3</sup> )	$\Delta P$	Su Rt $\Delta P$	$\Delta H$	Meter Temperature	Stack Temp.

Emission Point Height (ft):	200'
Sampling Point Height (ft):	150'
Stack Diameter (@ Exit):	78"
Stack Diameter (@ Port):	78"
Port Length (in.):	15"
Port Size (in.):	4"

A hand-drawn sketch of a clock face and a vertical structure. The clock face is on the left, showing the numbers 12, 4, and 1. To its right is a vertical structure, possibly a tower or a pole, with a small circle near the top. Further right is a circular arrow indicating a clockwise direction, with the number 20' written next to it.

Completeness \_\_\_\_\_ Legibility        Reasonableness \_\_\_\_\_

Certified by:

\_\_\_\_\_  
 Personnel (Signature/Date)


Team Leader (Signature/Date)

## Sampling Field Data Sheet

Plant: Conoco Phillips	Barometer ID: I3	Probe ID: G 7.5C
Location: Sweeney	Barometric Pressure (in. Hg): 30.04	Thermocouple ID: G7.5C
Unit: 39.1 SRU	Static Pressure (in HO): .10	Thermometer / Thermocouple: <input checked="" type="checkbox"/> / <input checked="" type="checkbox"/>
Source/ EPN: 1N	Ambient Temperature (F): 76°F	Pitot Tube ID: G 7.5C
Date: 11.18.09	Meter Pre Leak Check: .003 @ 15" VAC	PTCF: 0.24
Test ID: 3 Method #5	Meter Post Leak Check: .001 @ 10" VAC	Pump ID: 1
Test Operator: Gustav Vargas	Pitot (+) Tube Leak Check: 3" H2O @ 15 sec	Meter ID: 1
Team Leader: JOE MUIR	Pitot (-) Tube Leak Check: 3" H2O @ 15 sec	DGMCF: 1.006
	Nozzle ID: 440 MB .470	ΔH Std: 1.89
	Filter ID: 1144	K factor: 20.4

[illegible]

Source Measurements	
Emission Point Height (ft):	200'
Sampling Point Height (ft):	150'
Stack Diameter (@ Exit):	78"
Stack Diameter (@ Port):	78"
Port Length (in.):	18"
Port Size (in.):	4"



**QA/QC Check**

Completeness 1      Legibility 1      Reasonableness 1

**Certified by:**

Personnel (Signature/Date)

Team Leader (Signature/Date) \_\_\_\_\_

**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Impinger Catch**

Client:	Conoco Phillips
Location:	Sweeney
Unit:	39.1
Sampling Location:	SRV Incinerator
Date:	11/17/09
Test No:	1
Balance ID:	TL-3

Impinger Number	Solution Used	Amount of Solution (ml)	Impinger Tip Configuration	Final Weight (g)	Initial Weight (g)	Weight Difference (g)	Final Visual Observation
1	H <sub>2</sub> O	100	M	217.1	158.6	58.5	clean
2	H <sub>2</sub> O	100	GS	170.5	156.5	14.0	clean
3	-	-	M				
4	Silica Gel		M	305.9	301.0	4.9	blue / pink
5							
6							
7							
8							

Total Weight Gain of Impingers (g) = 377.4 77.4

Date: 11/17/09

Signature: **Independent Verification  
of Calculations**

Explanation: Someone other than the person who performed the original work and associated calculations must verify that these calculations have been performed correctly.

Date: 1/12/10

Signature: 

**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Impinger Catch**

Client:	Grace Phillips
Location:	Swamy
Unit:	39.1
Sampling Location:	SNV Indicator
Date:	11/18/09
Test No:	2
Balance ID:	TL-3

Impinger Number	Solution Used	Amount of Solution (ml)	Impinger Tip Configuration	Final Weight (g)	Initial Weight (g)	Weight Difference (g)	Final Visual Observation
1	H <sub>2</sub> O	100	M	228.4	156.2	72.2	
2	H <sub>2</sub> O	100	GS	171.3	158.9	12.4	
3	—	—	M				
4	5% Gel	—	M	343.7	335.8	7.9	
5							
6							
7							
8							

Total Weight Gain of Impingers (g) = **92.5**

Date: 11-18-09

Signature: **Independent Verification  
of Calculations**

Explanation: Someone other than the person who performed the original work and associated calculations must verify that these calculations have been performed correctly.

Date: 1/12/09

Signature: 



# ENTECH ENGINEERING INC.

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## Impinger Catch


Client:	Green Philip
Location:	Survey
Unit:	39.1
Sampling Location:	Stack SRV Incinerator
Date:	11/14/09
Test No:	T-3
Balance ID:	76-3

Impinger Number	Solution Used	Amount of Solution (ml)	Impinger Tip Configuration	Final Weight (g)	Initial Weight (g)	Weight Difference (g)
1	H <sub>2</sub> O	100	M	229.1	157.2	71.9
2	H <sub>2</sub> O	100	bs	170.1	157.7	12.4
3	-	-	M			
4	SR bel	-	M	277.4	268.2	9.2
5						
6						
7						
8						

Final Visual Observation
Clear
Clear
Blue / pink

Total Weight Gain of Impingers (g) = **93.5**


Date: 11/18/09

Signature (Personnel): 

### Independent Verification of Calculations

Explanation: Someone other than the person who performed the original work and associated calculations must verify that these calculations have been performed correctly.

Date: 1/12/09

Signature (Team Leader): 

TO MAKE CORRECTION, USE SINGLE LINE, MARK THROUGH THE INCORRECT DATA, INITIAL, AND DATA. (November 2004)  
DO NOT USE PEN OR LIQUID PAPER TO COMPLETE COVER INCORRECT DATA. [\\Enteng\F\Templates\DataSheets\Ver.2\RM4\Refere-2.wpd]

# ENTECH ENGINEERING INC.

P. O. Box 890746, Houston, Texas 77289-0746, (281) 332-3118

## Orsat Analysis Results

Client:	longco
Location:	Old Ocean Tr
Unit:	39.1
Sampling Location:	Stack
Date:	11/17/09
Leak Check:	✓
Personnel:	JM

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub>	O <sub>2</sub> (%)
T1	1	4.3		
	2	4.3		
	3	4.3		
	Average	4.3		

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub> (%)	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub>	O <sub>2</sub> (%)
T2	1	4.4		
	2	4.5		
	3	4.4		
	Average	4.43		

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub> (%)	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub>	O <sub>2</sub> (%)
T3	1	4.8		
	2	4.4		
	3	4.5		
	Average	4.47		

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub> (%)	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub>	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub> (%)	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub>	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

Test No.	Run No.	CO <sub>2</sub> (%)	O <sub>2</sub> +CO <sub>2</sub> (%)	O <sub>2</sub> (%)
	1			
	2			
	3			
	Average			

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Instructions:** Use black ink pens only. Do not obliterate data. If a reading is not used or incorrectly taken use a single line to cross out the data and initial the crossout. All other data should not be marked in any way signifying completion of that sample or for any other reason. A calibration check must be performed and recorded in the calibration check section on a daily basis for both the initial and final weights. Duplicate weights must agree to within 0.5 mg (4 decimal places).

Beaker ID (EB09- )			Initial Weight								
DATE	TIME	INITIAL	NUMBER	2078	2079	2080	2081	2083	2084	2085	2086
1/12/10	0920	JMY	1	92.7036	94.1571	98.4559	101.1250	97.7797	100.1572	100.7440	100.1288
1/13/10	1000	JMY	2	92.7046	94.1370	98.4557	101.1246	97.7798	100.1573	100.7444	100.1285
			3								
			4								
			5								
			6								
			7								
			8								
			9								
			10								
Average (g)				92.7039	94.1371	98.4558	101.1248	97.7796	100.1573	100.7442	

Beaker ID (EB09- )			Final Weight									
DATE	TIME	INITIAL	NUMBER	2078	2079	2080	2081	2083	2084	2085	2086	
12/11/09	08:00	DM	1	92.7103	99.1421	98.4660	101.1273	97.7842	100.1612	100.3548	100.1343	
12/14/09	09:00	DM	2	92.7113	99.1385	98.4690	101.1325	97.7823	100.1622	100.7570	100.1360	
12/15/09	15:00	MA	3	92.7109	99.1343	98.4646	101.1282	97.7893	100.1611	100.7500	100.1492	
12/16/09	15:48	MA	4		99.1340	98.4668	101.1269	97.7847	100.1616	100.7524	100.1390	
12/18/09	13:00	DM	5			99.4695	101.1265			100.7475	100.1341	
12/23/09	09:00	DM	6			99.4660				100.7529	100.1391	
12/24/09	09:00	DM	7			98.4658				100.7531	100.1388	
			8									
			9									
			10									
Average (g)				92.7111	99.1392	98.4659	101.1267	97.7845	100.1614	100.7530		

Total Weight Gain				
Beaker ID (E009- )	2078	2079	2080	2081
Difference (g)	0.0072	0.0021	0.0101	0.0019
				0.0050
				0.0041
				0.0088
				0.0085

Plant Name	Concord Mills	Concord Mills	Concord Mills	Concord Mills	Concord Mills	Concord Mills	Concord Mills
Location	Old Dam Tr	Old Dam Tr	Old Dam Tr	Old Dam Tr	Old Dam Tr	Old Dam Tr	Old Dam Tr
Unit	Unit 341	Unit 341	Unit 341	Unit 341	Unit 341	Unit 341	Unit 341
Sample ID	D37/184W	T1 FHV	T1E/184W	T1E/184W	T2E/1/184W	T2 FHV	T3 FHV
Test Date	11/14/09	11/17/09	11/17/09	11/18/09	11/18/09	11/18/09	11/18/09
Personnel Initials	Jm	Jm	Jm	Jm	Jm	Jm	Jm

AID

# ENTECH ENGINEERING INC.

P. O. Box 880746 - Houston, Texas 77289-0746 - (281)332-3118

## 2009 RM 5 Particulate Filter Tare Weights

Instructions: Use black ink pens only. Do not obliterate data. If a reading is not used or incorrectly taken use a single line to cross out the data and initial the crossout. All other data should not be marked in any way signifying completion of that sample or for any other reason. A calibration check must be performed and recorded in the calibration check section on a daily basis for both the initial and final weights. Duplicate weights must agree to within 0.5 mg (4 decimal places).

Initial Weight

DATE	TIME	INITIAL	NUMBER	1146	1145	1144	1143	1118	1110	1127	1141
11-13-09	08:00	JM	1	0.8200	0.7271	0.7949	0.7904				
11-13-09	14:42	JM	2	0.8207	0.7241	0.7977	0.7925				
11-16-09	08:00	JM	3	0.8197	0.7179	0.7960	0.7911				
11-16-09	16:00	JM	4	0.8193	0.7181	0.7963	0.7925	0.6517	0.6989	0.7457	0.6085
11-17-09	05:50	JM	5				0.7941	0.6520	0.6991	0.7461	0.6095
			6								
			7								
			8								
			9								
			10								
Average (g)				0.7962				0.6990	0.7027	0.6097	

Final Weight

DATE	TIME	INITIAL	NUMBER	1144					1110	1127	1141
12-11-09	09:00	JM	1	0.7974					0.6974	0.7401	0.6097
12-14-09	08:30	JM	2	0.7961					0.6947	0.7460	0.6095
12-15-09	15:00	MAB	3	0.8002					0.7017	0.8007	
12-16-09	15:15	MAB	4	0.8004					0.7049	0.8009	
12-18-09	13:00	JM	5						0.7044		
			6								
			7								
			8								
			9								
			10								
Average (g)				0.8003					0.7047	0.8003	0.6095

Total Weight Gain

Filter ID (EF09- )		1144	1110	1127	1141
Difference (g)		0.0041	0.0057	0.0049	0.0000

Plant Name		Comcast	Comcast	Comcast	Comcast
Location		Old Com Tr	Old Com Tr	Old Com Tr	Old Com Tr
Unit		Unit 391	Unit 391	Unit 391	Unit 391
Sample ID		Test 1	Test 2	Test 3	Test 4
Test Date		11/17/09	11/18/09	11/19/09	11/19/09
Personnel Initials		JM	JM	JM	JM

A11

## RESULTS

Run averages corrected for bias

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

Run	O2 %	CO ppm	NOx ppm	THC ppm	SO2 ppm
1	2.7093	55.49	13.980	0.19	-0.575
2	3.0015	60.14	13.672	0.75	0.127
3	1.8526	78.60	13.906	1.95	0.095

CALERR01 Calibration Error Test at Run 1 . STRATA Version 2.0									
SO2	O2	CO	NOX	THC	SO2	O2	CO	NOX	THC
Volts	Volts	ppm	ppm	ppm	ppm	Volts	Volts	Volts	
11-17-2009	11:29:33		1.3911	5.37	0.242	3.26	-0.628	1.3911	
0.0537 0.0242	0.326	-0.0628							
11-17-2009	11:30:33		-0.1232	5.37	0.079	3.29	0.107	-0.1232	
0.0537 0.0079	0.329	0.0107							
11-17-2009	11:31:33		-0.0528	5.88	0.081	3.19	-0.196	-0.0528	
0.0588 0.0081	0.319	-0.0196							
11-17-2009	11:32:33		-0.1557	5.54	0.145	3.19	-1.050	-0.1557	
0.0554 0.0145	0.319	-0.1050							
11-17-2009	11:33:32		-0.2159	3.81	0.278	3.21	-1.101	-0.2159	
0.0381 0.0278	0.321	-0.1101							
11-17-2009	11:34:33		-0.2506	4.59	0.277	3.15	-1.179	-0.2506	
0.0459 0.0277	0.315	-0.1179							
11-17-2009	11:35:32		-0.2092	4.49	0.276	3.27	-1.107	-0.2092	
0.0449 0.0276	0.327	-0.1107							
11-17-2009	11:36:33		1.4606	3.96	0.277	3.26	-1.022	1.4606	
0.0396 0.0277	0.326	-0.1022							
11-17-2009	11:37:32		4.9111	3.92	0.277	3.13	-1.062	4.9111	
0.0392 0.0277	0.313	-0.1062							
11-17-2009	11:38:33		5.0856	3.61	0.277	3.10	-1.018	5.0856	
0.0361 0.0277	0.310	-0.1018							
11-17-2009	11:39:33		-0.1744	4.23	0.113	3.15	-0.293	-0.1744	
0.0423 0.0113	0.315	-0.0293							
11-17-2009	11:40:33		-0.0516	4.68	0.080	18.23	-0.129	-0.0516	
0.0468 0.0080	1.823	-0.0129							
11-17-2009	11:41:33		-0.0057	4.75	0.079	2.15	-0.093	-0.0057	
0.0475 0.0079	0.215	-0.0093							
11-17-2009	11:42:32		0.0218	4.41	0.081	2.09	-0.166	0.0218	
0.0441 0.0081	0.209	-0.0166							
11-17-2009	11:43:33		0.0241	-0.33	0.082	1.21	-0.048	0.0241	
-0.0033 0.0082	0.121	-0.0048							
11-17-2009	11:44:32		0.0319	-0.23	0.082	0.05	-0.096	0.0319	
-0.0023 0.0082	0.005	-0.0096							
11-17-2009	11:45:33		0.0149	-1.05	0.037	0.03	-0.656	0.0149	
-0.0105 0.0037	0.003	-0.0656							
11-17-2009	11:46:32		3.8306	-0.36	11.237	31.70	17.425	3.8306	
-0.0036 1.1237	3.170	1.7425							
11-17-2009	11:47:33		4.8185	-0.62	13.685	85.12	21.650	4.8185	
-0.0062 1.3685	8.512	2.1650							
11-17-2009	11:48:33		4.8548	0.28	20.074	85.33	22.542	4.8548	
0.0028 2.0074	8.533	2.2542							
11-17-2009	11:49:33		4.8957	-0.23	25.478	85.49	22.308	4.8957	
-0.0023 2.5478	8.549	2.2308							
11-17-2009	11:50:33		4.9443	-0.13	24.728	88.32	22.158	4.9443	
-0.0013 2.4728	8.832	2.2158							
11-17-2009	11:51:32		4.9249	-0.30	20.939	88.35	20.174	4.9249	
-0.0030 2.0939	8.835	2.0174							
11-17-2009	11:52:33		9.5747	0.77	0.907	84.70	5.893	9.5747	
0.0077 0.0907	8.470	0.5893							
11-17-2009	11:53:32		9.3885	-0.09	44.630	53.60	40.296	9.3885	
-0.0009 4.4630	5.360	4.0296							
11-17-2009	11:54:33		9.4394	-0.04	54.091	53.19	44.876	9.4394	
-0.0004 5.4091	5.319	4.4876							
11-17-2009	11:55:32		9.4687	-0.05	54.131	52.77	47.259	9.4687	
-0.0005 5.4131	5.277	4.7259							
11-17-2009	11:56:33		9.4806	0.05	54.131	52.76	48.189	9.4806	
0.0005 5.4131	5.276	4.8190							
11-17-2009	11:57:33		8.3187	29.52	42.451	52.74	19.408	8.3187	
0.2952 4.2451	5.274	1.9408							

CALERR01									
11-17-2009	11:58:33	-0.0390	128.18	0.842	39.40	0.034	-0.0390		
1.2818 0.0842	3.940 0.0034								
11-17-2009	11:59:33	-0.0448	130.67	0.476	30.90	-0.836	-0.0448		
1.3067 0.0476	3.090 -0.0836								
11-17-2009	12:00:32	1.2487	247.34	0.396	30.83	-0.799	1.2487		
2.4734 0.0396	3.083 -0.0799								
11-17-2009	12:01:33	-0.0692	303.53	0.422	30.54	-1.114	-0.0692		
3.0353 0.0422	3.054 -0.1114								
11-17-2009	12:02:32	-0.0271	297.02	0.371	10.31	-1.165	-0.0271		
2.9702 0.0371	1.031 -0.1165								

# Calibration Error Test at Run 1

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

		Reference Cylinder Numbers			
	Zero	Low-range	Mid-range	High-range	
O2	N2	cc9824	cc55212		
CO	N2	ALM014290	cc111686		
NOx	N2	cc161053	cc241846	cc61285	
THC	AIR	cc161053	cc236179	cc151954	
SO2	N2	cc204173	cc144163	cc57866	

Date/Time	11-17-2009		12:02:42		PASSED
Analyte O2	CO	NOx	THC	SO2	
Units %	ppm	ppm	ppm	ppm	
Zero Ref Cyl	0.0000	0.00	0.000	0.00	0.000
Zero Avg	0.0277	-0.33	0.082	0.07	-0.051
Zero Error%	0.3	0.1	0.2	0.1	0.1
Low Ref Cyl				31.00	
Low Avg			30.93		
Low Error%				0.1	
Mid Ref Cyl	4.9400	132.00	24.600	52.60	22.200
Mid Avg 4.9458	130.54	24.745	52.76	22.167	
Mid Error%	0.1	0.5	0.3	0.2	0.1
High Ref Cyl	9.4900	297.00	53.700	88.40	47.300
High Avg	9.4831	297.09	54.116	88.45	48.223
High Error%	0.1	0.0	0.8	0.0	2.0

Initial System		Bias Check for Run 1		IBIAS01 STRATA Version 2.0							
SO2		O2	CO	NOX	THC	SO2	O2	CO	NOX	THC	
Volts	Volts	%	ppm	ppm	ppm	ppm	Volts	Volts	Volts		
11-17-2009		12:03:47		10.6242	39.34	1.057	2.53	-0.997	10.6242		
0.3934	0.1057	0.253	-0.0997								
11-17-2009		12:04:47		0.2923	-2.23	0.658	8.58	-0.555	0.2923		
-0.0223	0.0658	0.858	-0.0555								
11-17-2009		12:05:47		0.0976	-0.89	0.426	0.71	-0.154	0.0976		
-0.0089	0.0426	0.071	-0.0154								
11-17-2009		12:06:47		0.1104	-1.07	0.377	0.08	-0.084	0.1104		
-0.0107	0.0377	0.008	-0.0084								
11-17-2009		12:07:46		4.4406	-1.12	0.379	3.16	-0.577	4.4406		
-0.0112	0.0379	0.316	-0.0577								
11-17-2009		12:08:47		4.9161	-2.20	0.380	52.02	-0.784	4.9161		
-0.0220	0.0380	5.202	-0.0784								
11-17-2009		12:09:46		5.0379	0.29	0.294	52.29	-0.672	5.0379		
0.0029	0.0294	5.229	-0.0672								
11-17-2009		12:10:47		0.7344	116.66	0.340	52.23	-0.558	0.7344		
1.1666	0.0340	5.223	-0.0558								
11-17-2009		12:11:46		0.0905	131.58	0.272	9.73	-0.162	0.0905		
1.3158	0.0272	0.973	-0.0162								
11-17-2009		12:12:47		0.1108	120.28	0.248	2.10	-0.340	0.1108		
1.2028	0.0248	0.210	-0.0340								
11-17-2009		12:13:47		0.0825	-0.17	18.095	1.48	0.370	0.0825		
-0.0017	1.8095	0.148	0.0370								
11-17-2009		12:14:47		0.0880	-2.07	24.364	1.32	0.674	0.0880		
-0.0207	2.4364	0.132	0.0674								
11-17-2009		12:15:47		0.1332	-1.98	21.980	1.20	0.329	0.1332		
-0.0198	2.1980	0.120	0.0329								
11-17-2009		12:16:46		0.0903	-2.40	0.362	1.15	-0.299	0.0903		
-0.0240	0.0362	0.115	-0.0299								
11-17-2009		12:17:47		0.1012	-1.88	0.415	1.02	-0.241	0.1012		
-0.0188	0.0415	0.102	-0.0241								
11-17-2009		12:18:46		0.1054	-1.95	0.628	0.98	0.202	0.1054		
-0.0195	0.0628	0.098	0.0202								
11-17-2009		12:19:47		0.1076	-1.74	0.798	1.13	1.449	0.1076		
-0.0174	0.0798	0.113	0.1449								
11-17-2009		12:20:46		0.1123	-2.36	0.927	0.95	2.918	0.1123		
-0.0236	0.0927	0.095	0.2918								
11-17-2009		12:21:47		0.1193	-2.31	0.973	1.03	4.229	0.1193		
-0.0231	0.0973	0.103	0.4229								
11-17-2009		12:22:46		0.1382	-2.27	0.973	0.91	5.694	0.1382		
-0.0227	0.0973	0.091	0.5694								
11-17-2009		12:23:47		0.2391	-2.82	1.238	0.99	20.505	0.2391		
-0.0282	0.1238	0.099	2.0505								

Initial System Bias Check for Run 1

Operator: JMM  
Plant Name: Conoco  
Location: Old Ocean, TX  
Reference Cylinder Numbers  
Zero Span  
O2 N2 cc9824  
CO N2 ALM014290  
NOX N2 cc241846  
THC AIR cc236179  
SO2 N2 cc144163

Date/Time 11-17-2009 12:24:25 PASSED  
Analyte O2 CO NOX THC SO2  
Units % ppm ppm ppm ppm

A15



				IBIAS01	
Zero Ref Cyl	0.0000	0.00	0.000	0.00	0.000
Zero Cal	0.0277	-0.33	0.082	0.07	-0.051
Zero Avg	0.0898	-0.89	0.406	0.15	-0.073
Zero Bias%	0.7	0.2	0.6	0.1	0.0
Zero Drift%					
Span Ref Cyl	4.9400	132.00	24.600	52.60	22.200
Span Cal	4.9458	130.54	24.745	52.76	22.167
Span Avg	4.9272	132.07	24.342	52.13	22.905
Span Bias%	0.2	0.5	0.8	0.6	1.6
Span Drift%					

Test Run 1 Begin. STRATA Version 2.0

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

SO2	O2	CO	NOx	THC	SO2	O2	CO	NOx	THC
Volts	Volts	%	ppm	ppm	ppm	Volts	Volts	Volts	
11-17-2009		12:25:26	0.2513	-3.88	1.047	1.01	15.399	0.2513	
-0.0388	0.1047	0.101	1.5399						
11-17-2009		12:26:26	8.0024*	-1.11	1.065	0.85	1.299	8.0024*	
-0.0111	0.1065	0.085	0.1299						
11-17-2009		12:27:26	13.1815*		-0.04	0.660	0.85	-1.112	
13.1815*		-0.0004	0.0660	0.085	-0.1112				
11-17-2009		12:28:26	13.1815*		2.16	0.478	0.83	-1.396	
13.1815*		0.0216	0.0478	0.083	-0.1396				
11-17-2009		12:29:25	13.1815*		2.28	0.477	0.90	-1.598	
13.1815*		0.0228	0.0477	0.090	-0.1598				
11-17-2009		12:30:26	13.1815*		3.00	0.476	0.93	-1.394	
13.1815*		0.0300	0.0476	0.093	-0.1394				
11-17-2009		12:31:25	13.1815*		3.17	0.477	0.93	-1.466	
13.1815*		0.0317	0.0477	0.093	-0.1466				
11-17-2009		12:32:26	13.1815*		3.39	0.476	0.88	-1.490	
13.1815*		0.0339	0.0476	0.088	-0.1490				
11-17-2009		12:33:25	13.1815*		3.70	0.476	0.89	-1.625	
13.1815*		0.0370	0.0476	0.089	-0.1625				
11-17-2009		12:34:26	13.1815*		4.10	0.477	0.89	-1.681	
13.1815*		0.0410	0.0477	0.089	-0.1681				
11-17-2009		12:35:26	13.1815*		0.70	0.709	0.95	-1.362	
13.1815*		0.0070	0.0709	0.095	-0.1362				
11-17-2009		12:36:26	13.1815*		-2.32	1.175	0.88	-0.614	
13.1815*		-0.0232	0.1175	0.088	-0.0614				
11-17-2009		12:37:26	13.1815*		-2.74	1.171	0.90	-1.066	
13.1815*		-0.0274	0.1171	0.090	-0.1066				
11-17-2009		12:38:25	13.1815*		-2.89	1.072	0.81	-1.118	
13.1815*		-0.0289	0.1072	0.081	-0.1118				
11-17-2009		12:39:26	13.1815*		-2.65	1.053	0.83	-1.331	
13.1815*		-0.0265	0.1053	0.083	-0.1331				
11-17-2009		12:40:25	13.1815*		-1.45	0.883	0.80	-1.493	
13.1815*		-0.0145	0.0883	0.080	-0.1493				
11-17-2009		12:41:26	13.1815*		-0.36	0.505	0.78	-1.360	
13.1815*		-0.0036	0.0505	0.078	-0.1360				
11-17-2009		12:42:25	13.1815*		-0.37	0.475	0.79	-1.306	
13.1815*		-0.0037	0.0475	0.079	-0.1306				
11-17-2009		12:43:26	13.1815*		-0.39	0.405	0.80	-1.551	
13.1815*		-0.0039	0.0405	0.080	-0.1551				
11-17-2009		12:44:25	8.1696*	24.86	3.770	0.87	-1.398	8.1696*	
0.2486	0.3770	0.087	-0.1398						
11-17-2009		12:45:26	2.9224	45.89	12.739	0.99	-0.697	2.9224	
0.4589	1.2739	0.099	-0.0697						
11-17-2009		12:46:26	2.8922	33.34	13.510	1.02	-0.625	2.8922	
0.3334	1.3510	0.102	-0.0625						
11-17-2009		12:47:25	2.8336	49.05	13.114	0.93	-0.731	2.8336	
0.4905	1.3114	0.093	-0.0731						
11-17-2009		12:48:26	2.8575	42.32	12.814	0.74	-0.759	2.8575	
0.4232	1.2814	0.074	-0.0759						
11-17-2009		12:49:25	2.8532	44.32	12.713	0.75	-0.621	2.8532	
0.4432	1.2713	0.075	-0.0621						
11-17-2009		12:50:26	2.8351	58.57	12.639	0.72	-0.695	2.8351	
0.5857	1.2639	0.072	-0.0695						
11-17-2009		12:51:25	2.6814	66.41	12.718	1.02	-0.596	2.6814	
0.6641	1.2718	0.102	-0.0596						
11-17-2009		12:52:26	2.7118	53.30	13.048	0.94	-0.670	2.7118	

RUN01

0.5330	1.3048	0.094	-0.0670						
11-17-2009		12:53:26		2.7293	44.92	12.704	0.98	-0.702	2.7293
0.4492	1.2704	0.098	-0.0702						
11-17-2009		12:54:26		2.6486	39.88	12.811	0.69	-1.088	2.6486
0.3988	1.2811	0.069	-0.1088						
11-17-2009		12:55:26		2.7882	43.37	13.194	0.62	-0.750	2.7882
0.4337	1.3194	0.062	-0.0750						
11-17-2009		12:56:25		2.5390	45.81	14.246	0.89	-0.933	2.5390
0.4581	1.4246	0.089	-0.0933						
11-17-2009		12:57:26		2.4873	33.10	13.627	0.49	-0.761	2.4873
0.3310	1.3627	0.049	-0.0761						
11-17-2009		12:58:25		3.2573	29.47	10.240	0.38	-1.429	3.2573
0.2947	1.0240	0.038	-0.1429						
11-17-2009		12:59:26		2.7140	41.08	13.416	0.53	-0.806	2.7140
0.4108	1.3416	0.053	-0.0806						
11-17-2009		13:00:25		2.6138	37.83	13.300	0.50	-0.854	2.6138
0.3783	1.3300	0.050	-0.0854						
11-17-2009		13:01:26		2.6294	35.72	13.123	0.53	-0.669	2.6294
0.3572	1.3123	0.053	-0.0669						
11-17-2009		13:02:25		2.6360	35.32	13.324	0.46	-0.837	2.6360
0.3532	1.3324	0.046	-0.0837						
11-17-2009		13:03:26		2.5258	45.85	13.392	0.57	-0.742	2.5258
0.4585	1.3392	0.057	-0.0742						
11-17-2009		13:04:26		2.5567	35.01	13.335	0.63	-0.763	2.5567
0.3501	1.3335	0.063	-0.0763						
11-17-2009		13:05:25		2.6276	40.25	13.345	0.39	-0.766	2.6276
0.4025	1.3345	0.039	-0.0766						
11-17-2009		13:06:26		2.5351	45.45	13.407	0.65	-0.744	2.5351
0.4545	1.3407	0.065	-0.0744						
11-17-2009		13:07:25		2.3626	58.38	13.517	0.66	-0.750	2.3626
0.5838	1.3517	0.066	-0.0750						
11-17-2009		13:08:26		2.4715	41.51	13.907	0.59	-0.758	2.4715
0.4151	1.3907	0.059	-0.0758						
11-17-2009		13:09:25		2.4642	46.10	13.716	0.81	-0.677	2.4642
0.4610	1.3716	0.081	-0.0677						
11-17-2009		13:10:26		2.4330	54.60	13.608	0.50	-0.742	2.4330
0.5460	1.3608	0.050	-0.0742						
11-17-2009		13:11:25		2.4499	44.43	13.703	0.73	-0.635	2.4499
0.4443	1.3703	0.073	-0.0635						
11-17-2009		13:12:26		2.4885	34.36	13.922	0.59	-0.779	2.4885
0.3436	1.3922	0.059	-0.0779						
11-17-2009		13:13:26		4.9087*	38.46	13.526	0.49	-0.741	4.9087*
0.3846	1.3526	0.049	-0.0741						
11-17-2009		13:14:25		9.4339*	25.95	4.893	0.66	-1.380	9.4339*
0.2595	0.4893	0.066	-0.1380						
11-17-2009		13:15:26		2.8879	51.96	13.238	0.78	-0.991	2.8879
0.5196	1.3238	0.078	-0.0991						
11-17-2009		13:16:25		2.9744	48.63	12.936	0.73	-0.937	2.9744
0.4863	1.2936	0.073	-0.0937						
11-17-2009		13:17:26		4.5754	44.97	7.223	0.64	-1.526	4.5754
0.4497	0.7223	0.064	-0.1526						
11-17-2009		13:18:25		6.2539	43.11	6.643	1.29	-1.680	6.2539
0.4311	0.6643	0.129	-0.1680						
11-17-2009		13:19:26		7.2831	43.79	6.528	0.57	-1.607	7.2831
0.4379	0.6528	0.057	-0.1607						
11-17-2009		13:20:25		7.8909	42.15	6.661	0.54	-1.514	7.8909
0.4215	0.6661	0.054	-0.1514						
11-17-2009		13:21:26		8.2304	40.24	6.581	0.69	-1.688	8.2304
0.4024	0.6581	0.069	-0.1688						
11-17-2009		13:22:26		8.4587	38.02	6.564	0.25	-1.475	8.4587
0.3802	0.6564	0.025	-0.1475						
11-17-2009		13:23:26		8.5990	34.13	6.438	0.24	-1.651	8.5990
0.3413	0.6438	0.024	-0.1651						

11-17-2009	13:24:26	8.7396	RUN01 30.80	6.541	0.42	-1.537	8.7396
0.3080 0.6541	0.042 -0.1537						
11-17-2009	13:25:25	8.8355	30.31	6.480	0.43	-1.601	8.8355
0.3031 0.6480	0.043 -0.1601						
11-17-2009	13:26:26	8.8763	32.55	6.724	0.69	-1.618	8.8763
0.3255 0.6724	0.069 -0.1618						
11-17-2009	13:27:25	8.8710	35.78	6.656	0.79	-1.679	8.8710
0.3578 0.6656	0.079 -0.1679						
11-17-2009	13:28:26	8.8623	37.74	6.639	0.75	-1.691	8.8623
0.3774 0.6639	0.075 -0.1691						
11-17-2009	13:29:25	8.8566	40.23	6.819	0.73	-1.615	8.8566
0.4023 0.6819	0.073 -0.1615						
11-17-2009	13:30:26	8.8579	40.94	6.653	0.67	-1.634	8.8579
0.4094 0.6653	0.067 -0.1634						
11-17-2009	13:31:26	8.8620	40.99	6.606	0.64	-1.665	8.8620
0.4099 0.6606	0.064 -0.1665						
11-17-2009	13:32:26	8.8351	42.50	6.683	0.80	-1.617	8.8351
0.4250 0.6683	0.080 -0.1617						
11-17-2009	13:33:26	8.8171	41.64	6.728	0.40	-1.789	8.8171
0.4164 0.6728	0.040 -0.1789						
11-17-2009	13:34:25	8.8279	38.94	6.519	0.34	-1.678	8.8279
0.3894 0.6519	0.034 -0.1678						
11-17-2009	13:35:26	8.8615	36.59	6.451	0.74	-1.646	8.8615
0.3659 0.6451	0.074 -0.1646						
11-17-2009	13:36:25	8.8757	34.92	6.308	0.35	-1.737	8.8757
0.3492 0.6308	0.035 -0.1737						
11-17-2009	13:37:26	8.8839	35.92	6.265	0.29	-1.674	8.8839
0.3592 0.6265	0.029 -0.1674						
11-17-2009	13:38:25	8.8974	33.33	6.087	0.66	-1.713	8.8974
0.3333 0.6087	0.066 -0.1713						
11-17-2009	13:39:26	8.8944	35.24	6.168	0.28	-1.740	8.8944
0.3524 0.6168	0.028 -0.1740						
11-17-2009	13:40:26	8.9039	34.64	6.147	0.52	-1.754	8.9039
0.3464 0.6147	0.052 -0.1754						
11-17-2009	13:41:26	8.9012	34.60	5.945	0.54	-1.702	8.9012
0.3460 0.5945	0.054 -0.1702						
11-17-2009	13:42:26	8.9088	35.02	5.988	0.41	-1.703	8.9088
0.3502 0.5988	0.041 -0.1703						
11-17-2009	13:43:25	8.9318	35.57	6.074	0.44	-1.749	8.9318
0.3557 0.6074	0.044 -0.1749						
11-17-2009	13:44:26	8.9342	33.98	6.314	0.43	-1.802	8.9342
0.3398 0.6314	0.043 -0.1802						
11-17-2009	13:45:25	8.9355	33.25	6.161	0.64	-1.756	8.9355
0.3325 0.6161	0.064 -0.1756						
11-17-2009	13:46:26	8.9354	33.94	6.255	0.63	-1.758	8.9354
0.3394 0.6255	0.063 -0.1758						
11-17-2009	13:47:25	8.9115	36.41	6.462	0.64	-1.800	8.9115
0.3641 0.6462	0.064 -0.1800						
11-17-2009	13:48:26	8.8484	38.56	6.556	0.59	-1.830	8.8484
0.3856 0.6556	0.059 -0.1830						
11-17-2009	13:49:26	8.7896	40.22	6.676	1.05	-1.804	8.7896
0.4022 0.6676	0.105 -0.1804						
11-17-2009	13:50:26	8.7455	41.43	6.533	0.73	-1.752	8.7455
0.4143 0.6533	0.073 -0.1752						
11-17-2009	13:51:26	8.7282	41.67	6.630	0.64	-1.754	8.7282
0.4167 0.6630	0.064 -0.1754						
11-17-2009	13:52:25	8.7147	39.16	6.547	0.25	-1.788	8.7147
0.3916 0.6547	0.025 -0.1788						
11-17-2009	13:53:26	8.7474	35.84	6.573	0.53	-1.804	8.7474
0.3584 0.6573	0.053 -0.1804						
11-17-2009	13:54:25	8.7694	35.29	6.539	0.76	-1.771	8.7694
0.3529 0.6539	0.076 -0.1771						
11-17-2009	13:55:26	8.7385	37.29	6.711	0.50	-1.806	8.7385

RUN01

0.3729	0.6711	0.050	-0.1806						
11-17-2009		13:56:25		8.7192	37.55	6.544	0.56	-1.618	8.7192
0.3755	0.6544	0.056	-0.1618						
11-17-2009		13:57:26		8.7657	35.69	6.152	0.47	-1.790	8.7657
0.3569	0.6152	0.047	-0.1790						
11-17-2009		13:58:25		8.7916	34.85	5.748	0.80	-1.816	8.7916
0.3485	0.5748	0.080	-0.1816						
11-17-2009		13:59:26		8.8056	36.51	5.367	0.67	-1.816	8.8056
0.3651	0.5367	0.067	-0.1816						
11-17-2009		14:00:26		8.7741	39.17	5.711	0.67	-1.578	8.7741
0.3917	0.5711	0.067	-0.1578						
11-17-2009		14:01:25		8.7309	40.42	5.863	0.34	-1.730	8.7309
0.4042	0.5863	0.034	-0.1730						
11-17-2009		14:02:26		8.7192	38.34	5.844	0.42	-1.661	8.7192
0.3834	0.5844	0.042	-0.1661						
11-17-2009		14:03:25		8.7681	35.80	5.758	0.48	-1.766	8.7681
0.3580	0.5758	0.048	-0.1766						
11-17-2009		14:04:26		8.7824	35.92	5.814	0.78	-1.772	8.7824
0.3592	0.5814	0.078	-0.1772						
11-17-2009		14:05:25		8.7446	38.34	6.106	0.81	-1.675	8.7446
0.3834	0.6106	0.081	-0.1675						
11-17-2009		14:06:26		8.6786	39.00	6.009	0.53	-1.655	8.6786
0.3900	0.6009	0.053	-0.1655						
11-17-2009		14:07:25		8.6393	36.98	5.782	0.43	-1.768	8.6393
0.3698	0.5782	0.043	-0.1768						
11-17-2009		14:08:26		8.6509	34.15	5.638	0.29	-1.641	8.6509
0.3415	0.5638	0.029	-0.1641						
11-17-2009		14:09:26		8.6802	30.91	5.779	0.66	-1.736	8.6802
0.3091	0.5779	0.066	-0.1736						
11-17-2009		14:10:26		8.6782	34.22	5.702	0.95	-1.672	8.6782
0.3422	0.5702	0.095	-0.1672						
11-17-2009		14:11:26		8.6441	38.61	5.947	0.65	-1.689	8.6441
0.3861	0.5947	0.065	-0.1689						
11-17-2009		14:12:25		8.6212	40.46	6.129	0.42	-1.700	8.6212
0.4046	0.6129	0.042	-0.1700						
11-17-2009		14:13:26		8.6292	38.87	5.885	0.57	-1.725	8.6292
0.3887	0.5885	0.057	-0.1725						
11-17-2009		14:14:25		8.6723	36.83	5.612	0.44	-1.786	8.6723
0.3683	0.5612	0.044	-0.1786						
11-17-2009		14:15:26		8.6811	35.29	5.886	0.45	-1.660	8.6811
0.3529	0.5886	0.045	-0.1660						
11-17-2009		14:16:25		8.7020	35.47	5.928	0.48	-1.791	8.7020
0.3547	0.5928	0.048	-0.1791						
11-17-2009		14:17:26		8.7230	35.74	5.922	0.39	-1.822	8.7230
0.3574	0.5922	0.039	-0.1822						
11-17-2009		14:18:26		8.6977	35.00	6.087	0.62	-1.684	8.6977
0.3500	0.6087	0.062	-0.1684						
11-17-2009		14:19:26		8.6724	34.92	6.039	0.66	-1.689	8.6724
0.3492	0.6039	0.066	-0.1689						
11-17-2009		14:20:26		8.6562	37.76	6.076	0.79	-1.642	8.6562
0.3776	0.6076	0.079	-0.1642						
11-17-2009		14:21:25		8.6085	40.75	6.094	0.44	-1.706	8.6085
0.4075	0.6094	0.044	-0.1706						
11-17-2009		14:22:26		8.5726	40.11	5.986	0.86	-1.723	8.5726
0.4011	0.5986	0.086	-0.1723						
11-17-2009		14:23:25		8.5506	38.79	6.030	0.40	-1.806	8.5506
0.3879	0.6030	0.040	-0.1806						
11-17-2009		14:24:26		8.5452	38.79	6.047	0.49	-1.738	8.5452
0.3879	0.6047	0.049	-0.1738						
11-17-2009		14:25:25		8.5612	37.71	6.029	0.50	-1.856	8.5612
0.3771	0.6029	0.050	-0.1856						
11-17-2009		14:26:26		5.2971	45.22	8.131	0.45	-1.509	5.2971
0.4522	0.8131	0.045	-0.1509						

11-17-2009	14:27:26	2.7972	RUN01 51.29	13.403	0.63	-1.220	2.7972
0.5129 1.3403	0.063 -0.1220						
11-17-2009	14:28:26	2.9410	52.82	13.040	0.63	-1.077	2.9410
0.5282 1.3040	0.063 -0.1077						
11-17-2009	14:29:26	3.0188	50.65	12.795	1.01	-1.142	3.0188
0.5065 1.2795	0.101 -0.1142						
11-17-2009	14:30:25	2.9845	48.48	13.137	0.53	-1.041	2.9845
0.4848 1.3137	0.053 -0.1041						
11-17-2009	14:31:26	3.0382	57.57	13.110	0.70	-1.093	3.0382
0.5757 1.3110	0.070 -0.1093						
11-17-2009	14:32:25	3.0316	45.76	13.300	0.68	-1.056	3.0316
0.4576 1.3300	0.068 -0.1056						
11-17-2009	14:33:26	3.2211	47.01	13.116	0.59	-1.125	3.2211
0.4701 1.3116	0.059 -0.1125						
11-17-2009	14:34:25	3.2523	37.76	13.127	0.58	-1.057	3.2523
0.3776 1.3127	0.058 -0.1057						
11-17-2009	14:35:26	3.2122	25.99	13.425	0.18	-1.056	3.2122
0.2599 1.3425	0.018 -0.1056						
11-17-2009	14:36:26	3.1561	40.39	13.296	0.20	-1.006	3.1561
0.4039 1.3296	0.020 -0.1006						
11-17-2009	14:37:26	3.2082	34.99	13.365	0.26	-1.063	3.2082
0.3499 1.3365	0.026 -0.1063						
11-17-2009	14:38:26	3.2004	35.66	13.422	0.29	-1.137	3.2004
0.3566 1.3422	0.029 -0.1137						
11-17-2009	14:39:25	3.1317	62.63	13.285	0.61	-0.958	3.1317
0.6263 1.3285	0.061 -0.0958						
11-17-2009	14:40:26	3.0542	52.61	13.546	0.61	-1.059	3.0542
0.5261 1.3546	0.061 -0.1059						
11-17-2009	14:41:25	3.0218	36.34	13.419	0.64	-1.088	3.0218
0.3634 1.3419	0.064 -0.1088						
11-17-2009	14:42:26	3.1602	49.67	13.247	0.22	-1.140	3.1602
0.4967 1.3247	0.022 -0.1140						
11-17-2009	14:43:25	3.0895	66.72	13.258	0.60	-1.054	3.0895
0.6672 1.3258	0.060 -0.1054						
11-17-2009	14:44:26	2.9870	67.42	13.485	0.96	-0.912	2.9870
0.6742 1.3485	0.096 -0.0912						
11-17-2009	14:45:25	2.9371	59.55	13.688	0.99	-0.935	2.9371
0.5955 1.3688	0.099 -0.0935						
11-17-2009	14:46:26	2.8811	50.81	13.920	0.87	-1.169	2.8811
0.5081 1.3920	0.087 -0.1169						
11-17-2009	14:47:26	2.9746	30.62	13.785	0.32	-0.918	2.9746
0.3062 1.3785	0.032 -0.0918						
11-17-2009	14:48:26	3.0212	43.32	13.613	0.29	-0.989	3.0212
0.4332 1.3613	0.029 -0.0989						
11-17-2009	14:49:26	2.9886	42.45	13.481	0.52	-1.036	2.9886
0.4245 1.3481	0.052 -0.1036						
11-17-2009	14:50:25	2.9759	35.06	13.691	0.43	-0.957	2.9759
0.3506 1.3691	0.043 -0.0957						
11-17-2009	14:51:26	2.8402	49.89	13.677	0.51	-0.843	2.8402
0.4989 1.3677	0.051 -0.0843						
11-17-2009	14:52:26	2.7001	53.48	13.884	0.50	-0.967	2.7001
0.5348 1.3884	0.050 -0.0967						
11-17-2009	14:53:25	2.6675	56.73	14.046	0.61	-0.959	2.6675
0.5673 1.4046	0.061 -0.0959						
11-17-2009	14:54:26	2.6465	57.20	13.991	0.82	-0.930	2.6465
0.5720 1.3991	0.082 -0.0930						
11-17-2009	14:55:26	2.6691	59.22	13.948	0.84	-0.934	2.6691
0.5922 1.3948	0.084 -0.0934						
11-17-2009	14:56:25	2.7446	45.61	14.144	0.72	-1.066	2.7446
0.4561 1.4144	0.072 -0.1066						
11-17-2009	14:57:26	2.7372	60.34	13.866	0.72	-0.955	2.7372
0.6034 1.3866	0.072 -0.0955						
11-17-2009	14:58:25	2.7986	66.21	13.832	0.86	-0.994	2.7986

				RUN01				
0.6621	1.3832	0.086	-0.0994					
11-17-2009		14:59:26		2.8924	49.49	13.909	0.66	-1.015 2.8924
0.4949	1.3909	0.066	-0.1015					
11-17-2009		15:00:25		2.9270	52.66	13.808	0.52	-0.926 2.9270
0.5266	1.3808	0.052	-0.0926					
11-17-2009		15:01:26		2.9868	44.04	13.879	0.46	-1.101 2.9868
0.4404	1.3879	0.046	-0.1101					
11-17-2009		15:02:25		2.9675	45.84	13.762	0.39	-1.048 2.9675
0.4584	1.3762	0.039	-0.1048					
11-17-2009		15:03:26		2.8689	47.88	14.055	0.55	-0.983 2.8689
0.4788	1.4055	0.055	-0.0983					
11-17-2009		15:04:26		2.7953	53.99	14.094	0.99	-0.959 2.7953
0.5399	1.4094	0.099	-0.0959					
11-17-2009		15:05:26		2.7464	54.81	13.944	0.42	-0.932 2.7464
0.5481	1.3944	0.042	-0.0932					
11-17-2009		15:06:26		2.6527	48.03	14.228	0.65	-0.972 2.6527
0.4803	1.4228	0.065	-0.0972					
11-17-2009		15:07:26		2.6132	66.21	13.961	0.84	-0.895 2.6132
0.6621	1.3961	0.084	-0.0895					
11-17-2009		15:08:26		2.6434	50.28	14.145	0.80	-1.008 2.6434
0.5028	1.4145	0.080	-0.1008					
11-17-2009		15:09:25		2.6817	50.84	14.111	0.67	-0.817 2.6817
0.5084	1.4111	0.067	-0.0817					
Begin calculating run averages								
11-17-2009		15:11:03		2.6925	61.89	14.024	0.84	-0.965 2.6925
0.6189	1.4024	0.084	-0.0965					
11-17-2009		15:12:04		2.6182	60.10	14.164	0.95	-0.743 2.6182
0.6010	1.4164	0.095	-0.0743					
11-17-2009		15:13:03		2.7972	43.61	13.864	0.60	-0.857 2.7972
0.4361	1.3864	0.060	-0.0857					
11-17-2009		15:14:04		2.8341	43.08	13.718	0.37	-0.972 2.8341
0.4308	1.3718	0.037	-0.0972					
11-17-2009		15:15:03		2.7612	50.16	13.757	0.39	-0.923 2.7612
0.5016	1.3757	0.039	-0.0923					
11-17-2009		15:16:04		2.8170	44.31	13.620	0.42	-0.843 2.8170
0.4431	1.3620	0.042	-0.0843					
11-17-2009		15:17:03		2.7703	46.18	13.578	0.46	-0.907 2.7703
0.4618	1.3578	0.046	-0.0907					
11-17-2009		15:18:04		2.8424	62.83	13.531	0.44	-0.921 2.8424
0.6283	1.3531	0.044	-0.0921					
11-17-2009		15:19:03		2.8752	72.98	13.519	0.47	-0.996 2.8752
0.7298	1.3519	0.047	-0.0996					
11-17-2009		15:20:03		2.7872	82.17	13.417	1.41	-0.894 2.7872
0.8217	1.3417	0.141	-0.0894					
11-17-2009		15:21:04		2.6773	72.72	13.624	0.78	-0.888 2.6773
0.7272	1.3624	0.078	-0.0888					
11-17-2009		15:22:03		2.7407	54.37	13.604	0.91	-0.876 2.7407
0.5437	1.3604	0.091	-0.0876					
11-17-2009		15:23:04		2.8237	48.57	13.580	0.39	-0.907 2.8237
0.4857	1.3580	0.039	-0.0907					
11-17-2009		15:24:03		2.8393	57.72	13.444	0.39	-0.923 2.8393
0.5772	1.3444	0.039	-0.0923					
11-17-2009		15:25:04		2.8808	69.43	13.479	0.57	-0.904 2.8808
0.6943	1.3479	0.057	-0.0904					
11-17-2009		15:26:03		2.9290	61.45	13.395	0.60	-0.770 2.9290
0.6145	1.3395	0.060	-0.0770					
11-17-2009		15:27:03		2.9269	84.15	13.411	0.46	-0.792 2.9269
0.8415	1.3411	0.046	-0.0792					
11-17-2009		15:28:03		2.8405	66.82	13.789	0.60	-0.809 2.8405
0.6682	1.3789	0.060	-0.0809					
11-17-2009		15:29:03		2.7310	76.77	13.834	1.13	-0.759 2.7310
0.7677	1.3834	0.113	-0.0759					
11-17-2009		15:30:04		2.8599	50.56	13.967	0.70	-0.735 2.8599

RUN01

0.5056	1.3967	0.070	-0.0735						
11-17-2009		15:31:03		2.9722	63.39	13.733	0.39	-0.694	2.9722
0.6339	1.3733	0.039	-0.0694						
11-17-2009		15:32:04		2.9891	59.86	13.714	0.51	-0.746	2.9891
0.5986	1.3714	0.051	-0.0746						
11-17-2009		15:33:03		2.9560	54.26	13.856	0.42	-0.790	2.9560
0.5426	1.3856	0.042	-0.0790						
11-17-2009		15:34:04		3.0272	71.47	13.703	0.55	-0.718	3.0272
0.7147	1.3703	0.055	-0.0718						
11-17-2009		15:35:03		3.0341	52.73	13.692	0.55	-0.786	3.0341
0.5273	1.3692	0.055	-0.0786						
11-17-2009		15:36:04		2.9820	41.32	13.860	0.23	-0.765	2.9820
0.4132	1.3860	0.023	-0.0765						
11-17-2009		15:37:03		2.9928	48.94	13.809	0.25	-0.772	2.9928
0.4894	1.3809	0.025	-0.0772						
11-17-2009		15:38:03		2.9275	42.28	14.077	0.43	-0.795	2.9275
0.4228	1.4077	0.043	-0.0795						
11-17-2009		15:39:04		2.9912	38.74	13.912	0.18	-0.756	2.9912
0.3874	1.3912	0.018	-0.0756						
11-17-2009		15:40:03		2.9706	43.78	13.907	0.38	-0.710	2.9706
0.4378	1.3907	0.038	-0.0710						
11-17-2009		15:41:04		2.8338	57.71	13.732	0.33	-0.741	2.8338
0.5771	1.3732	0.033	-0.0741						
11-17-2009		15:42:03		2.7960	53.29	14.010	0.68	-0.782	2.7960
0.5329	1.4010	0.068	-0.0782						
11-17-2009		15:43:04		2.7147	46.06	14.255	0.60	-0.760	2.7147
0.4606	1.4255	0.060	-0.0760						
11-17-2009		15:44:03		2.7075	37.23	14.192	0.54	-0.733	2.7075
0.3723	1.4192	0.054	-0.0733						
11-17-2009		15:45:04		2.8684	39.86	13.919	0.24	-0.576	2.8684
0.3986	1.3919	0.024	-0.0576						
11-17-2009		15:46:03		2.8729	57.99	13.841	0.39	-0.631	2.8729
0.5799	1.3841	0.039	-0.0631						
11-17-2009		15:47:03		2.7629	58.53	13.894	0.81	-0.630	2.7629
0.5853	1.3894	0.081	-0.0630						
11-17-2009		15:48:04		2.7441	50.28	14.210	0.60	-0.704	2.7441
0.5028	1.4210	0.060	-0.0704						
11-17-2009		15:49:03		2.6808	51.22	14.144	0.96	-0.771	2.6808
0.5122	1.4144	0.096	-0.0771						
11-17-2009		15:50:04		2.7717	37.98	14.058	0.79	-0.543	2.7717
0.3798	1.4058	0.079	-0.0543						
11-17-2009		15:51:03		2.8861	47.50	13.928	0.48	-0.741	2.8861
0.4750	1.3928	0.048	-0.0741						
11-17-2009		15:52:04		2.7817	64.42	13.762	0.68	-0.810	2.7817
0.6442	1.3762	0.068	-0.0810						
11-17-2009		15:53:03		2.7275	62.03	13.944	0.99	-0.741	2.7275
0.6203	1.3944	0.099	-0.0741						
11-17-2009		15:54:04		2.6598	51.39	13.949	0.73	-0.670	2.6598
0.5139	1.3949	0.073	-0.0670						
11-17-2009		15:55:03		2.7109	75.44	13.877	0.91	-0.612	2.7109
0.7544	1.3877	0.091	-0.0612						
11-17-2009		15:56:03		2.6757	76.87	14.082	1.25	-0.734	2.6757
0.7687	1.4082	0.125	-0.0734						
11-17-2009		15:57:04		2.6753	62.67	14.030	1.07	-0.700	2.6753
0.6267	1.4030	0.107	-0.0700						
11-17-2009		15:58:03		2.6952	70.42	14.093	0.89	-0.694	2.6952
0.7042	1.4093	0.089	-0.0694						
11-17-2009		15:59:04		2.6926	57.59	14.222	0.92	-0.662	2.6926
0.5759	1.4222	0.092	-0.0662						
11-17-2009		16:00:03		2.6392	74.13	13.995	1.30	-0.681	2.6392
0.7413	1.3995	0.130	-0.0681						
11-17-2009		16:01:04		2.6820	49.08	14.044	1.05	-0.708	2.6820
0.4908	1.4044	0.105	-0.0708						



11-17-2009		16:02:03	2.7687	RUN01 61.32	14.077	0.81	-0.772	2.7687	
0.6132	1.4077	0.081 -0.0772							
11-17-2009		16:03:04	2.6367	63.07	14.047	0.94	-0.707	2.6367	
0.6307	1.4047	0.094 -0.0707							
11-17-2009		16:04:03	2.6416	47.30	14.090	1.05	-0.456	2.6416	
0.4730	1.4090	0.105 -0.0456							
11-17-2009		16:05:03	2.7495	41.25	14.090	0.62	-0.604	2.7495	
0.4125	1.4090	0.062 -0.0604							
11-17-2009		16:06:03	2.8438	50.23	13.695	0.49	-0.630	2.8438	
0.5023	1.3695	0.049 -0.0630							
11-17-2009		16:07:03	2.8537	63.42	13.579	0.81	-0.666	2.8537	
0.6342	1.3579	0.081 -0.0666							
11-17-2009		16:08:04	2.9261	52.91	13.647	0.97	-0.714	2.9261	
0.5291	1.3647	0.097 -0.0714							
11-17-2009		16:09:03	2.8965	50.39	13.718	0.71	-0.580	2.8965	
0.5039	1.3718	0.071 -0.0580							
11-17-2009		16:10:04	2.9093	50.36	13.610	0.70	-0.702	2.9093	
0.5036	1.3610	0.070 -0.0702							
Average of Test Run			O2	CO	NOx	THC	SO2	O2	CO
NOx	THC	SO2							
ppm		%	ppm	ppm	ppm	%	ppm	ppm	ppm
11-17-2009		16:10:04	2.8115	56.48	13.839	0.67	-0.756	2.8115	
0.5648	1.3839	0.067 -0.0756							
Test Run 1 End									

FBIAS01  
STRATA Version 2.0

Final System Bias Check for Run 1

SO2		O2	CO	NOX	THC	SO2	O2	CO	NOX	THC
Volts	Volts	%	ppm	ppm	ppm	ppm	Volts	Volts	Volts	
11-17-2009		16:11:09		2.8919	55.18	13.726	0.57	-0.623	2.8919	
0.5518	1.3726	0.057	-0.0623							
11-17-2009		16:12:09		2.1690	41.19	13.504	6.77	-0.631	2.1690	
0.4119	1.3504	0.677	-0.0631							
11-17-2009		16:13:08		0.2245	2.17	0.984	0.87	-0.352	0.2245	
0.0217	0.0984	0.087	-0.0352							
11-17-2009		16:14:09		0.3646	2.28	0.467	0.54	-0.459	0.3646	
0.0228	0.0467	0.054	-0.0459							
11-17-2009		16:15:10		4.7248	1.87	0.500	8.66	-1.047	4.7248	
0.0187	0.0500	0.866	-0.1047							
11-17-2009		16:16:09		5.0530	1.53	0.425	51.17	-1.316	5.0530	
0.0153	0.0425	5.117	-0.1316							
11-17-2009		16:17:09		5.0013	5.62	0.369	52.25	-1.305	5.0013	
0.0562	0.0369	5.225	-0.1305							
11-17-2009		16:18:09		0.6003	130.47	0.363	43.53	-0.977	0.6003	
1.3047	0.0363	4.353	-0.0977							
11-17-2009		16:19:09		0.2582	134.30	0.334	22.43	-0.695	0.2582	
1.3430	0.0334	2.243	-0.0695							
11-17-2009		16:20:09		0.2633	9.20	16.935	2.46	-0.218	0.2633	
0.0920	1.6935	0.246	-0.0218							
11-17-2009		16:21:09		0.2843	1.86	24.012	1.71	0.273	0.2843	
0.0186	2.4012	0.171	0.0273							
11-17-2009		16:22:09		1.3287	12.54	15.322	10.47	2.991	1.3287	
0.1254	1.5322	1.047	0.2991							
11-17-2009		16:23:09		1.2919	28.41	0.560	1.46	15.308	1.2919	
0.2841	0.0560	0.146	1.5308							
11-17-2009		16:24:09		0.7283	37.64	0.173	0.92	20.673	0.7283	
0.3764	0.0173	0.092	2.0673							
11-17-2009		16:25:09		1.8045	54.79	0.094	0.85	22.570	1.8045	
0.5479	0.0094	0.085	2.2570							

Final System Bias Check for Run 1

Operator: JMM  
 Plant Name: Conoco  
 Location: Old Ocean, TX  
 Reference Cylinder Numbers  
 Zero Span  
 O2 N2 cc9824  
 CO N2 ALM014290  
 NOX N2 cc241846  
 THC AIR cc236179  
 SO2 N2 cc144163

Date/Time	11-17-2009	16:25:15	PASSED		
Analyte O2	CO	NOX	THC	SO2	
Units %	ppm	ppm	ppm	ppm	
Zero Ref Cyl	0.0000	0.00	0.000	0.00	0.000
Zero Cal	0.0277	-0.33	0.082	0.07	-0.051
Zero Avg	0.2382	2.34	0.527	0.81	-0.255
Zero Bias%	2.2	0.9	0.8	0.7	0.4
Zero Drift%	1.6	1.1	0.2	0.7	-0.4
Span Ref Cyl	4.9400	132.00	24.600	52.60	22.200
Span Cal	4.9458	130.54	24.745	52.76	22.167
Span Avg	5.0551	134.63	23.652	52.19	22.550
Span Bias%	1.2	1.4	2.0	0.6	0.8
Span Drift%	1.3	0.9	-1.3	0.1	-0.8
Ini Zero Avg	0.0898	-0.89	0.406	0.15	-0.073

				FBIAS01	
Ini Span Avg	4.9272	132.07	24.342	52.13	22.905
Run Avg	2.8115	56.48	13.839	0.67	-0.756
Co	0.1640	0.73	0.466	0.48	-0.164
Cm	4.9911	133.35	23.997	52.16	22.727
Correct Avg	2.7093	55.49	13.980	0.19	-0.575

CALERR02 Calibration Error Test at Run 2 . STRATA Version 2.0										
SO2		O2	CO	NOX	THC	SO2	O2	CO	NOX	THC
		%	ppm	ppm	ppm	ppm	Volts	Volts	Volts	
Volts	Volts									
11-18-2009		09:01:02		0.7489	48.72	1.845	1.54	-2.135	0.7489	
0.4872	0.1845	0.154	-0.2135							
11-18-2009		09:02:03		0.6980	39.63	0.419	2.78	-1.398	0.6980	
0.3963	0.0419	0.278	-0.1398							
11-18-2009		09:03:02		0.8948	39.65	0.147	0.80	-1.176	0.8948	
0.3965	0.0147	0.080	-0.1176							
11-18-2009		09:04:02		0.0514	18.96	0.079	0.64	-1.195	0.0514	
0.1896	0.0079	0.064	-0.1195							
11-18-2009		09:05:02		0.0911	9.50	0.080	0.55	0.152	0.0911	
0.0950	0.0080	0.055	0.0152							
11-18-2009		09:06:02		0.1159	8.67	0.080	0.50	0.298	0.1159	
0.0867	0.0080	0.050	0.0298							
11-18-2009		09:07:02		0.1339	4.90	0.044	0.47	0.223	0.1339	
0.0490	0.0044	0.047	0.0223							
11-18-2009		09:08:02		0.1443	0.52	-0.144	0.44	0.556	0.1443	
0.0052	-0.0144	0.044	0.0556							
11-18-2009		09:09:02		2.3652	5.65	-0.194	0.41	2.747	2.3652	
0.0565	-0.0194	0.041	0.2747							
11-18-2009		09:10:02		5.0814	7.54	18.798	12.81	5.235	5.0814	
0.0754	1.8798	1.281	0.5235							
11-18-2009		09:11:03		4.9976	7.11	27.534	90.46	20.692	4.9976	
0.0711	2.7534	9.046	2.0692							
11-18-2009		09:12:02		4.9569	7.97	27.490	88.16	23.171	4.9569	
0.0797	2.7490	8.816	2.3171							
11-18-2009		09:13:02		4.9670	7.34	25.855	88.39	22.172	4.9670	
0.0734	2.5855	8.839	2.2172							
11-18-2009		09:14:02		4.9561	7.71	24.760	88.45	21.863	4.9561	
0.0771	2.4760	8.845	2.1863							
11-18-2009		09:15:02		6.8948	9.34	22.758	88.49	5.268	6.8948	
0.0934	2.2758	8.849	0.5268							
11-18-2009		09:16:02		9.4039	9.17	54.640	58.41	19.454	9.4039	
0.0917	5.4640	5.841	1.9454							
11-18-2009		09:17:02		9.4851	9.59	54.716	52.79	45.989	9.4851	
0.0959	5.4716	5.279	4.5989							
11-18-2009		09:18:02		9.4979	10.04	53.859	52.73	47.490	9.4979	
0.1004	5.3859	5.273	4.7490							
11-18-2009		09:19:02		7.7876	30.00	53.894	52.69	38.936	7.7876	
0.3000	5.3894	5.269	3.8936							
11-18-2009		09:20:03		0.1818	116.64	8.885	49.52	1.833	0.1818	
1.1664	0.8885	4.952	0.1833							
11-18-2009		09:21:02		0.1402	133.32	0.519	30.86	0.007	0.1402	
1.3332	0.0519	3.086	0.0007							
11-18-2009		09:22:02		0.1469	132.40	0.464	30.82	-0.187	0.1469	
1.3240	0.0464	3.082	-0.0187							
11-18-2009		09:23:02		0.1527	132.74	0.466	30.83	-0.298	0.1527	
1.3274	0.0466	3.083	-0.0298							
11-18-2009		09:24:02		0.1543	132.65	0.464	30.82	-0.320	0.1543	
1.3265	0.0464	3.082	-0.0320							
11-18-2009		09:25:02		0.4967	233.00	0.462	30.83	-0.384	0.4967	
2.3300	0.0462	3.083	-0.0384							
11-18-2009		09:26:02		0.1419	295.51	0.461	30.82	-0.330	0.1419	
2.9551	0.0461	3.082	-0.0330							

Calibration Error Test at Run 2  
 Operator: JMM  
 Plant Name: Conoco  
 Location: Old Ocean, TX  
 Reference Cylinder Numbers

A27

		CALERR02		
	Zero	Low-range	Mid-range	High-range
O2	N2	cc9824	cc55212	
CO	N2	ALM014290	cc111686	
NOx	N2	cc161053	cc241846	cc61285
THC	AIR	cc161053	cc236179	cc151954
SO2	N2	cc204173	cc144163	cc57866

Date/Time	11-18-2009		09:26:18		PASSED
Analyte O2	CO	NOx	THC	SO2	
Units %	ppm	ppm	ppm	ppm	
Zero Ref Cyl	0.0000	0.00	0.000	0.00	0.000
Zero Avg	0.1396	2.22	0.038	0.46	0.171
Zero Error%	1.5	0.7	0.1	0.5	0.4
Low Ref Cyl				31.00	
Low Avg			30.85		
Low Error%				0.1	
Mid Ref Cyl	4.9400	132.00	24.600	52.60	22.200
Mid Avg 4.9706	133.21	24.683	52.73	22.126	
Mid Error%	0.3	0.4	0.2	0.1	0.2
High Ref Cyl	9.4900	297.00	53.700	88.40	47.300
High Avg	9.4999	295.48	53.865	88.42	47.519
High Error%	0.1	0.5	0.3	0.0	0.5

IBIAS02

Initial System Bias Check for Run 2 . STRATA Version 2.0

SO2		O2	CO	NOx	THC	SO2	O2	CO	NOx	THC
volts	volts	%	ppm	ppm	ppm	ppm	volts	volts	volts	
11-18-2009		09:27:28		8.3558	110.32	0.528	18.22	-0.430	8.3558	
1.1032	0.0528	1.822	-0.0430							
11-18-2009		09:28:28		9.1650	11.84	0.670	3.44	-0.514	9.1650	
0.1184	0.0670	0.344	-0.0514							
11-18-2009		09:29:28		-0.8743	12.00	0.764	4.18	0.103	-0.8743	
0.1200	0.0764	0.418	0.0103							
11-18-2009		09:30:28		-0.0068	11.04	0.757	0.24	0.253	-0.0068	
0.1104	0.0757	0.024	0.0253							
11-18-2009		09:31:28		0.0612	12.01	0.663	0.21	0.169	0.0612	
0.1201	0.0663	0.021	0.0169							
11-18-2009		09:32:28		3.8197	12.60	0.681	0.37	0.003	3.8197	
0.1260	0.0681	0.037	0.0003							
11-18-2009		09:33:28		4.9630	12.69	0.626	51.06	-0.248	4.9630	
0.1269	0.0626	5.106	-0.0248							
11-18-2009		09:34:29		5.1023	13.62	0.523	52.33	-0.132	5.1023	
0.1362	0.0523	5.233	-0.0132							
11-18-2009		09:35:28		3.5821	78.84	0.463	52.41	-0.199	3.5821	
0.7884	0.0463	5.241	-0.0199							
11-18-2009		09:36:29		0.9386	131.32	0.370	24.31	0.235	0.9386	
1.3132	0.0370	2.431	0.0235							
11-18-2009		09:37:28		1.1350	64.52	7.657	2.68	0.334	1.1350	
0.6452	0.7657	0.268	0.0334							
11-18-2009		09:38:28		1.0324	15.60	23.664	2.18	1.083	1.0324	
0.1560	2.3664	0.218	0.1083							
11-18-2009		09:39:28		1.0657	17.25	23.856	1.84	1.102	1.0657	
0.1725	2.3856	0.184	0.1102							
11-18-2009		09:40:28		2.1277	20.39	16.371	1.49	2.992	2.1277	
0.2039	1.6371	0.149	0.2992							
11-18-2009		09:41:28		4.2757	55.89	4.283	1.81	21.609	4.2757	
0.5589	0.4283	0.181	2.1609							

Initial System Bias Check for Run 2

Operator: JMM  
 Plant Name: Conoco  
 Location: Old Ocean, TX  
 Reference Cylinder Numbers  
 Zero Span  
 O2 N2 cc9824  
 CO N2 ALM014290  
 NOx N2 cc241846  
 THC AIR cc236179  
 SO2 N2 cc144163

Date/Time	11-18-2009	09:42:15	PASSED
Analyte O2	CO	NOx	THC
Units %	ppm	ppm	ppm
Zero Ref Cyl	0.0000	0.00	0.000
Zero Cal	0.1396	2.22	0.038
Zero Avg	0.0155	10.99	0.723
Zero Bias%	1.3	3.0	1.3
Zero Drift%			0.2
Span Ref Cyl	4.9400	132.00	24.600
Span Cal	4.9706	133.21	24.683
Span Avg	5.0865	130.98	23.864
Span Bias%	1.2	0.7	1.5
Span Drift%			0.4

Test Run 2 Begin. STRATA Version 2.0

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

SO2		O2	CO	NOX	THC	SO2	O2	CO	NOX	THC
Volts	Volts	%	ppm	ppm	ppm	ppm	Volts	Volts	Volts	
11-18-2009		09:43:16		5.6882	46.30	2.848	1.12	13.480	5.6882	
0.4630	0.2848	0.112	1.3480							
11-18-2009		09:44:16		8.0246*	71.63	5.202	0.87	0.142	8.0246*	
0.7163	0.5202	0.087	0.0142							
11-18-2009		09:45:15		3.0165	78.94	11.002	1.85	0.026	3.0165	
0.7894	1.1002	0.185	0.0026							
11-18-2009		09:46:16		2.9385	75.53	11.676	0.85	0.142	2.9385	
0.7553	1.1676	0.085	0.0142							
11-18-2009		09:47:15		2.9371	90.75	11.859	1.53	0.118	2.9371	
0.9075	1.1859	0.153	0.0118							
11-18-2009		09:48:16		2.8276	99.22	12.303	1.63	0.185	2.8276	
0.9922	1.2303	0.163	0.0185							
11-18-2009		09:49:15		2.7444	91.28	12.615	2.12	0.226	2.7444	
0.9128	1.2615	0.212	0.0226							
11-18-2009		09:50:16		2.7794	71.62	12.862	1.22	0.206	2.7794	
0.7162	1.2862	0.122	0.0206							
11-18-2009		09:51:15		2.8980	67.30	12.796	0.97	0.161	2.8980	
0.6730	1.2796	0.097	0.0161							
11-18-2009		09:52:16		2.8810	59.48	13.014	0.62	0.269	2.8810	
0.5948	1.3014	0.062	0.0269							
11-18-2009		09:53:16		2.8600	75.32	12.862	0.91	0.200	2.8600	
0.7532	1.2862	0.091	0.0200							
11-18-2009		09:54:15		2.9329	84.52	12.953	0.72	0.289	2.9329	
0.8452	1.2953	0.072	0.0289							
11-18-2009		09:55:16		2.8108	90.69	13.233	1.39	0.303	2.8108	
0.9069	1.3233	0.139	0.0303							
11-18-2009		09:56:15		2.7636	82.59	13.493	1.77	0.498	2.7636	
0.8259	1.3493	0.177	0.0498							
11-18-2009		09:57:16		2.8059	96.29	13.569	1.18	0.349	2.8059	
0.9629	1.3569	0.118	0.0349							
11-18-2009		09:58:15		2.8290	88.08	13.672	1.99	0.484	2.8290	
0.8808	1.3672	0.199	0.0484							
11-18-2009		09:59:16		2.8220	75.68	13.912	1.82	0.267	2.8220	
0.7568	1.3912	0.182	0.0267							
11-18-2009		10:00:15		2.9117	71.01	13.776	1.00	0.495	2.9117	
0.7101	1.3776	0.100	0.0495							
11-18-2009		10:01:16		2.8201	69.70	13.971	0.85	0.402	2.8201	
0.6970	1.3971	0.085	0.0402							
11-18-2009		10:02:16		2.7846	60.42	14.094	1.03	0.369	2.7846	
0.6042	1.4094	0.103	0.0369							
11-18-2009		10:03:15		2.8349	82.36	13.837	0.90	0.439	2.8349	
0.8236	1.3837	0.090	0.0439							
11-18-2009		10:04:16		2.7533	91.48	13.840	1.96	0.611	2.7533	
0.9148	1.3840	0.196	0.0611							
11-18-2009		10:05:15		2.7210	91.59	13.815	1.67	0.531	2.7210	
0.9159	1.3815	0.167	0.0531							
11-18-2009		10:06:16		2.7213	79.16	13.721	2.21	0.663	2.7213	
0.7916	1.3721	0.221	0.0663							
11-18-2009		10:07:15		2.7804	69.60	13.680	1.03	0.524	2.7804	
0.6960	1.3680	0.103	0.0524							
11-18-2009		10:08:16		2.8644	82.32	13.374	1.20	0.529	2.8644	
0.8232	1.3374	0.120	0.0529							
11-18-2009		10:09:15		2.8558	66.39	13.327	1.04	0.452	2.8558	
0.6639	1.3327	0.104	0.0452							
11-18-2009		10:10:16		2.8514	84.42	13.374	0.85	0.464	2.8514	

				RUN02			
0.8442	1.3374	0.085	0.0464				
11-18-2009		10:11:16		2.8422	79.77	13.261	2.00
0.7977	1.3261	0.200	0.0350			0.350	2.8422
11-18-2009		10:12:15		2.8378	83.13	13.202	1.16
0.8313	1.3202	0.116	0.0466			0.466	2.8378
11-18-2009		10:13:16		2.8952	91.99	13.162	1.19
0.9199	1.3162	0.119	0.0485			0.485	2.8952
11-18-2009		10:14:15		2.9278	80.76	13.209	1.77
0.8076	1.3209	0.177	0.0483			0.483	2.9278
11-18-2009		10:15:16		2.8996	99.35	12.993	1.50
0.9935	1.2993	0.150	0.0514			0.514	2.8996
11-18-2009		10:16:15		2.8954	78.25	13.162	1.69
0.7825	1.3162	0.169	0.0483			0.483	2.8954
11-18-2009		10:17:16		2.9535	97.13	13.133	1.05
0.9713	1.3133	0.105	0.0456			0.456	2.9535
11-18-2009		10:18:15		2.9420	98.48	13.094	1.87
0.9848	1.3094	0.187	0.0482			0.482	2.9420
11-18-2009		10:19:16		2.9768	87.46	13.018	1.02
0.8746	1.3018	0.102	0.0385			0.385	2.9768
11-18-2009		10:20:15		2.9517	71.06	13.040	1.38
0.7106	1.3040	0.138	0.0465			0.465	2.9517
11-18-2009		10:21:15		2.9356	81.23	13.052	0.96
0.8123	1.3052	0.096	0.0530			0.530	2.9356
11-18-2009		10:22:16		2.9591	92.87	13.124	1.52
0.9287	1.3124	0.152	0.0515			0.515	2.9591
11-18-2009		10:23:15		2.9210	90.46	13.034	1.73
0.9046	1.3034	0.173	0.0581			0.581	2.9210
11-18-2009		10:24:16		2.9862	83.11	13.033	1.34
0.8311	1.3033	0.134	0.0415			0.415	2.9862
11-18-2009		10:25:15		2.9937	89.67	12.908	1.48
0.8967	1.2908	0.148	0.0649			0.649	2.9937
11-18-2009		10:26:16		2.9773	90.75	12.993	1.24
0.9075	1.2993	0.124	0.0567			0.567	2.9773
11-18-2009		10:27:15		3.0596	103.84	12.891	1.21
1.0384	1.2891	0.121	0.0614			0.614	3.0596
11-18-2009		10:28:16		3.0895	81.98	13.041	1.28
0.8198	1.3041	0.128	0.0597			0.597	3.0895
11-18-2009		10:29:16		3.0498	80.61	12.951	0.64
0.8061	1.2951	0.064	0.0613			0.613	3.0498
11-18-2009		10:30:15		3.0717	110.98	13.083	1.09
1.1098	1.3083	0.109	0.0702			0.702	3.0717
11-18-2009		10:31:16		2.8976	101.75	13.295	2.08
1.0175	1.3295	0.208	0.0681			0.681	2.8976
11-18-2009		10:32:15		2.8794	116.77	13.460	2.03
1.1677	1.3460	0.203	0.0717			0.717	2.8794
11-18-2009		10:33:16		2.8806	109.92	13.396	3.11
1.0992	1.3396	0.311	0.0584			0.584	2.8806
11-18-2009		10:34:15		2.9690	82.89	13.356	1.61
0.8289	1.3356	0.161	0.0629			0.629	2.9690
11-18-2009		10:35:16		3.0761	73.90	13.144	1.29
0.7390	1.3144	0.129	0.0563			0.563	3.0761
11-18-2009		10:36:15		3.1194	66.94	13.285	0.54
0.6694	1.3285	0.054	0.0581			0.581	3.1194
11-18-2009		10:37:16		3.0851	73.05	13.081	0.92
0.7305	1.3081	0.092	0.0627			0.627	3.0851
11-18-2009		10:38:15		3.0981	73.20	13.214	0.67
0.7320	1.3214	0.067	0.0684			0.684	3.0981
11-18-2009		10:39:16		3.0685	76.95	13.261	0.63
0.7695	1.3261	0.063	0.0547			0.547	3.0685
11-18-2009		10:40:16		3.0244	85.84	13.141	1.23
0.8584	1.3141	0.123	0.0596			0.596	3.0244
11-18-2009		10:41:15		3.0556	88.28	13.051	1.44
0.8828	1.3051	0.144	0.0594			0.594	3.0556



11-18-2009	10:42:16	3.0006	RUN02 93.04	12.824	1.36	0.604	3.0006
0.9304 1.2824	0.136 0.0604						
11-18-2009	10:43:15	3.0305	102.02	12.753	1.01	0.597	3.0305
1.0202 1.2753	0.101 0.0597						
11-18-2009	10:44:16	3.0232	91.20	12.813	1.18	0.608	3.0232
0.9120 1.2813	0.118 0.0608						
11-18-2009	10:45:15	3.0562	97.79	12.832	1.25	0.690	3.0562
0.9779 1.2832	0.125 0.0690						
11-18-2009	10:46:16	3.0669	81.07	12.765	1.04	0.803	3.0669
0.8107 1.2765	0.104 0.0803						
11-18-2009	10:47:15	3.0300	99.77	12.633	1.08	0.595	3.0300
0.9977 1.2633	0.108 0.0595						
11-18-2009	10:48:15	3.1059	97.72	12.715	1.08	0.621	3.1059
0.9772 1.2715	0.108 0.0621						
11-18-2009	10:49:16	3.0711	124.79	12.609	1.68	0.613	3.0711
1.2479 1.2609	0.168 0.0613						
11-18-2009	10:50:15	3.0866	108.15	12.449	1.31	0.668	3.0866
1.0815 1.2449	0.131 0.0668						
11-18-2009	10:51:16	3.0802	108.09	12.536	1.41	0.731	3.0802
1.0809 1.2536	0.141 0.0731						
11-18-2009	10:52:15	3.0740	104.61	12.448	1.35	0.640	3.0740
1.0461 1.2448	0.135 0.0640						
11-18-2009	10:53:16	3.0919	90.55	12.614	1.35	0.598	3.0919
0.9055 1.2614	0.135 0.0598						
11-18-2009	10:54:15	3.1501	98.03	12.593	0.98	0.580	3.1501
0.9803 1.2593	0.098 0.0580						
11-18-2009	10:55:16	3.1556	99.96	12.564	1.02	0.577	3.1556
0.9996 1.2564	0.102 0.0577						
11-18-2009	10:56:15	3.1111	103.33	12.716	1.29	0.559	3.1111
1.0333 1.2716	0.129 0.0559						
11-18-2009	10:57:16	3.1431	92.35	12.841	1.00	0.763	3.1431
0.9235 1.2841	0.100 0.0763						
11-18-2009	10:58:16	3.3006	103.36	12.862	0.76	0.713	3.3006
1.0336 1.2862	0.076 0.0713						
11-18-2009	10:59:15	3.2818	94.51	12.814	0.82	0.804	3.2818
0.9451 1.2814	0.082 0.0804						
11-18-2009	11:00:16	3.3088	96.15	12.893	0.82	0.806	3.3088
0.9615 1.2893	0.082 0.0806						
11-18-2009	11:01:15	3.2923	90.08	13.125	0.88	0.747	3.2923
0.9008 1.3125	0.088 0.0747						
11-18-2009	11:02:16	3.2847	75.59	13.308	0.58	0.705	3.2847
0.7559 1.3308	0.058 0.0705						
11-18-2009	11:03:15	3.2446	66.24	13.309	0.63	0.684	3.2446
0.6624 1.3309	0.063 0.0684						
11-18-2009	11:04:16	3.2195	71.93	13.523	0.68	0.754	3.2195
0.7193 1.3523	0.068 0.0754						
11-18-2009	11:05:15	3.0668	69.38	13.748	0.91	0.705	3.0668
0.6938 1.3748	0.091 0.0705						
11-18-2009	11:06:16	3.0491	67.16	13.860	0.99	0.772	3.0491
0.6716 1.3860	0.099 0.0772						
11-18-2009	11:07:16	3.0481	49.74	14.111	0.62	0.736	3.0481
0.4974 1.4111	0.062 0.0736						
11-18-2009	11:08:15	3.1593	51.25	13.741	0.40	0.813	3.1593
0.5125 1.3741	0.040 0.0813						
11-18-2009	11:09:16	3.1685	56.26	13.611	0.33	0.666	3.1685
0.5626 1.3611	0.033 0.0666						
11-18-2009	11:10:15	3.1945	51.09	13.689	0.41	0.843	3.1945
0.5109 1.3689	0.041 0.0843						
11-18-2009	11:11:16	3.1805	68.14	13.416	0.54	0.584	3.1805
0.6814 1.3416	0.054 0.0584						
11-18-2009	11:12:15	3.2806	63.78	13.363	0.89	0.703	3.2806
0.6378 1.3363	0.089 0.0703						
11-18-2009	11:13:16	3.2463	69.60	13.188	0.42	0.664	3.2463

RUN02

0.6960	1.3188	0.042	0.0664						
11-18-2009		11:14:15		3.2435	64.31	13.171	0.55	0.611	3.2435
0.6431	1.3171	0.055	0.0611						
11-18-2009		11:15:16		3.2857	82.99	13.192	0.49	0.724	3.2857
0.8299	1.3192	0.049	0.0724						
Begin calculating run averages									
11-18-2009		11:16:52		3.1610	63.63	13.581	0.96	0.824	3.1610
0.6363	1.3581	0.096	0.0824						
11-18-2009		11:17:53		3.2261	64.71	13.369	0.59	0.831	3.2261
0.6471	1.3369	0.059	0.0831						
11-18-2009		11:18:52		3.2242	61.91	13.492	0.49	0.796	3.2242
0.6191	1.3492	0.049	0.0796						
11-18-2009		11:19:53		3.2092	67.01	13.486	0.48	0.666	3.2092
0.6701	1.3486	0.048	0.0666						
11-18-2009		11:20:52		3.2466	63.22	13.642	0.56	0.580	3.2466
0.6322	1.3642	0.056	0.0580						
11-18-2009		11:21:53		3.2239	57.45	13.499	0.55	0.646	3.2239
0.5745	1.3499	0.055	0.0646						
11-18-2009		11:22:52		3.2497	53.00	13.508	0.33	0.812	3.2497
0.5300	1.3508	0.033	0.0812						
11-18-2009		11:23:52		3.1591	70.30	13.579	0.51	0.611	3.1591
0.7030	1.3579	0.051	0.0611						
11-18-2009		11:24:52		3.1170	68.62	13.812	0.93	0.646	3.1170
0.6862	1.3812	0.093	0.0646						
11-18-2009		11:25:52		3.0327	65.99	13.949	0.99	0.545	3.0327
0.6599	1.3949	0.099	0.0545						
11-18-2009		11:26:53		3.1023	69.51	13.778	0.89	0.711	3.1023
0.6951	1.3778	0.089	0.0711						
11-18-2009		11:27:52		3.1805	47.16	13.588	0.78	0.713	3.1805
0.4716	1.3588	0.078	0.0713						
11-18-2009		11:28:53		3.1545	52.70	13.362	0.42	0.727	3.1545
0.5270	1.3362	0.042	0.0727						
11-18-2009		11:29:52		3.1253	50.79	13.589	0.31	0.729	3.1253
0.5079	1.3589	0.031	0.0729						
11-18-2009		11:30:53		3.0684	59.69	13.475	0.57	0.683	3.0684
0.5969	1.3475	0.057	0.0683						
11-18-2009		11:31:52		3.0840	73.32	13.300	1.12	0.788	3.0840
0.7332	1.3300	0.112	0.0788						
11-18-2009		11:32:53		3.1132	63.20	13.403	0.66	0.901	3.1132
0.6320	1.3403	0.066	0.0901						
11-18-2009		11:33:52		3.1516	64.00	13.361	0.49	0.747	3.1516
0.6400	1.3361	0.049	0.0747						
11-18-2009		11:34:52		3.1107	77.42	13.297	0.85	0.751	3.1107
0.7742	1.3297	0.085	0.0751						
11-18-2009		11:35:53		3.1301	83.90	13.406	1.49	0.748	3.1301
0.8390	1.3406	0.149	0.0748						
11-18-2009		11:36:52		2.9835	84.11	13.600	1.52	0.645	2.9835
0.8411	1.3600	0.152	0.0645						
11-18-2009		11:37:53		2.9748	89.60	13.637	1.52	0.633	2.9748
0.8960	1.3637	0.152	0.0633						
11-18-2009		11:38:52		2.9699	84.12	13.754	1.63	0.777	2.9699
0.8412	1.3754	0.163	0.0777						
11-18-2009		11:39:53		3.0575	74.69	13.880	1.11	0.680	3.0575
0.7469	1.3880	0.111	0.0680						
11-18-2009		11:40:52		3.0147	82.16	13.803	1.33	0.731	3.0147
0.8216	1.3803	0.133	0.0731						
11-18-2009		11:41:53		3.0488	71.52	13.929	1.42	0.561	3.0488
0.7152	1.3929	0.142	0.0561						
11-18-2009		11:42:52		2.9865	77.31	14.018	1.39	0.649	2.9865
0.7731	1.4018	0.139	0.0649						
11-18-2009		11:43:52		2.9450	73.12	13.946	1.41	0.798	2.9450
0.7312	1.3946	0.141	0.0798						
11-18-2009		11:44:53		3.0176	71.65	14.016	0.74	0.748	3.0176

RUN02

0.7165	1.4016	0.074	0.0748						
11-18-2009		11:45:52		2.9903	68.38	13.934	1.34	0.774	2.9903
0.6838	1.3934	0.134	0.0774						
11-18-2009		11:46:53		2.9673	77.18	14.001	0.96	0.709	2.9673
0.7718	1.4001	0.096	0.0709						
11-18-2009		11:47:52		2.9888	71.39	14.034	1.48	0.646	2.9888
0.7139	1.4034	0.148	0.0646						
11-18-2009		11:48:53		2.9710	70.44	14.017	0.99	0.636	2.9710
0.7044	1.4017	0.099	0.0636						
11-18-2009		11:49:52		2.9999	70.98	14.000	0.97	0.662	2.9999
0.7098	1.4000	0.097	0.0662						
11-18-2009		11:50:53		2.9870	57.35	14.150	0.83	0.665	2.9870
0.5735	1.4150	0.083	0.0665						
11-18-2009		11:51:52		3.0213	58.90	14.044	0.96	0.749	3.0213
0.5890	1.4044	0.096	0.0749						
11-18-2009		11:52:52		3.0774	68.09	13.856	0.62	0.718	3.0774
0.6809	1.3856	0.062	0.0718						
11-18-2009		11:53:52		3.0482	58.67	14.023	1.27	0.812	3.0482
0.5867	1.4023	0.127	0.0812						
11-18-2009		11:54:52		3.0448	46.92	14.000	0.49	0.713	3.0448
0.4692	1.4000	0.049	0.0713						
11-18-2009		11:55:53		3.0853	68.78	13.719	0.64	0.686	3.0853
0.6878	1.3719	0.064	0.0686						
11-18-2009		11:56:52		3.1326	57.47	13.736	0.60	0.619	3.1326
0.5747	1.3736	0.060	0.0619						
11-18-2009		11:57:53		3.1214	61.59	13.597	0.45	0.692	3.1214
0.6159	1.3597	0.045	0.0692						
11-18-2009		11:58:52		3.1005	68.89	13.620	0.54	0.678	3.1005
0.6889	1.3620	0.054	0.0678						
11-18-2009		11:59:53		3.1154	64.50	13.566	0.85	0.721	3.1154
0.6450	1.3566	0.085	0.0721						
11-18-2009		12:00:52		3.1186	70.77	13.492	0.56	0.660	3.1186
0.7077	1.3492	0.056	0.0660						
11-18-2009		12:01:52		3.0293	72.72	13.812	0.99	0.770	3.0293
0.7272	1.3812	0.099	0.0770						
11-18-2009		12:02:52		2.9453	70.46	13.728	0.94	0.707	2.9453
0.7046	1.3728	0.094	0.0707						
11-18-2009		12:03:52		3.0710	64.37	13.771	1.43	0.644	3.0710
0.6437	1.3771	0.143	0.0644						
11-18-2009		12:04:53		3.0135	52.67	13.769	0.75	0.521	3.0135
0.5267	1.3769	0.075	0.0521						
11-18-2009		12:05:52		3.0961	73.31	13.586	0.60	0.691	3.0961
0.7331	1.3586	0.060	0.0691						
11-18-2009		12:06:53		3.0948	67.61	13.596	0.60	0.683	3.0948
0.6761	1.3596	0.060	0.0683						
11-18-2009		12:07:52		3.0199	72.66	13.579	0.90	0.473	3.0199
0.7266	1.3579	0.090	0.0473						
11-18-2009		12:08:53		3.0603	73.03	13.568	0.94	0.717	3.0603
0.7303	1.3568	0.094	0.0717						
11-18-2009		12:09:52		3.1469	55.82	13.491	0.58	0.580	3.1469
0.5582	1.3491	0.058	0.0580						
11-18-2009		12:10:53		3.1398	71.85	13.442	0.50	0.642	3.1398
0.7185	1.3442	0.050	0.0642						
11-18-2009		12:11:52		3.0718	71.87	13.532	0.90	0.575	3.0718
0.7187	1.3532	0.090	0.0575						
11-18-2009		12:12:52		3.1567	72.28	13.415	0.72	0.614	3.1567
0.7228	1.3415	0.072	0.0614						
11-18-2009		12:13:53		3.1571	76.81	13.593	0.97	0.566	3.1571
0.7681	1.3593	0.097	0.0566						
11-18-2009		12:14:52		3.0604	59.98	13.753	0.75	0.487	3.0604
0.5998	1.3753	0.075	0.0487						
11-18-2009		12:15:53		3.1332	77.26	13.458	0.57	0.550	3.1332
0.7726	1.3458	0.057	0.0550						

Average of Test Run				O2	RUN02 CO	NOx	THC	SO2	O2	CO
NOx	THC	SO2		ppm	ppm	ppm	%	ppm	ppm	ppm
ppm		%	ppm							
11-18-2009		12:15:53		3.0839	67.65	13.682	0.86	0.684	3.0839	
0.6765	1.3682	0.086	0.0684							
Test Run 2 End										

FBIAS02  
STRATA Version 2.0

Final System Bias Check for Run 2

SO2	O2	CO	NOx	THC	SO2	O2	CO	NOx	THC
Volts	Volts	%	ppm	ppm	ppm	Volts	Volts	Volts	
11-18-2009	12:18:11		1.0384	14.22	0.603	8.26	0.830	1.0384	
0.1422 0.0603	0.826 0.0830		0.8857	14.14	0.415	0.09	0.849	0.8857	
11-18-2009	12:19:11		0.1466	15.66	0.353	0.02	0.912	0.1466	
0.1414 0.0415	0.009 0.0849		0.1640	16.42	0.265	0.00	0.802	0.1640	
11-18-2009	12:20:11		0.0624	15.78	0.261	-0.01	0.464	0.0624	
0.1566 0.0353	0.002 0.0912		2.7803	15.91	0.275	-0.02	0.261	2.7803	
11-18-2009	12:21:12		0.1591 0.0275	-0.002 0.0261					
0.1642 0.0265	0.000 0.0802		4.8965	13.57	0.258	5.98	-0.068	4.8965	
11-18-2009	12:22:11		0.1357 0.0258	0.598 -0.0068					
0.1578 0.0261	-0.001 0.0464		4.9517	13.62	0.260	51.50	-0.137	4.9517	
11-18-2009	12:23:12		0.1362 0.0260	5.150 -0.0137					
0.1793 0.0261	5.173 -0.0090		5.0548	17.93	0.261	51.73	-0.090	5.0548	
11-18-2009	12:24:11		0.1793 0.0261	5.173 -0.0090					
1.2965 0.0257	5.165 0.0153		0.4445	129.65	0.257	51.65	0.153	0.4445	
11-18-2009	12:25:12		1.2965 0.0257	5.165 0.0153					
1.3425 0.0257	0.370 0.0490		-0.2007	134.25	0.257	3.70	0.490	-0.2007	
11-18-2009	12:26:11		1.3425 0.0257	0.370 0.0490					
0.4921 0.8178	0.189 0.0490		0.2318	49.21	8.178	1.89	0.490	0.2318	
11-18-2009	12:27:12		0.4921 0.8178	0.189 0.0490					
0.1383 2.4541	0.107 0.1439		-0.0591	13.83	24.541	1.07	1.439	-0.0591	
11-18-2009	12:28:11		0.1383 2.4541	0.107 0.1439					
0.1359 2.4639	0.093 0.1501		0.1922	13.59	24.639	0.93	1.501	0.1922	
11-18-2009	12:29:11		0.1359 2.4639	0.093 0.1501					
0.4301 0.6463	0.122 1.4719		2.8179	43.01	6.463	1.22	14.719	2.8179	
11-18-2009	12:30:11		0.4301 0.6463	0.122 1.4719					
0.6795 0.3919	0.148 2.2640		4.7786	67.95	3.919	1.48	22.640	4.7786	
11-18-2009	12:31:11		0.6795 0.3919	0.148 2.2640					

Final System Bias Check for Run 2

Operator: JMM  
Plant Name: Conoco  
Location: Old Ocean, TX

Reference Cylinder Numbers

O2	N2	cc9824
CO	N2	ALM014290
NOx	N2	cc241846
THC	AIR	cc236179
SO2	N2	cc144163

Date/Time	11-18-2009	12:33:17	PASSED
Analyte O2	CO	NOx	THC
Units %	ppm	ppm	ppm
Zero Ref Cyl	0.0000	0.00	0.000
Zero Cal	0.1396	2.22	0.038
Zero Avg	0.1593	16.14	0.313
Zero Bias%	0.2	4.7	0.5
Zero Drift%	1.5	1.7	-0.8
Span Ref Cyl	4.9400	132.00	24.600
Span Cal	4.9706	133.21	24.683
Span Avg	4.9520	133.58	24.543
Span Bias%	0.2	0.1	0.3
Span Drift%	-1.4	0.9	1.3

FBIAS02

Ini Zero Avg	0.0155	10.99	0.723	0.23	0.244
Ini Span Avg	5.0865	130.98	23.864	52.29	22.471
Run Avg 3.0839	67.65	13.682	0.86	0.684	
Co 0.0874	13.57	0.518	0.12	0.558	
Cm 5.0192	132.28	24.204	51.94	22.562	
Correct Avg	3.0015	60.14	13.672	0.75	0.127

Test Run 3 Begin. STRATA Version 2.0

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

SO2		O2	CO	NOx	THC	SO2	O2	CO	NOx	THC
Volts	Volts	%	ppm	ppm	ppm	ppm	Volts	Volts	Volts	
11-18-2009		12:34:17		4.8856	54.51	3.229	1.51	12.157	4.8856	
0.5451	0.3229	0.151	1.2157							
11-18-2009		12:35:18		8.0911*	49.02	4.428	1.65	-0.153	8.0911*	
0.4902	0.4428	0.165	-0.0153							
11-18-2009		12:36:17		2.2050	79.76	12.594	0.97	0.234	2.2050	
0.7976	1.2594	0.097	0.0234							
11-18-2009		12:37:18		2.2023	82.59	12.826	0.92	0.834	2.2023	
0.8259	1.2826	0.092	0.0834							
11-18-2009		12:38:17		2.1398	90.14	12.823	0.99	1.242	2.1398	
0.9014	1.2823	0.099	0.1242							
11-18-2009		12:39:18		2.1163	112.32	13.039	1.45	1.218	2.1163	
1.1232	1.3039	0.145	0.1218							
11-18-2009		12:40:17		1.9646	103.03	13.226	2.30	1.144	1.9646	
1.0303	1.3226	0.230	0.1144							
11-18-2009		12:41:17		1.9850	101.55	13.238	1.35	1.036	1.9850	
1.0155	1.3238	0.135	0.1036							
11-18-2009		12:42:17		1.9749	106.05	13.374	1.90	1.012	1.9749	
1.0605	1.3374	0.190	0.1012							
11-18-2009		12:43:17		1.9539	99.51	13.471	1.96	1.023	1.9539	
0.9951	1.3471	0.196	0.1023							
11-18-2009		12:44:18		1.9502	90.84	13.473	1.39	0.962	1.9502	
0.9084	1.3473	0.139	0.0962							
11-18-2009		12:45:17		1.9608	70.92	13.731	1.03	0.748	1.9608	
0.7092	1.3731	0.103	0.0748							
11-18-2009		12:46:18		1.9574	68.60	13.656	0.62	0.958	1.9574	
0.6860	1.3656	0.062	0.0958							
11-18-2009		12:47:17		2.0858	85.00	13.299	1.03	0.844	2.0858	
0.8500	1.3299	0.103	0.0844							
11-18-2009		12:48:18		2.1081	70.48	13.428	0.76	0.912	2.1081	
0.7048	1.3428	0.076	0.0912							
11-18-2009		12:49:17		2.0854	76.60	13.410	1.04	0.797	2.0854	
0.7660	1.3410	0.104	0.0797							
11-18-2009		12:50:17		2.0536	81.82	13.382	0.70	0.842	2.0536	
0.8182	1.3382	0.070	0.0842							
11-18-2009		12:51:17		2.0308	82.71	13.270	1.16	0.797	2.0308	
0.8271	1.3270	0.116	0.0797							
11-18-2009		12:52:17		2.0646	87.51	13.364	1.00	0.779	2.0646	
0.8751	1.3364	0.100	0.0779							
11-18-2009		12:53:18		2.1277	97.40	13.284	1.63	0.743	2.1277	
0.9740	1.3284	0.163	0.0743							
11-18-2009		12:54:17		2.0198	100.48	13.507	1.27	0.844	2.0198	
1.0048	1.3507	0.127	0.0844							
11-18-2009		12:55:18		1.8375	115.15	13.714	2.24	0.794	1.8375	
1.1515	1.3714	0.224	0.0794							
11-18-2009		12:56:17		1.8646	104.46	13.785	2.16	0.876	1.8646	
1.0446	1.3785	0.216	0.0876							
11-18-2009		12:57:18		2.0743	80.05	13.979	1.77	0.878	2.0743	
0.8005	1.3979	0.177	0.0878							
11-18-2009		12:58:17		2.4202	51.97	13.792	0.95	0.811	2.4202	
0.5197	1.3792	0.095	0.0811							
11-18-2009		12:59:17		3.0699	98.31	13.644	1.69	0.729	3.0699	
0.9831	1.3644	0.169	0.0729							
11-18-2009		13:00:17		2.8037	96.65	13.573	2.32	0.747	2.8037	
0.9665	1.3573	0.232	0.0747							
11-18-2009		13:01:17		2.8057	76.59	13.612	1.11	0.826	2.8057	

RUN03

0.7659	1.3612	0.111	0.0826						
11-18-2009		13:02:18		2.8893	85.78	13.508	1.14	0.695	2.8893
0.8578	1.3508	0.114	0.0695						
11-18-2009		13:03:17		2.7887	123.84	13.392	1.76	0.777	2.7887
1.2384	1.3392	0.176	0.0777						
11-18-2009		13:04:18		2.7138	86.53	13.791	3.08	0.768	2.7138
0.8653	1.3791	0.308	0.0768						
11-18-2009		13:05:17		2.9104	87.53	13.703	1.63	0.744	2.9104
0.8753	1.3703	0.163	0.0744						
11-18-2009		13:06:18		3.0244	105.79	13.692	2.28	0.660	3.0244
1.0579	1.3692	0.228	0.0660						
11-18-2009		13:07:17		3.0275	103.13	13.626	1.67	0.760	3.0275
1.0313	1.3626	0.167	0.0760						
11-18-2009		13:08:17		2.9665	115.13	13.624	2.27	0.710	2.9665
1.1513	1.3624	0.227	0.0710						
11-18-2009		13:09:17		3.0265	113.15	13.701	2.36	0.677	3.0265
1.1315	1.3701	0.236	0.0677						
11-18-2009		13:10:17		2.9473	99.22	13.738	2.42	0.695	2.9473
0.9922	1.3738	0.242	0.0695						
11-18-2009		13:11:17		3.1021	81.11	13.525	1.55	0.628	3.1021
0.8111	1.3525	0.155	0.0628						
11-18-2009		13:12:17		3.1843	69.97	13.536	0.89	0.746	3.1843
0.6997	1.3536	0.089	0.0746						
11-18-2009		13:13:18		3.1729	88.42	13.335	0.99	0.777	3.1729
0.8842	1.3335	0.099	0.0777						
11-18-2009		13:14:17		3.1601	82.61	13.427	1.08	0.658	3.1601
0.8261	1.3427	0.108	0.0658						
11-18-2009		13:15:18		3.0853	110.06	13.315	1.38	0.788	3.0853
1.1006	1.3315	0.138	0.0788						
11-18-2009		13:16:17		3.0685	77.63	13.675	1.93	0.631	3.0685
0.7763	1.3675	0.193	0.0631						
11-18-2009		13:17:17		3.1045	80.65	13.447	1.12	0.725	3.1045
0.8065	1.3447	0.112	0.0725						
11-18-2009		13:18:17		3.1472	79.75	13.452	1.33	0.646	3.1472
0.7975	1.3452	0.133	0.0646						
11-18-2009		13:19:17		3.1379	90.64	13.537	1.39	0.759	3.1379
0.9064	1.3537	0.139	0.0759						
11-18-2009		13:20:18		3.0283	115.81	13.428	1.85	0.692	3.0283
1.1581	1.3428	0.185	0.0692						
11-18-2009		13:21:17		3.0366	109.65	13.657	2.42	0.812	3.0366
1.0965	1.3657	0.242	0.0812						
11-18-2009		13:22:18		2.9593	94.33	13.516	2.61	0.640	2.9593
0.9433	1.3516	0.261	0.0640						
11-18-2009		13:23:17		3.0855	79.21	13.542	1.60	0.712	3.0855
0.7921	1.3542	0.160	0.0712						
11-18-2009		13:24:18		3.0763	74.95	13.478	0.77	0.825	3.0763
0.7495	1.3478	0.077	0.0825						
11-18-2009		13:25:17		3.0513	88.17	13.402	1.46	0.727	3.0513
0.8817	1.3402	0.146	0.0727						
11-18-2009		13:26:17		3.1292	86.03	13.147	1.11	0.728	3.1292
0.8603	1.3147	0.111	0.0728						
11-18-2009		13:27:17		3.2397	97.73	13.256	0.89	0.778	3.2397
0.9773	1.3256	0.089	0.0778						
11-18-2009		13:28:17		3.1633	99.58	13.097	1.68	0.826	3.1633
0.9958	1.3097	0.168	0.0826						
11-18-2009		13:29:17		3.2347	93.67	13.269	1.41	0.676	3.2347
0.9367	1.3269	0.141	0.0676						
11-18-2009		13:30:17		3.2446	85.88	13.350	1.49	0.683	3.2446
0.8588	1.3350	0.149	0.0683						
11-18-2009		13:31:18		3.2206	84.15	13.553	0.55	0.831	3.2206
0.8415	1.3553	0.055	0.0831						
11-18-2009		13:32:17		3.0838	100.17	13.515	2.17	0.661	3.0838
1.0017	1.3515	0.217	0.0661						



11-18-2009	13:33:18	3.2358	RUN03 87.68	13.429	1.99	0.713	3.2358
0.8768 1.3429	0.199 0.0713						
11-18-2009	13:34:17	3.2811	74.84	13.520	1.08	0.882	3.2811
0.7484 1.3520	0.108 0.0882						
11-18-2009	13:35:18	3.2020	104.38	13.501	1.21	0.748	3.2020
1.0438 1.3501	0.121 0.0748						
11-18-2009	13:36:17	3.0734	99.31	13.685	2.18	0.663	3.0734
0.9931 1.3685	0.218 0.0663						
11-18-2009	13:37:17	3.0182	94.51	13.798	1.56	0.824	3.0182
0.9451 1.3798	0.156 0.0824						
11-18-2009	13:38:17	3.2207	70.49	13.909	1.62	0.693	3.2207
0.7049 1.3909	0.162 0.0693						
11-18-2009	13:39:17	3.1658	82.16	13.690	0.90	0.712	3.1658
0.8216 1.3690	0.090 0.0712						
11-18-2009	13:40:18	3.1811	69.00	13.697	1.35	0.710	3.1811
0.6900 1.3697	0.135 0.0710						
11-18-2009	13:41:17	3.2035	84.08	13.533	0.70	0.692	3.2035
0.8408 1.3533	0.070 0.0692						
11-18-2009	13:42:18	3.2484	79.25	13.631	1.18	0.632	3.2484
0.7925 1.3631	0.118 0.0632						
11-18-2009	13:43:17	3.1699	93.81	13.557	1.33	0.801	3.1699
0.9381 1.3557	0.133 0.0801						
11-18-2009	13:44:18	3.1842	86.97	13.496	1.71	0.678	3.1842
0.8697 1.3496	0.171 0.0678						
11-18-2009	13:45:17	3.2637	92.90	13.304	1.53	0.709	3.2637
0.9290 1.3304	0.153 0.0709						
11-18-2009	13:46:17	3.2123	80.15	13.314	1.90	0.798	3.2123
0.8015 1.3314	0.190 0.0798						
11-18-2009	13:47:17	3.1449	86.62	13.178	0.84	0.718	3.1449
0.8662 1.3178	0.084 0.0718						
11-18-2009	13:48:17	3.1263	74.73	13.293	1.60	0.660	3.1263
0.7473 1.3293	0.160 0.0660						
11-18-2009	13:49:18	3.0939	72.74	13.200	1.10	0.725	3.0939
0.7274 1.3200	0.110 0.0725						
11-18-2009	13:50:17	3.1709	85.17	13.160	1.13	0.774	3.1709
0.8517 1.3160	0.113 0.0774						
11-18-2009	13:51:18	3.1104	107.20	13.185	1.52	0.757	3.1104
1.0720 1.3185	0.152 0.0757						
11-18-2009	13:52:17	2.9699	121.17	13.197	2.15	0.897	2.9699
1.2117 1.3197	0.215 0.0897						
11-18-2009	13:53:18	2.9426	117.40	13.575	2.82	0.867	2.9426
1.1740 1.3575	0.282 0.0867						
11-18-2009	13:54:17	2.9418	111.83	13.514	2.47	0.700	2.9418
1.1183 1.3514	0.247 0.0700						
11-18-2009	13:55:17	3.0041	107.32	13.636	2.49	0.711	3.0041
1.0732 1.3636	0.249 0.0711						
11-18-2009	13:56:17	2.9270	101.56	13.710	2.81	0.733	2.9270
1.0156 1.3710	0.281 0.0733						
11-18-2009	13:57:17	2.9840	103.38	13.685	1.99	0.742	2.9840
1.0338 1.3685	0.199 0.0742						
11-18-2009	13:58:18	2.9823	103.97	13.830	2.20	0.798	2.9823
1.0397 1.3830	0.220 0.0798						
11-18-2009	13:59:17	2.9446	103.58	13.711	2.44	0.660	2.9446
1.0358 1.3711	0.244 0.0660						
11-18-2009	14:00:18	3.0083	96.01	13.767	2.11	0.698	3.0083
0.9601 1.3767	0.211 0.0698						
11-18-2009	14:01:17	2.9500	82.17	13.714	1.56	0.733	2.9500
0.8217 1.3714	0.156 0.0733						
11-18-2009	14:02:18	3.1466	80.38	13.522	0.88	0.807	3.1466
0.8038 1.3522	0.088 0.0807						
11-18-2009	14:03:17	3.2159	88.46	13.468	1.15	0.722	3.2159
0.8846 1.3468	0.115 0.0722						
11-18-2009	14:04:17	3.1647	93.16	13.395	1.45	0.720	3.1647

RUN03

0.9316	1.3395	0.145	0.0720						
11-18-2009		14:05:17		3.0562	72.19	13.645	1.57	0.753	3.0562
0.7219	1.3645	0.157	0.0753						
11-18-2009		14:06:17		3.1246	79.05	13.554	0.65	0.627	3.1246
0.7905	1.3554	0.065	0.0627						
11-18-2009		14:07:18		3.0937	70.28	13.678	1.35	0.786	3.0937
0.7028	1.3678	0.135	0.0786						
11-18-2009		14:08:17		3.0584	83.69	13.468	0.97	0.757	3.0584
0.8369	1.3468	0.097	0.0757						
11-18-2009		14:09:18		3.1086	96.58	13.482	1.51	0.816	3.1086
0.9658	1.3482	0.151	0.0816						
11-18-2009		14:10:17		3.0320	95.61	13.194	1.97	0.740	3.0320
0.9561	1.3194	0.197	0.0740						
11-18-2009		14:11:18		3.0849	101.79	13.416	1.47	0.756	3.0849
1.0179	1.3416	0.147	0.0756						
11-18-2009		14:12:17		3.0036	92.35	13.282	2.61	0.734	3.0036
0.9235	1.3282	0.261	0.0734						
11-18-2009		14:13:17		3.0597	103.20	13.154	1.20	0.844	3.0597
1.0320	1.3154	0.120	0.0844						
11-18-2009		14:14:17		3.0709	82.88	13.303	1.75	0.640	3.0709
0.8288	1.3303	0.175	0.0640						
11-18-2009		14:15:17		3.0443	100.37	13.036	1.28	0.768	3.0443
1.0037	1.3036	0.128	0.0768						
11-18-2009		14:16:18		3.1402	100.07	13.231	1.44	0.735	3.1402
1.0007	1.3231	0.144	0.0735						
11-18-2009		14:17:17		3.1395	99.76	13.233	1.85	0.853	3.1395
0.9976	1.3233	0.185	0.0853						
11-18-2009		14:18:18		3.1440	114.50	13.149	1.15	0.771	3.1440
1.1450	1.3149	0.115	0.0771						
11-18-2009		14:19:17		3.2479	116.83	13.233	1.50	0.763	3.2479
1.1683	1.3233	0.150	0.0763						
11-18-2009		14:20:18		3.2288	105.58	13.165	1.66	0.872	3.2288
1.0558	1.3165	0.166	0.0872						
11-18-2009		14:21:17		3.2085	106.82	13.198	1.01	0.678	3.2085
1.0682	1.3198	0.101	0.0678						
11-18-2009		14:22:18		3.2789	98.24	13.280	1.67	0.715	3.2789
0.9824	1.3280	0.167	0.0715						
11-18-2009		14:23:17		3.2752	121.06	13.299	0.68	0.774	3.2752
1.2106	1.3299	0.068	0.0774						
11-18-2009		14:24:17		3.1570	109.41	13.502	2.55	0.629	3.1570
1.0941	1.3502	0.255	0.0629						
11-18-2009		14:25:17		3.0845	114.65	13.714	2.14	0.587	3.0845
1.1465	1.3714	0.214	0.0587						
11-18-2009		14:26:17		3.0846	84.74	13.816	1.58	0.732	3.0846
0.8474	1.3816	0.158	0.0732						
11-18-2009		14:27:18		3.2474	88.05	13.550	1.18	0.512	3.2474
0.8805	1.3550	0.118	0.0512						
11-18-2009		14:28:17		3.2019	81.06	13.428	0.99	0.665	3.2019
0.8106	1.3428	0.099	0.0665						
11-18-2009		14:29:18		3.1919	77.95	13.678	1.26	0.698	3.1919
0.7795	1.3678	0.126	0.0698						
11-18-2009		14:30:17		3.1154	96.33	13.449	1.29	0.711	3.1154
0.9633	1.3449	0.129	0.0711						
11-18-2009		14:31:18		3.1428	111.49	13.546	2.33	0.709	3.1428
1.1149	1.3546	0.233	0.0709						
11-18-2009		14:32:17		2.8997	107.59	13.772	2.04	0.836	2.8997
1.0759	1.3772	0.204	0.0836						
11-18-2009		14:33:17		2.9691	103.94	13.592	1.59	0.847	2.9691
1.0394	1.3592	0.159	0.0847						
11-18-2009		14:34:17		3.1712	86.24	13.571	1.89	0.746	3.1712
0.8624	1.3571	0.189	0.0746						
11-18-2009		14:35:17		3.1367	105.58	13.519	1.60	0.818	3.1367
1.0558	1.3519	0.160	0.0818						

11-18-2009	14:36:18	3.0545	RUN03 93.48	13.575	1.57	0.774	3.0545
0.9348 1.3575	0.157 0.0774						
11-18-2009	14:37:17	3.1431	93.75	13.597	1.90	0.712	3.1431
0.9375 1.3597	0.190 0.0712						
11-18-2009	14:38:18	3.2208	95.35	13.339	1.55	0.779	3.2208
0.9535 1.3339	0.155 0.0779						
11-18-2009	14:39:17	3.2910	101.62	13.425	1.24	0.627	3.2910
1.0162 1.3425	0.124 0.0627						
11-18-2009	14:40:18	3.1949	94.63	13.565	1.70	0.819	3.1949
0.9463 1.3565	0.170 0.0819						
11-18-2009	14:41:17	3.1425	83.59	13.457	1.32	0.714	3.1425
0.8359 1.3457	0.132 0.0714						
11-18-2009	14:42:17	3.1917	82.42	13.506	0.65	0.733	3.1917
0.8242 1.3506	0.065 0.0733						
11-18-2009	14:43:17	3.2011	92.63	13.679	0.91	0.757	3.2011
0.9263 1.3679	0.091 0.0757						
11-18-2009	14:44:17	3.0192	104.59	13.625	2.15	0.669	3.0192
1.0459 1.3625	0.215 0.0669						
11-18-2009	14:45:18	2.9896	87.00	13.757	1.70	0.828	2.9896
0.8700 1.3757	0.170 0.0828						
11-18-2009	14:46:17	3.0567	69.21	13.838	1.49	0.728	3.0567
0.6921 1.3838	0.149 0.0728						
11-18-2009	14:47:18	3.1662	80.67	13.562	0.71	0.696	3.1662
0.8067 1.3562	0.071 0.0696						
11-18-2009	14:48:17	3.1317	100.49	13.712	1.33	0.839	3.1317
1.0049 1.3712	0.133 0.0839						
11-18-2009	14:49:18	2.8992	89.14	13.738	2.32	0.739	2.8992
0.8914 1.3738	0.232 0.0739						
11-18-2009	14:50:17	3.1226	97.90	13.498	1.21	0.754	3.1226
0.9790 1.3498	0.121 0.0754						
11-18-2009	14:51:17	3.1623	84.89	13.615	1.92	0.747	3.1623
0.8489 1.3615	0.192 0.0747						
11-18-2009	14:52:17	3.0404	104.13	13.675	1.61	0.847	3.0404
1.0413 1.3675	0.161 0.0847						
11-18-2009	14:53:17	2.9609	77.34	14.056	1.49	0.762	2.9609
0.7734 1.4056	0.149 0.0762						
11-18-2009	14:54:18	3.0497	75.81	13.882	1.25	0.877	3.0497
0.7581 1.3882	0.125 0.0877						
11-18-2009	14:55:17	3.0898	101.13	13.734	1.30	0.711	3.0898
1.0113 1.3734	0.130 0.0711						
11-18-2009	14:56:18	3.1455	96.00	13.832	1.92	0.766	3.1455
0.9600 1.3832	0.192 0.0766						
11-18-2009	14:57:17	2.9561	103.51	13.914	2.01	0.690	2.9561
1.0351 1.3914	0.201 0.0690						
11-18-2009	14:58:18	2.9312	94.02	14.009	2.12	0.770	2.9312
0.9402 1.4009	0.212 0.0770						
11-18-2009	14:59:17	3.0426	86.35	14.039	1.53	0.797	3.0426
0.8635 1.4039	0.153 0.0797						
11-18-2009	15:00:17	3.0361	99.48	13.937	1.73	0.665	3.0361
0.9948 1.3937	0.173 0.0665						
11-18-2009	15:01:17	2.9704	94.12	14.027	1.72	0.808	2.9704
0.9412 1.4027	0.172 0.0808						
11-18-2009	15:02:17	3.0069	76.70	14.246	2.07	0.894	3.0069
0.7670 1.4246	0.207 0.0894						
11-18-2009	15:03:18	3.1637	78.12	13.808	1.13	0.829	3.1637
0.7812 1.3808	0.113 0.0829						
11-18-2009	15:04:17	3.4456	65.56	14.015	0.80	0.903	3.4456
0.6556 1.4015	0.080 0.0903						
11-18-2009	15:05:18	5.5288*	10.98	4.213	0.81	0.435	5.5288*
0.1098 0.4213	0.081 0.0435						
11-18-2009	15:06:17	1.1413	7.74	0.385	1.33	0.818	1.1413
0.0774 0.0385	0.133 0.0818						
11-18-2009	15:07:18	4.3863	7.53	0.367	2.00	0.121	4.3863

				RUN03					
0.0753	0.0367	0.200	0.0121						
11-18-2009		15:08:17		0.1316	8.35	0.264	1.05	0.390	0.1316
0.0835	0.0264	0.105	0.0390						
11-18-2009		15:09:17		4.9511*	9.04	0.395	1.39	0.266	4.9511*
0.0904	0.0395	0.139	0.0266						
11-18-2009		15:10:17		5.4962*	102.12	10.042	2.29	0.010	5.4962*
1.0212	1.0042	0.229	0.0010						
11-18-2009		15:11:17		2.0551	96.87	13.721	1.93	0.294	2.0551
0.9687	1.3721	0.193	0.0294						
11-18-2009		15:12:17		1.8700	97.30	13.960	1.82	0.186	1.8700
0.9730	1.3960	0.182	0.0186						
11-18-2009		15:13:17		1.9226	103.34	13.899	1.97	0.212	1.9226
1.0334	1.3899	0.197	0.0212						
11-18-2009		15:14:18		1.9316	76.22	14.095	8.05	0.385	1.9316
0.7622	1.4095	0.805	0.0385						
Begin calculating run averages									
11-18-2009		15:16:03		1.9028	97.52	14.098	11.65	0.619	1.9028
0.9752	1.4098	1.165	0.0619						
11-18-2009		15:17:04		1.8770	77.90	14.224	2.40	0.719	1.8770
0.7790	1.4224	0.240	0.0719						
11-18-2009		15:18:03		1.9254	83.83	14.063	1.87	0.861	1.9254
0.8383	1.4063	0.187	0.0861						
11-18-2009		15:19:04		1.9920	71.65	13.835	2.03	0.797	1.9920
0.7165	1.3835	0.203	0.0797						
11-18-2009		15:20:03		1.9721	70.71	14.067	2.23	0.663	1.9721
0.7071	1.4067	0.223	0.0663						
11-18-2009		15:21:04		1.9737	82.01	13.849	1.96	0.744	1.9737
0.8201	1.3849	0.196	0.0744						
11-18-2009		15:22:03		1.9362	106.40	13.778	1.67	0.820	1.9362
1.0640	1.3778	0.167	0.0820						
11-18-2009		15:23:03		1.7774	108.30	14.056	2.00	0.778	1.7774
1.0830	1.4056	0.200	0.0778						
11-18-2009		15:24:03		1.8144	102.75	14.151	2.60	0.944	1.8144
1.0275	1.4151	0.260	0.0944						
11-18-2009		15:25:03		1.8349	102.82	14.048	2.24	0.826	1.8349
1.0282	1.4048	0.224	0.0826						
11-18-2009		15:26:04		1.8305	103.05	14.266	2.59	0.798	1.8305
1.0305	1.4266	0.259	0.0798						
11-18-2009		15:27:03		1.8470	100.91	14.096	2.86	0.854	1.8470
1.0091	1.4096	0.286	0.0854						
11-18-2009		15:28:04		1.8336	110.72	14.194	2.51	0.756	1.8336
1.1072	1.4194	0.251	0.0756						
11-18-2009		15:29:03		1.7700	105.49	14.395	2.74	0.849	1.7700
1.0549	1.4395	0.274	0.0849						
11-18-2009		15:30:04		1.7685	109.97	14.293	2.24	0.851	1.7685
1.0997	1.4293	0.224	0.0851						
11-18-2009		15:31:03		1.7862	106.02	14.296	2.69	0.853	1.7862
1.0602	1.4296	0.269	0.0853						
11-18-2009		15:32:03		1.8115	105.94	14.462	2.79	0.809	1.8115
1.0594	1.4462	0.279	0.0809						
11-18-2009		15:33:03		1.7798	104.12	14.379	2.26	0.881	1.7798
1.0412	1.4379	0.226	0.0881						
11-18-2009		15:34:03		1.8549	103.15	14.429	2.65	0.921	1.8549
1.0315	1.4429	0.265	0.0921						
11-18-2009		15:35:04		1.7941	110.80	14.493	2.62	0.838	1.7941
1.1080	1.4493	0.262	0.0838						
11-18-2009		15:36:03		1.8121	107.37	14.479	2.53	0.952	1.8121
1.0737	1.4479	0.253	0.0952						
11-18-2009		15:37:04		1.8013	108.78	14.349	2.99	0.872	1.8013
1.0878	1.4349	0.299	0.0872						
11-18-2009		15:38:03		1.8750	83.43	14.424	2.38	0.811	1.8750
0.8343	1.4424	0.238	0.0811						
11-18-2009		15:39:04		1.8090	80.43	14.365	1.98	0.821	1.8090

RUN03

0.8043	1.4365	0.198	0.0821						
11-18-2009		15:40:03		1.9473	86.33	14.174	1.51	0.811	1.9473
0.8633	1.4174	0.151	0.0811						
11-18-2009		15:41:04		2.0256	69.84	14.095	1.51	0.629	2.0256
0.6984	1.4095	0.151	0.0629						
11-18-2009		15:42:03		1.9779	72.77	14.029	0.98	0.610	1.9779
0.7277	1.4029	0.098	0.0610						
11-18-2009		15:43:03		1.9762	65.99	14.126	1.42	0.806	1.9762
0.6599	1.4126	0.142	0.0806						
11-18-2009		15:44:04		1.9599	80.56	14.064	0.82	0.719	1.9599
0.8056	1.4064	0.082	0.0719						
11-18-2009		15:45:03		1.8946	93.84	14.015	1.67	0.821	1.8946
0.9384	1.4015	0.167	0.0821						
11-18-2009		15:46:04		1.8833	69.91	14.755	2.09	0.674	1.8833
0.6991	1.4755	0.209	0.0674						
11-18-2009		15:47:03		1.9463	80.76	14.260	3.04	0.872	1.9463
0.8076	1.4260	0.304	0.0872						
11-18-2009		15:48:04		1.9791	79.41	13.776	1.84	0.839	1.9791
0.7941	1.3776	0.184	0.0839						
11-18-2009		15:49:03		2.0341	75.25	13.809	1.66	1.184	2.0341
0.7525	1.3809	0.166	0.1184						
11-18-2009		15:50:03		2.0055	74.33	13.816	1.67	1.346	2.0055
0.7433	1.3816	0.167	0.1346						
11-18-2009		15:51:03		2.0359	68.43	13.725	1.38	1.091	2.0359
0.6843	1.3725	0.138	0.1091						
11-18-2009		15:52:03		2.0559	77.04	13.619	1.01	1.079	2.0559
0.7704	1.3619	0.101	0.1079						
11-18-2009		15:53:04		1.9950	95.89	13.648	1.41	1.145	1.9950
0.9589	1.3648	0.141	0.1145						
11-18-2009		15:54:03		2.0772	70.84	13.694	2.06	0.823	2.0772
0.7084	1.3694	0.206	0.0823						
11-18-2009		15:55:04		2.1317	67.91	13.682	0.72	0.872	2.1317
0.6791	1.3682	0.072	0.0872						
11-18-2009		15:56:03		2.0930	68.10	13.652	0.90	0.971	2.0930
0.6810	1.3652	0.090	0.0971						
11-18-2009		15:57:04		2.0856	70.02	13.751	0.98	0.879	2.0856
0.7002	1.3751	0.098	0.0879						
11-18-2009		15:58:03		2.0618	78.35	13.638	1.09	0.740	2.0618
0.7835	1.3638	0.109	0.0740						
11-18-2009		15:59:04		1.9783	87.89	13.658	1.33	0.786	1.9783
0.8789	1.3658	0.133	0.0786						
11-18-2009		16:00:03		1.9805	88.52	13.861	1.53	0.913	1.9805
0.8852	1.3861	0.153	0.0913						
11-18-2009		16:01:03		1.9708	87.88	13.928	1.62	0.796	1.9708
0.8788	1.3928	0.162	0.0796						
11-18-2009		16:02:04		1.9506	82.51	13.935	1.83	0.705	1.9506
0.8251	1.3935	0.183	0.0705						
11-18-2009		16:03:03		2.0234	82.57	13.843	1.39	0.744	2.0234
0.8257	1.3843	0.139	0.0744						
11-18-2009		16:04:04		1.9604	82.05	13.994	2.10	0.867	1.9604
0.8205	1.3994	0.210	0.0867						
11-18-2009		16:05:03		1.9641	84.38	13.923	1.40	0.844	1.9641
0.8438	1.3923	0.140	0.0844						
11-18-2009		16:06:04		1.9556	83.10	13.939	1.59	0.819	1.9556
0.8310	1.3939	0.159	0.0819						
11-18-2009		16:07:03		1.9528	76.67	14.064	1.24	0.832	1.9528
0.7667	1.4064	0.124	0.0832						
11-18-2009		16:08:04		1.9067	75.54	13.987	1.50	0.823	1.9067
0.7554	1.3987	0.150	0.0823						
11-18-2009		16:09:03		2.0140	75.13	14.003	1.66	0.771	2.0140
0.7513	1.4003	0.166	0.0771						
11-18-2009		16:10:03		1.9327	81.47	13.827	1.07	0.747	1.9327
0.8147	1.3827	0.107	0.0747						

11-18-2009		16:11:03		1.9293		80.42		13.837		1.62		0.812		1.9293	
0.8042	1.3837	0.162	0.0812	1.9930	68.42	13.924	1.40	0.780	1.9930						
0.6842	1.3924	0.140	0.0780	1.9966	63.02	13.797	1.05	0.696	1.9966						
0.6302	1.3797	0.105	0.0696	2.0635	56.60	13.679	0.76	0.706	2.0635						
0.5660	1.3679	0.076	0.0706	2.0447	62.80	13.612	0.67	0.537	2.0447						
0.6280	1.3612	0.067	0.0537												
Average of Test Run		SO2		O2	CO	NOx	THC	SO2	O2	CO					
NOx	THC	%	ppm	ppm	ppm	ppm	%	ppm	ppm	ppm					
ppm															
11-18-2009		16:15:04		1.9360		85.14		14.021		1.97		0.829		1.9360	
0.8514	1.4021	0.197	0.0829												
Test Run 3 End															

A45

FBIAS03  
STRATA Version 2.0

Final System Bias Check for Run 3

SO2	O2	CO	NOX	THC	SO2	O2	CO	NOX	THC
Volts	Volts	%	ppm	ppm	ppm	Volts	Volts	Volts	
11-18-2009	16:16:10		1.9680	49.63	12.552	0.60	0.641	1.9680	
0.4963	1.2552	0.060	0.0641						
11-18-2009	16:17:10		0.0403	7.39	0.892	7.11	0.687	0.0403	
0.0739	0.0892	0.711	0.0687						
11-18-2009	16:18:10		0.0976	11.42	0.305	0.16	0.618	0.0976	
0.1142	0.0305	0.016	0.0618						
11-18-2009	16:19:10		0.1053	14.10	0.252	0.38	0.612	0.1053	
0.1410	0.0252	0.038	0.0612						
11-18-2009	16:20:09		3.6291	13.68	0.251	3.99	0.261	3.6291	
0.1368	0.0251	0.399	0.0261						
11-18-2009	16:21:10		4.9367	13.75	0.252	2.01	-0.373	4.9367	
0.1375	0.0252	0.201	-0.0373						
11-18-2009	16:22:09		4.9368	13.93	0.221	2.64	-0.240	4.9368	
0.1393	0.0221	0.264	-0.0240						
11-18-2009	16:23:10		4.9358	14.24	0.152	20.37	-0.363	4.9358	
0.1424	0.0152	2.037	-0.0363						
11-18-2009	16:24:10		4.9385	13.80	0.154	51.50	-0.283	4.9385	
0.1380	0.0154	5.150	-0.0283						
11-18-2009	16:25:10		4.5691	35.18	0.152	12.31	-0.401	4.5691	
0.3518	0.0152	1.231	-0.0401						
11-18-2009	16:26:10		-0.0017	132.19	0.148	2.97	0.155	-0.0017	
1.3219	0.0148	0.297	0.0155						
11-18-2009	16:27:10		0.0571	121.73	0.148	2.47	0.057	0.0571	
1.2173	0.0148	0.247	0.0057						
11-18-2009	16:28:10		0.0586	15.41	17.888	2.50	0.762	0.0586	
0.1541	1.7888	0.250	0.0762						
11-18-2009	16:29:10		0.0228	14.64	24.647	2.53	1.146	0.0228	
0.1464	2.4647	0.253	0.1146						
11-18-2009	16:30:10		0.7472	14.75	23.993	2.55	0.967	0.7472	
0.1475	2.3993	0.255	0.0967						
11-18-2009	16:31:09		9.9747	15.17	6.002	2.54	0.980	9.9747	
0.1517	0.6002	0.254	0.0980						
11-18-2009	16:32:10		12.9857	22.60	0.570	2.57	22.254	12.9857	
0.2260	0.0570	0.257	2.2254						
11-18-2009	16:33:09		13.1607	23.56	0.553	2.58	7.332	13.1607	
0.2356	0.0553	0.258	0.7332						
11-18-2009	16:34:10		13.1607	23.83	0.605	2.58	17.739	13.1607	
0.2383	0.0605	0.258	1.7739						

Final System Bias Check for Run 3

Operator: JMM  
Plant Name: Conoco  
Location: Old Ocean, TX  
Reference Cylinder Numbers  
Zero Span  
O2 N2 cc9824  
CO N2 ALM014290  
NOx N2 cc241846  
THC AIR cc236179  
SO2 N2 cc144163

Date/Time	11-18-2009	16:34:59	PASSED
Analyte O2	CO	NOX	THC
Units %	ppm	ppm	ppm
Zero Ref Cyl	0.0000	0.00	0.000
Zero Cal	0.1396	2.22	0.038
Zero Avg	0.1019	14.04	0.265
Zero Bias%	0.4	4.0	0.4

				FBIAS03	
Zero Drift%	-0.6	-0.7	-0.1	0.1	-0.6
Span Ref Cyl	4.9400	132.00	24.600	52.60	22.200
Span Cal	4.9706	133.21	24.683	52.73	22.126
Span Avg	4.9375	131.90	24.618	51.28	22.043
Span Bias%	0.3	0.4	0.1	1.4	0.2
Span Drift%	-0.2	-0.6	0.1	-0.3	-1.3
Ini Zero Avg	0.1593	16.14	0.313	0.01	0.872
Ini Span Avg	4.9520	133.58	24.543	51.60	22.653
Run Avg 1.9360	85.14	14.021	1.97	0.829	
Co 0.1306	15.09	0.289	0.07	0.736	
Cm 4.9447	132.74	24.581	51.44	22.348	
Correct Avg	1.8526	78.60	13.906	1.95	0.095



RUN02

Test Run 2 Begin. STRATA Version 2.0

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

SO2	O2	CO	NOx	THC	SO2	O2	CO	NOx	THC
Volts	Volts	%	ppm	ppm	ppm	Volts	Volts	Volts	
11-17-2009	16:26:15		3.1704	57.68	0.227	1.06	22.839	3.1704	
0.5768	0.0227	0.106	2.2839						
11-17-2009	16:27:16		4.8172	62.53	21.972	0.58	22.889	4.8172	
0.6253	2.1972	0.058	2.2889						
11-17-2009	16:28:15		8.7625	58.90	40.230	0.75	22.772	8.7625	
0.5890	4.0230	0.075	2.2772						
11-17-2009	16:29:16		11.0774	59.12	42.548	0.74	22.836	11.0774	
0.5912	4.2548	0.074	2.2836						
11-17-2009	16:30:15		12.2621	58.36	49.508	0.97	22.609	12.2621	
0.5836	4.9508	0.097	2.2609						
11-17-2009	16:31:15		12.7904	62.50	51.089	1.25	3.918	12.7904	
0.6250	5.1089	0.125	0.3918						

RUN04

Test Run 4 Begin. STRATA Version 2.0

Operator: JMM

Plant Name: Conoco

Location: Old Ocean, TX

SO2	O2	CO	NOx	THC	SO2	O2	CO	NOx	THC	
Volts	Volts	%	ppm	ppm	ppm	Volts	Volts	Volts		
11-18-2009	13.1607*	16:36:00	0.2381	0.0340	0.255	1.5404	23.81	0.340	2.55	15.404
11-18-2009	13.1607*	16:37:00	0.2050	0.0054	0.254	-0.0098	20.50	0.054	2.54	-0.098
11-18-2009	13.1485*	16:38:00	0.2047	0.0056	0.252	-0.0589	20.47	0.056	2.52	-0.589
11-18-2009	12.9550*	16:39:00	0.2042	0.5395	0.252	-0.0514	20.42	5.395	2.52	-0.514
11-18-2009	13.1609*	16:40:00	0.2239	0.9899	0.253	-0.0693	22.39	9.899	2.53	-0.693
11-18-2009	13.1609*	16:41:00	0.2276	3.7278	0.253	-0.0556	22.76	37.278	2.53	-0.556
11-18-2009	13.1609*	16:41:59	0.1998	4.5028	0.216	-0.0683	19.98	45.028	2.16	-0.683
11-18-2009	13.1609*	16:43:00	0.1912	4.9445	-0.217	-0.0602	19.12	49.445	-2.17	-0.602
11-18-2009	13.1609*	16:43:59	0.1918	5.0488	-0.456	-0.0639	19.18	50.488	-4.56	-0.639
11-18-2009	13.1609*	16:45:00	0.1917	5.0269	-0.456	-0.0672	19.17	50.269	-4.56	-0.672

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**APPENDIX B.**

**LABORATORY ANALYSES**

# **ENTECH ENGINEERING INC.**

P. O. Box 890746, Houston, Texas 77289-0746, (281)332-3118

## **Particulate Beaker Tare Weights**

		Initial Weight				
Beaker Number:	2078	2079	2080	2081	2083	2084
1	92.7036	99.1371	98.4559	101.1250	97.7793	100.1572
2	92.7041	99.1370	98.4557	101.1246	97.7798	100.1573
3						
4						
5						
Average (g)	92.7039	99.1371	98.4558	101.1248	97.7796	100.1573
		Final Weight				
	2078	2079	2080	2081	2083	2084
1	92.7113	99.1393	98.4660	101.1268	97.7843	100.1611
2	92.7108	99.1390	98.4658	101.1265	97.7847	100.1616
3						
4						
5						
Average (g)	92.7111	99.1392	98.4659	101.1267	97.7845	100.1614
		Total Weight Gain				
	2078	2079	2080	2081	2083	2084
Difference (g)	0.0072	0.0021	0.0101	0.0019	0.0050	0.0041
Total PM (mg)	7.2	2.1	10.1	1.9	5.0	4.1
Plant Name	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company
Location	Old Ocean, Tx	Old Ocean, Tx	Old Ocean, Tx	Old Ocean, Tx	Old Ocean, Tx	Old Ocean, Tx
Unit	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator
ID	Test 3 Impinger 1/BHW	Test 1 FHW	Test 1 Impinger 1/BHW	DI Blank 100mL	Test 2 Impinger 1/BHW	Test 3 FHW
Test Date	11/18/09	11/17/09	11/17/09	11/18/09	11/18/09	11/18/09

# ENTECH ENGINEERING INC.

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## Particulate Filter Tare Weights

Initial Weight				
Filter ID	1127	1110	1144	1141
1	0.7957	0.6988	0.7960	0.6085
2	0.7961	0.6991	0.7963	0.6085
3				
4				
5				
Average (g)	0.7959	0.6990	0.7962	0.6085
Final Weight				
Filter ID	1127	1110	1144	1141
1	0.8007	0.7049	0.8002	0.6085
2	0.8009	0.7044	0.8004	0.6085
3				
4				
5				
Average (g)	0.8008	0.7047	0.8003	0.6085
Total Weight Gain				
	1127	1110	1144	1141
Difference (g)	0.0049	0.0057	0.0041	0.0000
Total corrected for blank (g)	0.0049	0.0057	0.0041	0.0000
Total PM (mg)	4.9	5.7	4.1	0.0
Plant Name	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company	ConocoPhillips Company
Location	Old Ocean, TX	Old Ocean, TX	Old Ocean, TX	Old Ocean, TX
Unit	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator	Unit 39.1 Tail Gas Incinerator
ID	Test 1	Test 2	Test 3	Blank
Test Date	11/17/09	11/18/09	11/18/09	11/17/09

B2

# **GC VOC Analysis**

Pre-Calibration Standards  
(EPA Method 18)



**AIR LIQUIDE**Scott Specialty Gases  
Air Liquide America Specialty Gases LLC**CUSTOM CLASS**

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800 Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: Custom Class Calibration Standard****Product Information**Project No.: 04-64960-002  
Item No.: 0402B700100ZAL  
P.O. No.: 52180Cylinder Number: ALM066814  
Cylinder Size: AL  
Certification Date: 29May2008  
Expiration Date: 29May2010**Customer**ENTECH ENGINEERING INC.  
ROBERT STENBERG  
100 EAST NASA ROAD ONE  
SUITE 407  
WEBSTER, TX 77598**CERTIFIED CONCENTRATION**

<u>Component Name</u>	<u>Concentration (Moles)</u>	<u>Accuracy (+/-%)</u>
N-BUTANE	15.2 PPM	2
ETHANE	15.6 PPM	2
N-HEXANE	15.5 PPM	2
METHANE	16.2 PPM	2
N-PENTANE	14.7 PPM	2
PROPANE	15.8 PPM	2
NITROGEN	BALANCE	

**TRACEABILITY****Description**

ANALYTICAL TRACEABILITY

**Traceability Type**

GAS STANDARDS

**Traceable To**

APPROVED BY:

  
CRISSA MARTIN

DATE: 05/29/08



**AIR LIQUIDE**Scott Specialty Gases  
Air Liquide America Specialty Gases LLC**CUSTOM CLASS**

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800 Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: Custom Class Calibration Standard****Product Information**Project No.: 04-64960-001  
Item No.: 0402B700101ZAL  
P.O. No.: 52180Cylinder Number: ALM031355  
Cylinder Size: AL  
Certification Date: 29May2008  
Expiration Date: 29May2010**Customer**ENTECH ENGINEERING INC.  
ROBERT STENBERG  
100 EAST NASA ROAD ONE  
SUITE 407  
WEBSTER, TX 77598**CERTIFIED CONCENTRATION**

<u>Component Name</u>	<u>Concentration (Moles)</u>	<u>Accuracy (+/-%)</u>
N-BUTANE	101. PPM	2
ETHANE	101. PPM	2
N-HEXANE	96. PPM	2
METHANE	95. PPM	2
N-PENTANE	97. PPM	2
PROPANE	97. PPM	2
NITROGEN	BALANCE	

**TRACEABILITY**

<u>Description</u>	<u>Traceability Type</u>	<u>Traceable To</u>
ANALYTICAL TRACEABILITY	GAS STANDARDS	SRS, NIST

APPROVED BY:

  
CRISSA MARTIN

DATE:

05/29/08

**AIR LIQUIDE**Scott Specialty Gases  
Air Liquide America Specialty Gases LLC**CUSTOM CLASS**

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800 Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: Custom Class Calibration Standard****Product Information**Project No.: 04-64960-003  
Item No.: 0402B700101ZAL  
P.O. No.: 52180Cylinder Number: ALM043423  
Cylinder Size: AL  
Certification Date: 29May2008  
Expiration Date: 28May2010**Customer**ENTECH ENGINEERING INC.  
ROBERT STENBERG  
100 EAST NASA ROAD ONE  
SUITE 407  
WEBSTER, TX 77598**CERTIFIED CONCENTRATION****Component Name****Concentration  
(Moles)****Accuracy  
(+/-%)**

N-BUTANE	1,010. PPM	2
ETHANE	1,010. PPM	2
N-HEXANE	980. PPM	2
METHANE	1,020. PPM	2
N-PENTANE	990. PPM	2
PROPANE	1,000. PPM	2
NITROGEN	BALANCE	

**TRACEABILITY****Description****Traceability Type****Traceable To**

ANALYTICAL TRACEABILITY

GAS STANDARDS

SRS

APPROVED BY:

  
CRISSA MARTIN

DATE: 05/29/08

# ENTECH ENGINEERING INC.

## Gas Chromatograph Analysis Standard Curve Data

GC #1 FID-2 ( Range 1)

Methane Standard (ALM066814; ALM031355; ALM043423)


08/27/09

(Retention Time = 1.062 min)

Concentration (ppmv)	Area 1 (AU)	Area 2 (AU)	Area 3 (AU)	Average Area (AU)	Precision Error (%)
0.0	0	0	0	0.0	NA
16.2	2641	2644	2618	2634.3	0.25%
95.0	19034	18757	18565	18785.3	1.32%
1020.0	200507	197983	202325	200271.7	0.12%

## Regression Data (Zero-Forced)

Constant	0.0000
Std Err of Y Est	1.653
R Squared	0.999989
No. of Observations	4
Degrees of Freedom	3
X Coefficient(s)	5.092950E-03
Std Err of Coef.	8.218916E-06



Operator

km

Date

Sep 01, 2009

BS

# ENTECH ENGINEERING INC.

## Gas Chromatograph Analysis Standard Curve Data GC #1 FID-2 ( Range 1) Ethane Standard (ALM066814; ALM031355; ALM043423)


08/27/09

(Retention Time = 2.990 min)

Concentration (ppmv)	Area 1 (AU)	Area 2 (AU)	Area 3 (AU)	Average Area (AU)	Precision Error (%)
0.0	0	0	0	0.0	NA
15.6	4860	4898	4841	4866.3	0.13%
101.0	35725	35188	34838	35250.3	1.35%
1010.0	395934	391092	400509	395845.0	0.02%

### Regression Data (Zero-Forced)

Constant	0.0000
Std Err of Y Est	6.618
R Squared	0.999816
No. of Observations	4
Degrees of Freedom	3
X Coefficient(s)	2.554070E-03
Std Err of Coef.	1.665118E-05




Operator km Date Sep 01, 2009

89

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 15 PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\15PA2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:26:45 AM  
PRINTED: 9/1/2009 8:26:50 AM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.062	2641
Ethane	2.990	4860
Propane	8.359	7191
n-Butane	11.994	10212
n-Pentane	15.882	12994
n-Hexane	24.042	15659
Totals		53557

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 15 PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\15PA3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 8/31/2009 6:00:57 PM  
PRINTED: 9/1/2009 8:27:00 AM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.062	2644
Ethane	2.990	4898
Propane	8.364	7210
n-Butane	11.996	10263
n-Pentane	15.882	13078
n-Hexane	24.050	15845

Totals		53938
--------	--	-------

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 15 PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\15PA4.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:27:10 AM  
PRINTED: 9/1/2009 8:27:14 AM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.055	2618
Ethane	2.980	4841
Propane	8.358	7237
n-Butane	11.988	10123
n-Pentane	15.856	12926
n-Hexane	24.015	15678

Totals		53423
--------	--	-------

B12

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\100PA1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:27:43 AM  
PRINTED: 9/1/2009 8:27:48 AM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.062	19034
Ethane	2.986	35725
Propane	8.359	52661
n-Butane	11.991	74462
n-Pentane	15.871	91881
n-Hexane	24.028	106642

Totals		380405
--------	--	--------



# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\100PA2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:27:59 AM  
PRINTED: 9/1/2009 8:28:02 AM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.057	18757
Ethane	2.973	35188
Propane	8.351	51943
n-Butane	11.986	73465
n-Pentane	15.862	90786
n-Hexane	24.013	105449

Totals		375588
--------	--	--------

B14

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100PPM PARAFFIN  
FILE: E:\Calibration\GC#1\CL082709\100PA3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:28:25 AM  
PRINTED: 9/1/2009 8:28:27 AM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.060	18565
Ethane	2.990	34838
Propane	8.354	51231
n-Butane	11.979	72588
n-Pentane	15.849	90012
n-Hexane	24.002	104695

Totals		371929
--------	--	--------

B15

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 1000PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\1000PA1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:28:37 AM  
PRINTED: 9/1/2009 8:28:40 AM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.058	200507
Ethane	2.986	395934
Propane	8.351	587622
n-Butane	11.978	784763
n-Pentane	15.842	961837
n-Hexane	23.978	1118892

Totals		4049555
--------	--	---------

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 1000PPM PARAFFIN  
FILE: E:\Calibration\GC#1\ICL082709\1000PA2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:28:51 AM  
PRINTED: 9/1/2009 8:28:53 AM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.060	197983
Ethane	3.005	391092
Propane	8.353	580256
n-Butane	11.976	775704
n-Pentane	15.832	953281
n-Hexane	23.962	1115965

Totals		4014281
--------	--	---------

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 1000PPM PARAFFIN  
FILE: E:\Calibration\GC#1\CL082709\1000PA3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/1/2009 8:29:01 AM  
PRINTED: 9/1/2009 8:29:04 AM  
USER: System



## FID-2 Results

Name	Retention Time	Area
Methane	1.062	202325
Ethane	2.986	400509
Propane	8.361	595947
n-Butane	11.993	797032
n-Pentane	15.879	980913
n-Hexane	24.038	1152943

Totals		4129669
--------	--	---------

B18

# **GC VOC Analysis**

Post-Calibration Standards  
(EPA Method 18)

**GC #1 FID (Channel B@Range 1) - Initial Calibration: August 27, 2009**  
**Paraffin Standards Post Check Result**  
**11/20/09**

Sample ID No.	GC Run No.	Area Counts	Concentration	Area Counts	Concentration	Area Counts	Concentration	Area Counts	Concentration	Area Counts	Concentration
Initial Cal. Standard Avg.		Methane	Methane	Ethane	Propane	n-Butane	n-Butane	n-Pentane	n-Pentane		
		area counts	ppmv	area counts	ppmv	area counts	ppmv	area counts	ppmv		
1	1	18785.3	97.5	35250.3	51945.0	73505.0	101.00	90893.0	97.00		
2	2	18997.4	96.8	36435.7	54457.1	74701.1	96.07	91797.7	94.18		
3	3	18903.3	96.3	37083.3	53277.7	75639.9	97.28	92184.4	94.57		
		Average =	96.8		54753.3	75310.0	96.85	92583.3	94.98		
		Standard Response Factor =	197.74				96.73		94.58		
		QA Response Factor =	200.14				727.77		937.04		
		Mean Response Factor =	198.94				744.72		950.39		
		Percent Difference, % =	1.21				736.25		943.72		
		Pass/Fail Criterion (<5%) =	Pass				2.30		1.41		
							Pass		Pass		


020

Sample ID No.	GC Run No.	Area Counts	Concentration
Initial Cal. Standard Avg.		n-Hexane	n-Hexane
		area counts	ppmv
1	1	105595.3	96.000
2	2	101790.0	88.374
3	3	102862.7	89.305
		108239.9	93.973
		Average =	90.551
		Standard Response Factor =	1099.95
		QA Response Factor =	1086.43
		Mean Response Factor =	1093.19
		Percent Difference, % =	1.24
		Pass/Fail Criterion (<5%) =	Pass

Operator SP Date 11-20-09

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100ppm paraffin  
FILE: E:\Calibration\GC#1\post\_112009\100pa2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/23/2009 10:47:35 AM  
PRINTED: 11/23/2009 10:47:46 AM  
USER: System



## FID-2 Results

Name	Retention Time	Area
Methane	1.035	19140
Ethane	2.896	36876
Propane	8.264	54457
n-Butane	11.931	74701
n-Pentane	15.757	91797
n-Hexane	23.802	101790


Totals		378761
--------	--	--------

821



# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100ppm paraffin  
FILE: E:\Calibration\GC#1\post\_112009\100pa3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/23/2009 10:48:52 AM  
PRINTED: 11/23/2009 10:48:58 AM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.035	18996
Ethane	2.928	36435
Propane	8.263	53277
n-Butane	11.921	75639
n-Pentane	15.736	92184
n-Hexane	23.778	102862

Totals		379393
--------	--	--------

B22

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 100ppm paraffin  
FILE: E:\Calibration\GC#1\post\_112009\100pa4.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/23/2009 10:50:08 AM  
PRINTED: 11/23/2009 10:50:13 AM  
USER: System



## FID-2 Results

Name	Retention Time	Area
Methane	1.038	18903
Ethane	2.925	37083
Propane	8.266	54753
n-Butane	11.926	75310
n-Pentane	15.752	92583
n-Hexane	23.785	108239

Totals		386871
--------	--	--------

# **GC VOC Analysis**

QA/QC Results and Raw Data  
(EPA Method 18)

**ENTECH ENGINEERING INC.**

P.O. Box 890746, Houston, Texas 77289-0746, (281)332-3118

**GC#1 - Paraffins & Olefins Minimum Detection Limits Study (EPA Method 18) - September 02, 2009**

Initial Calibration Date: August 27, 2009 - GC#1 FID-2 (Range1), Haysep Q Column


ID	Run 1	Run 2	Run 3	Concentration 1 (ppmv)	Concentration 2 (ppmv)	Concentration 3 (ppmv)	Average Conc. (ppmv)	Detection Limit	Recovered Concentration (ppmv)	Spiking Concentration (ppmv)	Spike Percent Recovery	Control Limit (70-130)%
Methane	239	226	240	1.22	1.15	1.22	1.20	0.24	1.20	1.53	78%	Pass
Ethane	447	410	447	1.14	1.05	1.14	1.11	0.22	1.11	1.52	73%	Pass
Propane	713	693	729	1.21	1.18	1.24	1.21	0.24	1.21	1.50	81%	Pass
n-Butane	1024	999	1065	1.32	1.31	1.39	1.34	0.27	1.34	1.52	88%	Pass
n-Pentane	1289	1222	1203	1.32	1.25	1.23	1.27	0.25	1.27	1.49	86%	Pass
n-Hexane	1585	1489	1518	1.38	1.29	1.32	1.33	0.27	1.33	1.47	90%	Pass
Ethylene	469	481	463	1.27	1.30	1.25	1.27	0.25	1.27	1.50	85%	Pass
Propylene	596	696	623	1.26	1.26	1.13	1.22	0.24	1.22	1.52	81%	Pass
1-Butene	983	962	991	1.31	1.29	1.33	1.31	0.26	1.31	1.49	88%	Pass
1-Pentene	1324	1367	1350	1.44	1.49	1.47	1.47	0.29	1.47	1.50	98%	Pass
1-Hexene	1685	1680	1617	1.59	1.58	1.52	1.56	0.31	1.56	1.49	105%	Pass

Note: The MDL Standard Prepared by Diluting 1.5ml of Paraffin/Olefins Standard (1000ppm) to 1L of Nitrogen Bag.

Operator km Date Sep 03/2009

# Entech Engineering Inc. League City, Texas

SAMPLE ID: paraffin MDL  
FILE: E:\Calibration\GC#1\Post\_090209\para\_md11.dat  
METHOD: E:\Calibration\GC#1\FID-2 30min\_MDL.met  
ACQUIRED: 9/2/2009 7:14:27 PM  
PRINTED: 9/2/2009 7:14:33 PM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.067	239
Ethane	2.998	447
Propane	8.394	713
n-Butane	12.043	1024
n-Pentane	15.909	1289
n-Hexane	24.207	1585
Totals		5297

# Entech Engineering Inc. League City, Texas

SAMPLE ID: paraffin MDL  
FILE: E:\Calibration\GC#1\Post\_090209\para\_md12.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/3/2009 8:32:02 AM  
PRINTED: 9/3/2009 8:32:29 AM  
USER: System




## FID-2 Results

Name	Retention Time	Area
Methane	1.063	226
Ethane	2.990	410
Propane	8.381	693
n-Butane	11.899	999
n-Pentane	15.901	1222
n-Hexane	24.042	1489

Totals		5039
--------	--	------

# Entech Engineering Inc. League City, Texas

SAMPLE ID: paraffin MDL  
FILE: E:\Calibration\GC#1\Post\_090209\para\_md13.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 9/3/2009 8:40:47 AM  
PRINTED: 9/3/2009 8:40:58 AM  
USER: System



## FID-2 Results

Name	Retention Time	Area
Methane	1.067	240
Ethane	2.998	447
Propane	8.368	729
n-Butane	12.059	1065
n-Pentane	15.934	1203
n-Hexane	24.068	1518

Totals		5202
--------	--	------

# **GC VOC Analysis**

Gas Dilution System Validation  
(EPA Method 205)



# ENTECH ENGINEERING INC.

## Gas Chromatograph Analysis Standard Curve Data

GC #1 FID (Channel B @ Range 1)


Propane Standard (Air Liquide CC115715, 490ppm Propane certified)

10/14/08

Dilution System : ENGASDL002

Dilution Factor	Predicted Concentration (ppmv)	Area Count				Precision Error			Calculated Concentration (ppmv)	% Difference (%)
		Run 1	Run 2	Run 3	Average	Run 1 (%)	Run 2 (%)	Run 3 (%)		
1:50	9.80	4979	4952	5015	4982.0	-0.06	-0.60	0.66	9.80	0.00
1:10	49.00	25288	24650	25176	25038.0	1.00	-1.55	0.55	48.90	-0.20

Constant	0.0625
R Squared	1.0000
No. of Observations	2
Degrees of Freedom	0
X Coefficient(s)	1.954527E-03



Operator

pm

Date

Oct 20 / 2008

30



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: Entech Engineering  
P.O. Number: 51831  
Item Number:  
Notes:

Cyl. Number: CC115715

Shipping Order #: 25000221  
Transfer #: 25000221  
LOT #: LPX220350  
Valve: CGA350  
Cyl. Pressure\*: 1900psig

\*Cylinder should not be used when gas pressure is below 150 psig

Assay Date: 7-May-07

Expiration Date: 6-May-10

Component	Requested Concentration	Assay Concentration
Propane	500 ppm	490 ±5 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS290	GMIS290	497.0	ppm	2.52	C3H8	N2	CC61240	5/3/2008	N.A.

#### Analysis Information:

Component 1: Propane		First Triad Analysis On: 5/7/2007					Second Triad Analysis On:			
Analyzer Information		Zero	Trial 1	Trial 2	Trial 3	Units	Zero	Trial 1	Trial 2	Trial 3
Manufacturer:	HP	0.00	0.00	0.00	0.00		Reference			
Model Number:	6890	474.10	477.14	475.61	475.61		Candidate			
Serial Number:	8295	469.16	469.72	468.19	468.19		Result			
Analytical Principle:	GC-FID	480.25	490.84	489.23	489.23	ppm				
MPC Calibrated:	05/07/07									
		Mean Result:		490.11	ppm		Mean Result:			

Analyst Signature: \_\_\_\_\_

Warren Pereira


Calculated by: \_\_\_\_\_

Warren Pereira

B31

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\9.8r1  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 9.8ppm propane  
Acquired : Oct 20, 2008 13:18:53  
Printed : Oct 20, 2008 13:31:04  
User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
	1.00	41	0.0
Propane	8.16	4979	9.7
Totals :		5020	9.7

B32

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\9.8r2  
 Method : f:\gclab2\gc#1\Fid\_m205.met  
 Sample ID : 9.8ppm propane  
 Acquired : Oct 20, 2008 13:33:34  
 Printed : Oct 20, 2008 13:46:17  
 User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	42	0.0
	8.13	4952	9.6
Totals :		4994	9.6

833

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\9.8r3  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 9.8ppm propane  
Acquired : Oct 20, 2008 13:48:23  
Printed : Oct 20, 2008 13:58:35  
User : System




## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	38	0.0
	8.14	5015	9.7
Totals :		5053	9.7

B34

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\49r5  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 49ppm propane  
Acquired : Oct 17, 2008 14:58:35  
Printed : Oct 17, 2008 16:08:28  
User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	40	0.0
	8.17	25288	49.1
Totals :		25328	49.1

635

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\49r2  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 49ppm propane  
Acquired : Oct 14, 2008 14:58:24  
Printed : Oct 14, 2008 15:11:06  
User : System




## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00	50	0.0
	8.14	24650	47.9
Totals :		24700	47.9

B36

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\49r1  
Method : f:\gclab2\gc#1\fid\_m205.met  
Sample ID : 49ppm propane  
Acquired : Oct 14, 2008 14:43:35  
Printed : Oct 20, 2008 12:57:13  
User : System



Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00	42	0.0
	8.15	25176	48.9
Totals :		25218	48.9

B37



# ENTECH ENGINEERING INC.

P.O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## Gas Dilution System Validation Study (EPA Method 205)

Gas Dilution System ID: ENGASDL002  
Calibration Standard: Air Liquide CC115715  
Calibration Date: October 14, 2008  
Propane X Coefficient: 1.954527E-03

Validation Gas Standard: Air Liquide CC161053  
Validation Date: October 17, 2008  
Certified Concentration: 31.0 ppm

Instrument: GC#1 FID-2, Range 1 Column: 80/100 Haysep Q, 6 ft X 1/8 in, S.S. (316) Sample Injection Loop: 1000 ul

ID	Run 1 Area	Run 2 Area	Run 3 Area	Average Area	Precision Error (%)			Average Conc. (ppmv)	Certified Std. Conc. (ppmv)	Percent Difference	Control Limit ( < 2 )%
					Run 1	Run 2	Run 3				
Propane	15537	15718	15642	15632	0.61%	0.55%	0.06%	30.55	31.00	1.45%	Pass

Operator

pm

Date

Oct 20 / 2008

B38



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: Entech Engineering  
P.O. Number: 51831  
Item Number:  
Notes:

Cyl. Number: CC161053

Shipping Order #: 25000357  
Transfer #: 25000357  
LOT #: LPX220344  
Valve: CGA350  
Cyl. Pressure\*: 1900psig

\*Cylinder should not be used when gas pressure is below 150 psig

Assay Date: 3-May-07

Expiration Date: 2-May-10

Component	Requested Concentration	Assay Concentration
Propane	30 ppm	31.0 ±0.5 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS281	GMIS281	49.9	ppm	0.184	C3H8	N2	CC28041	5/3/2008	N.A.

#### Analysis Information:

Component 1: Propane		First Triad Analysis On: 5/3/2007					Second Triad Analysis On:			
Analyzer Information		Trial 1	Trial 2	Trial 3	Units		Trial 1	Trial 2	Trial 3	Units
Manufacturer:	HP	0.00	0.00	0.00						
Model Number:	6890									
Serial Number:	8295									
Analytical Principle:	GC-FID									
MPC Calibrated:	04/05/07									
Zero		47.99	47.96	47.93			Zero			
Reference		29.79	29.77	29.84			Reference			
Candidate		31.00	30.98	31.05	ppm		Candidate			
Result							Result			
Mean Result:		31.01		ppm			Mean Result:			

Analyst Signature: Warren Pereira Warren Pereira

Calculated by: Warren Pereira Warren Pereira

B39


Mix Assayed At: Air Liquide America

11426 Fairmont Pkwy, La Porte, TX, 77571

Phone: (281)474-8400 Fax: (281)474-8419

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\31r4  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 31ppm propane  
Acquired : Oct 17, 2008 10:06:24  
Printed : Oct 17, 2008 10:31:37  
User : System



## Channel B Results


Peak	Retention Time	Area	Concentration, ppmv
Propane	8.17	15537	30.2
Totals :		15537	30.2

B40

**Entech Engineering Inc.**  
**Webster, Texas**

Page 1 of 1

File : f:\gclab2\gc#1\calibr~1\2008\m205val\31r5  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 31ppm propane  
Acquired : Oct 17, 2008 10:46:51  
Printed : Oct 17, 2008 11:11:43  
User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	8.17	15642	30.4
Totals :		15642	30.4

B41

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\31r6  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 31ppm propane  
Acquired : Oct 17, 2008 11:14:00  
Printed : Oct 20, 2008 14:12:01  
User : System



Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	8.17	15718	30.5
Totals :		15718	30.5

B42

# ENTECH ENGINEERING INC.


## Gas Chromatograph Analysis Standard Curve Data GC #1 FID (Channel B @ Range 1) Propane Standard (AirGas SG1014905, 8.49% Propane certified)

10/20/08

Dilution System : ENGASDL003

Dilution Factor	Predicted Concentration (ppmv)	Area Count				Precision Error			Calculated Concentration (ppmv)	% Difference (%)
		Run 1	Run 2	Run 3	Average	Run 1 (%)	Run 2 (%)	Run 3 (%)		
1:1000	84.90	43801	43004	44511	43772.0	0.07	-1.75	1.69	85.96	1.25
1:100	849.00	436916	432059	429614	432863.0	0.94	-0.19	-0.75	850.06	0.12

Constant	-1.0598
R Squared	1.0000
No. of Observations	2
Degrees of Freedom	0
X Coefficient(s)	1.963808E-03



Operator pm Date Nov 05/2008

B43

**AIR LIQUIDE**

# CERTIFICATE OF ANALYSIS

Customer : Entech Engineering, Inc.

P.O. Number : 51915

Doc. # : 26216901-1B

Mix/Lot # : LPX225415

Item Number :

Valid Until : 11 August, 2012

Specification : Custom Certified

Phase : GAS

Cyl. Size : 30AL Valve : CGA 350

Pressure : 767 Psia

Volume : 55 SCF

Cylinder Number: **CC78830**

Requested Concentrations		Actual Concentration Mole	Analytical Uncertainty	Equipment Used		
Mole	Component			Scale	Analyt. Inst.	Calibration Standard
Balance	NITROGEN	Balance		1399		
8.5 %	PROPANE	8.49 %	+/-2%	1399		

This mixture was prepared and certified by weight using one or more scales certified against weights traceable to N.I.S.T.

Comments:

Dewpoint calculated to 40 F, unless otherwise stated. Improper storage or use may affect the accuracy of this standard.

Certified by


Date: 13-Aug-2007

11426 Fairmont Pkwy -- LaPorte, TX 77571  
Phone (281) 474-8400 Fax (281) 474-8419 USA (800) 248-1427

ISO: 9001-2000

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\84.9r5  
Method : f:\gclab2\gc#1\fid\_m205.met  
Sample ID : 84.9ppm propane  
Acquired : Oct 20, 2008 17:04:15  
Printed : Nov 3, 2008 11:52:05  
User : System



## Channel B Results


Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	23	0.0
	8.12	44511	86.4
Totals :		44534	86.4

B45



**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\84.9r9  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 84.9ppm  
Acquired : Nov 3, 2008 16:07:01  
Printed : Nov 3, 2008 16:29:20  
User : System




## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00	32	0.0
	8.16	43004	83.5
Totals :		43036	83.5

B46

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\84.9r7  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 84.9ppm  
Acquired : Nov 3, 2008 15:34:49  
Printed : Nov 3, 2008 15:47:20  
User : System




## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	35	0.0
	8.17	43801	85.1
Totals :		43836	85.1

B47

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\849r4  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 849ppm propane  
Acquired : Oct 21, 2008 16:45:17  
Printed : Oct 21, 2008 16:55:02  
User : System




## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00 8.12	29 429614	0.0 834.2
Totals :		429643	834.2

*B48*

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\849r5  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 849ppm  
Acquired : Nov 3, 2008 12:46:33  
Printed : Nov 3, 2008 13:17:05  
User : System



## Channel B Results


Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	51	0.0
	8.16	436916	848.4
Totals :		436967	848.4

B 49

**Entech Engineering Inc.**  
**Webster, Texas**

Page 1 of 1

File : f:\gclab2\gc#1\calibr~1\2008\m205val\849r6  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 849ppm  
Acquired : Nov 3, 2008 13:40:15  
Printed : Nov 3, 2008 14:11:59  
User : System



Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	45	0.0
	8.17	432059	839.0
Totals :		432104	839.0

B50

# ENTECH ENGINEERING INC.

P.O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## Gas Dilution System Validation Study (EPA Method 205)

Gas Dilution System ID: ENGASDL003  
Calibration Standard: AirGas CC78830  
Calibration Date: October 20, 2008  
Propane X Coefficient: 1.963808E-03

Validation Gas Standard: Air Liquide (CC115715)  
Validation Date: November 04, 2008  
Certified Concentration: 490 ppmv

Instrument: GC#1 FID-2, Range 1 Column: 80/100 Haysep Q, 6 ft X 1/8 in, S.S. (316) Sample Injection Loop: 1000 ul

ID	Run 1 Area	Run 2 Area	Run 3 Area	Average Area	Precision Error (%)			Average Conc. (ppmv)	Certified Std. Conc. (ppmv)	Percent Difference	Control Limit ( < 2 )%
					Run 1	Run 2	Run 3				
Propane	242063	245641	249971	245892	1.56%	0.10%	1.66%	482.88	490.00	1.45%	Pass

Operator km Date Nov 05/2008

B51

# CERTIFICATE of ANALYSIS

## Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

<b>Customer:</b> Entech Engineering		<b>Cyl. Number:</b> CC115715	<b>Shipping Order #:</b> 25000221
<b>P.O. Number:</b> 51831			<b>Transfer #:</b> 25000221
<b>Item Number:</b>			<b>LOT #:</b> LPX220350
<b>Notes:</b>			<b>Valve:</b> CGA350
			<b>Cyl. Pressure:*</b> 1900psig
<b>Assay Date:</b> 7-May-07		<b>Expiration Date:</b> 6-May-10	<small>*Cylinder should not be used when gas pressure is below 150 psig</small>



Component	Requested Concentration	Assay Concentration
Propane	500 ppm	490 ±5 ppm
Nitrogen	Balance	Balance

**Reference Standard(s) Employed For Analysis:**

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS290	GMIS290	497.0	ppm	2.52	C3H8	N2	CC61240	5/3/2008	N.A.


**Analysis Information:**

Component 1: Propane												
Analyzer Information		First Triad Analysis On: 5/7/2007					Second Triad Analysis On:					
Manufacturer:	HP		Trial 1	Trial 2	Trial 3	Units		Trial 1	Trial 2	Trial 3	Units	
Model Number:	6890	Zero	0.00	0.00	0.00		Zero					
Serial Number:	8295	Reference	474.10	477.14	475.61		Reference					
Analytical Principle:	GC-FID	Candidate	469.16	469.72	468.19		Candidate					
MPC Calibrated:	05/07/07	Result	490.25	490.84	489.23	ppm	Result					
		Mean Result: 490.11					ppm	Mean Result:				

Analyst Signature:  Warren Pereira      Calculated by:  Warren Pereira

B52

File : f:\gclab2\gc#1\calibr~1\2008\m205val\490r5  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 490ppm  
Acquired : Nov 4, 2008 12:35:42  
Printed : Nov 5, 2008 15:38:41  
User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00 8.13	19 242063	0.0 470.0
Totals :		242082	470.0

B-53



File : f:\gclab2\gc#1\calibr~1\2008\m205val\490r7  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 490ppm  
Acquired : Nov 4, 2008 15:22:50  
Printed : Nov 5, 2008 15:39:28  
User : System




Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.01	40	0.0
	8.12	245641	477.0
Totals :		245681	477.0

B54

**Entech Engineering Inc.**  
**Webster, Texas**

File : f:\gclab2\gc#1\calibr~1\2008\m205val\490r8  
Method : f:\gclab2\gc#1\Fid\_m205.met  
Sample ID : 490ppm  
Acquired : Nov 4, 2008 16:24:22  
Printed : Nov 5, 2008 15:40:05  
User : System



## Channel B Results

Peak	Retention Time	Area	Concentration, ppmv
Propane	1.00	34	0.0
	8.13	249971	485.4
Totals :		250005	485.4

*455*

# **GC VOC Analysis**

Sample Results and Raw Data  
(EPA Method 18)

*B-56*

**ENTECH ENGINEERING INC.****ConocoPhillips. - Sweeny Refinery, Old Ocean, Texas - November 17, 2009****ES09-11-20832 / stack T1**

	GC #1	X Coefficient	DF	Area Counts				Concentration (%)				High Standard (%)
				Run 1	Run 2	Run 3	Average	Run 1	Run 2	Run 3	Averages	
Methane	5.09E-03	1		385	397	385	389	1.961	2.022	1.961	1.981	1020.000
Ethane	2.55E-03	1		0	0	0	0	0.000	0.000	0.000	0.000	1010.000

GC#1 FID initial Calibration Date: 08-27-2009.

DF = Dilution Factor.

B57

**ENTECH ENGINEERING INC.****ConocoPhillips. - Sweeny Refinery, Old Ocean, Texas - November 18, 2009****ES09-11-20851 / stack T2**

	GC #1	X Coefficient	DF	Area Counts				Concentration (%)				High Standard (%)
				Run 1	Run 2	Run 3	Average	Run 1	Run 2	Run 3	Averages	
Methane	5.09E-03	1		397	415	433	415	2.022	2.114	2.205	2.114	1020.000
Ethane	2.55E-03	1		0	0	0	0	0.000	0.000	0.000	0.000	1010.000

GC#1 FID initial Calibration Date: 08-27-2009.

DF = Dilution Factor.

**ES09-11-20852 / stack T3**

	GC #1	X Coefficient	DF	Area Counts				Concentration (%)				High Standard (%)
				Run 1	Run 2	Run 3	Average	Run 1	Run 2	Run 3	Averages	
Methane	5.09E-03	1		944	927	914	928	4.808	4.721	4.655	4.728	1020.000
Ethane	2.55E-03	1		0	0	0	0	0.000	0.000	0.000	0.000	1010.000

GC#1 FID initial Calibration Date: 08-27-2009.

DF = Dilution Factor.

B58



**ENTECH ENGINEERING INC.**  
P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3178

P. O. Box 890746 • Houston, Texas 77289-0746 • (281)332-3118

## Chain of Custody


[illegible]

Original

—

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20832 stack t1  
FILE: E:\Sample\2009\ConocoPhillips Company\111709\20832r1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/18/2009 10:35:21 PM  
PRINTED: 11/19/2009 4:28:41 PM  
USER: System



**FID-2 Results**


Name	Retention Time	Area
Methane	1.030	385
Totals		385

*AG1*



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20832 stack t1  
FILE: E:\Sample\2009\ConocoPhillips Company\111709\20832r2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/18/2009 9:23:42 PM  
PRINTED: 11/19/2009 4:44:10 PM  
USER: System



**FID-2 Results**


Name	Retention Time	Area
Methane	1.032	397

Totals		397
--------	--	-----

B62

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20832 stack t1  
FILE: E:\Sample\2009\ConocoPhillips Company\111709\20832r3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/18/2009 10:01:22 PM  
PRINTED: 11/19/2009 4:46:29 PM  
USER: System




**FID-2 Results**

Name	Retention Time	Area
Methane	1.027	385
Totals		385

B63

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20851 stack t2  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20851r1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 6:19:41 PM  
PRINTED: 11/19/2009 6:19:59 PM  
USER: System




**FID-2 Results**

Name	Retention Time	Area
Methane	1.032	397
Totals		397

B64

**Entech Engineering Inc. League City, Texas**


SAMPLE ID: 20851 stack t2  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20851r2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 6:52:17 PM  
PRINTED: 11/19/2009 6:52:30 PM  
USER: System



FID-2 Results		
Name	Retention Time	Area
Methane	1.037	415
Totals		415

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20851 stack t2  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20851r3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 7:25:55 PM  
PRINTED: 11/19/2009 7:26:07 PM  
USER: System




**FID-2 Results**

Name	Retention Time	Area
Methane	1.035	433
Totals		433

B66

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 20852 stack t3  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20852r1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 7:59:53 PM  
PRINTED: 11/19/2009 7:59:58 PM  
USER: System



## FID-2 Results


Name	Retention Time	Area
Methane	1.035	944

Totals		944
--------	--	-----

B67

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20852 stack t3  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20852r2.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 8:33:25 PM  
PRINTED: 11/19/2009 8:33:34 PM  
USER: System




**FID-2 Results**

Name	Retention Time	Area
Methane	1.037	927
Totals		927

B68

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 20852 stack t3  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\20852r3.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 9:06:48 PM  
PRINTED: 11/19/2009 9:06:54 PM  
USER: System



**FID-2 Results**

Name	Retention Time	Area
Methane	1.040	914
Totals		914


B69



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: blk  
FILE: E:\Sample\2009\ConocoPhillips Company\111709\blkr1.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/18/2009 7:59:03 PM  
PRINTED: 11/18/2009 7:59:17 PM  
USER: System

710



**FID-2 Results**


<b>Name</b>	<b>Retention Time</b>	<b>Area</b>
-------------	-----------------------	-------------

B70

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: blk  
FILE: E:\Sample\2009\ConocoPhillips Company\111809\blk.dat  
METHOD: E:\Calibration\GC#1\FID-2 paraffin 30min.met  
ACQUIRED: 11/19/2009 5:46:48 PM  
PRINTED: 11/19/2009 5:47:02 PM  
USER: System

710



FID-2 Results


Name	Retention Time	Area
------	----------------	------

B 71

**Sulfur Analysis - GC / FPD**  
Column Resolution Check  
(EPA method 15)

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 5ppm @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\Std\_Resolution Chk.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD.met  
ACQUIRED: 11/17/2009 10:24:25 AM  
PRINTED: 11/17/2009 10:29:04 AM  
USER: System




## FPD Results

Name	Retention Time	Area
COS	1.737	3508864
H2S	2.187	2878128
Totals		6386992

473

# Entech Engineering Inc. League City, Texas

SAMPLE ID: resolution check  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\res\_chk.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 8:47:27 AM  
PRINTED: 11/18/2009 8:56:11 AM  
USER: System



## FPD Results

Name	Retention Time	Area
COS	2.035	3402408
H2S	2.505	2806964
CS2	4.251	6332231
	5.695	6766

Totals		12548369
--------	--	----------

74

**Sulfur Analysis - GC / FPD**  
Pre-Test Calibration  
(EPA method 15)

*B-75*

**P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118**

Gas Chromatograph Information			
GC Type:	Shimadzu GC-14A		
Detector:	Photo Ionization Detector (FPD). Channel: A		
GC ID #:	6	DAS:	Shimadzu EZChrom
GC Serial #:	C10552911593		
Column Type:	Chromosil 310 Column, 8'x1/8" OD Teflon.		
Date in Service:	Oct. 14, 2005	Cond.:	Used
Location:	Entech GC Lab		
Gas Chromatograph Operating Parameters			
GC Oven Temperature Parameters:			
Isothermal Column Temperature (C):	37	Injector Temperature (C):	220
Initial Column Temperature (C):	37	Detector Temperature (C):	220
Duration of initial temperature (min):	na	TCD Oven Temperature (C):	na
Program Rate (C/min):	na	Methanizer Temperature (C):	na
Final Temperature (C):	na		
Final Time (min):	na		
Carrier Gas Mass Flow Setting (Kpa):	100	Carrier Gas Type:	Nitrogen
Air Flow (Kpa):	50	Hydrogen Flow (Kpa):	50
Detector Range:	2	Back-flush Time (min):	NA
Sample Size (ul):	250	TCD Bridge Current (mA):	NA
Calibration Data			
	Retention	ID	Regression Data: $y = mx + b$
	Time	Time Band	Where: $x = \text{Area}$ & $y = \text{molar ppm}$
ID	(min)	(+/- min)	Range      m      b      r
* See Data Calculations Section			

~~6~~ 96

# ENTECH ENGINEERING INC.

## Gas Chromatograph Analysis Standard Curve Data

GC #6 FPD ( Range 2)

H<sub>2</sub>S Standard - Dynacalibrator


11/17/09

(Retention Time = 2.49 min)

Concentration (ppmv)	Area 1 (AU)	Area 2 (AU)	Area 3 (AU)	Avg. Area (Square Root) (AU)	Precision Error (%)
0.00	0	0	0	0.0	NA
0.27	819	826	827	28.7	0.60%
0.86	8139	8131	8140	90.2	0.07%
4.34	209378	212058	204178	456.7	2.79%

## Regression Data (Zero-Forced)

Constant	0.0000
Std Err of Y Est	0.002
R Squared	0.999999
No. of Observations	4
Degrees of Freedom	3
X Coefficient(s)	9.504556E-03
Std Err of Coef.	4.825793E-06



Permeation Device: Metronics Dynacal H<sub>2</sub>S Rate: 3590 ng/ml @ 30C

Operator

*Sh6*

Date


*11-18-09*

*877*



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 8 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET8\_R1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:07:16 AM  
PRINTED: 11/17/2009 11:22:35 AM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.545	819
Totals		819

B 78

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 8 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709ical\H2S\_SET8\_R2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:12:16 AM  
PRINTED: 12/2/2009 4:40:10 PM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.518	826

Totals		826
--------	--	-----

179

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 8 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709ical\H2S\_SET8\_R3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:16:29 AM  
PRINTED: 12/2/2009 4:36:54 PM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.507	827
Totals		827

B 80

# Entech Engineering Inc. League City, Texas

SAMPLE ID: Dyna Set 3 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET3\_RR1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:47:31 AM  
PRINTED: 11/17/2009 11:51:13 AM  
USER: System



## FPD Results


Name	Retention Time	Area
H2S	2.490	8139

Totals		8139
--------	--	------

481

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 3 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET3\_RR2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:51:47 AM  
PRINTED: 11/17/2009 11:55:49 AM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.492	8131
Totals		8131

B82

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 3 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET3\_RR3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:56:15 AM  
PRINTED: 11/17/2009 11:59:39 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.495	8140

Totals		8140
--------	--	------

B83

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 1 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET1\_R1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 10:40:25 AM  
PRINTED: 11/17/2009 10:44:53 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.498	209378

Totals		209378
--------	--	--------

1684

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 1 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET1\_R4.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 10:56:14 AM  
PRINTED: 11/17/2009 10:59:47 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.498	212058
Totals		212058

B85



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: Dyna Set 1 @ Range 2, N2@1.5, 37C  
FILE: E:\GC DATA\2009\Calibration\GC#6\111709\H2S\_SET1\_R5.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 11:00:07 AM  
PRINTED: 11/17/2009 11:03:49 AM  
USER: System



FPD Results

Name	Retention Time	Area
H2S	2.498	204178
Totals		204178

# ENTECH ENGINEERING INC.

## Gas Chromatograph Analysis Standard Curve Data

GC #6 FPD ( Range 2)

H<sub>2</sub>S Standard - Dynacalibrator


11/18/09

(Retention Time = 2.49 min)

Concentration (ppmv)	Area 1 (AU)	Area 2 (AU)	Area 3 (AU)	Avg. Area (Square Root) (AU)	Precision Error (%)
0.00	0	0	0	0.0	NA
0.27	2600	2558	2578	50.8	0.90%
1.46	75391	75246	75444	274.5	0.10%
4.34	659483	666388	666240	814.9	0.30%

### Regression Data (Zero-Forced)

Constant	0.0000
Std Err of Y Est	0.001
R Squared	1.000000
No. of Observations	4
Degrees of Freedom	3
X Coefficient(s)	5.325114E-03
Std Err of Coef.	1.337038E-06




Permeation Device: Metronics Dynacal H<sub>2</sub>S Rate: 3590 ng/ml @ 30C

Operator Shy Date 11-18-09

887

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET8  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\SP1SET8R1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 10:21:04 AM  
PRINTED: 11/18/2009 10:43:53 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.545	2600

Totals		2600
--------	--	------

488

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET8  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\SP1SET8R2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 10:25:36 AM  
PRINTED: 11/18/2009 10:30:47 AM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.495	2558
Totals		2558

889

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET8  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\SP1SET8R3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 10:31:19 AM  
PRINTED: 11/18/2009 10:34:36 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.520	2578

Totals		2578
--------	--	------

B 90

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET2  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ica\SP1SET2R2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 9:52:15 AM  
PRINTED: 11/18/2009 10:03:39 AM  
USER: System




**FPD Results**

Name	Retention Time	Area
H2S	2.522	75246
Totals		75246

B91

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET2  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ica\SP1SET2R3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 9:48:13 AM  
PRINTED: 11/18/2009 10:04:41 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.513	75444

Totals		75444
--------	--	-------

B92

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET1  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ica\SP1SET1R2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 9:13:04 AM  
PRINTED: 11/18/2009 9:36:53 AM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.505	659483
Totals		659483



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: DYNACAL\_SP1SET1  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ica\SP1SET1R6.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 9:29:06 AM  
PRINTED: 11/18/2009 9:32:34 AM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.512	666388

Totals		666388
--------	--	--------

B94

# Entech Engineering Inc. League City, Texas

SAMPLE ID: DYNACAL\_SP1SET1  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\SP1SET1R7.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 9:32:57 AM  
PRINTED: 11/18/2009 9:37:19 AM  
USER: System



## FPD Results

Name	Retention Time	Area
H2S	2.512	666240

Totals		666240
--------	--	--------

B-95-

**Sulfur Analysis - GC / FPD**  
Post-Test Calibration  
(EPA method 15)

**ENTECH ENGINEERING INC.**

**P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118**

**3C #6 FPD Analysis - Initial Calibration: November 17, 2009**  
**H2S Standards Post Check Result - November 17, 2009**  
**(FPD@Range 2)**

		Area Counts	Concentration
Sample ID No.	GC Run No.	H2S	H2S
		area counts (sq. root)	0.86
Initial Cal. Standard Avg.		90.2	ppmv
Calibration Post-Check Run	1	91	0.86
	2	91	0.86
	3	91	0.86
		Average =	0.9
		Standard Response Factor =	104.88
		QA Response Factor =	105.57
		Mean Response Factor =	105.23
		Percent Difference, % =	0.65
		Pass/Fail Criterion (<5%) =	Pass

Operator


*SLY*

Date

*11-17-09**P 97*

# Entech Engineering Inc. League City, Texas

SAMPLE ID: SP1 SET3 POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\POSTCHK1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 6:57:04 PM  
PRINTED: 11/17/2009 7:01:30 PM  
USER: System



## FPD Results


Name	Retention Time	Area
H2S	2.500	8216

Totals		8216
--------	--	------

B98

# Entech Engineering Inc. League City, Texas

SAMPLE ID: SP1 SET3 POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\POSTCHK2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 6:51:58 PM  
PRINTED: 11/17/2009 6:56:17 PM  
USER: System



## FPD Results


Name	Retention Time	Area
H2S	2.505	8270

Totals		8270
--------	--	------

B99

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: SP1 SET3 POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\POSTCHK3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 7:01:57 PM  
PRINTED: 11/17/2009 7:06:22 PM  
USER: System



FPD Results

Name	Retention Time	Area
H2S	2.508	8241
Totals		8241

B100

**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**3C #6 FPD Analysis - Initial Calibration: November 18, 2009**  
**H2S Standards Post Check-1 Result - November 18, 2009**  
**(FPD@Range 2)**

		Area Counts	Concentration
Sample ID No.	GC Run No.	H2S	H2S
		area counts (sq. root)	1.46
Initial Cal. Standard Avg.		274.5	ppmv
Calibration Post-Check Run	1	276.7	1.47
	2	275.9	1.47
	3	276.1	1.47
Average =		1.5	
Standard Response Factor =		188.01	
QA Response Factor =		189.19	
Mean Response Factor =		188.60	
Percent Difference, % =		0.62	
Pass/Fail Criterion (<5%) =		Pass	


Operator Shy Date 11-18-09

*B101*



# Entech Engineering Inc. League City, Texas

SAMPLE ID: STD CHK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDCHK1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 2:45:09 PM  
PRINTED: 11/18/2009 2:49:22 PM  
USER: System



## FPD Results


Name	Retention Time	Area
H2S	2.718	76572

Totals		76572
--------	--	-------

6102

# Entech Engineering Inc. League City, Texas

SAMPLE ID: STD CHK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDCHK2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 2:33:48 PM  
PRINTED: 11/18/2009 2:39:36 PM  
USER: System




## FPD Results

Name	Retention Time	Area
H2S	2.735	76217
Totals		76217

B103

# Entech Engineering Inc. League City, Texas

SAMPLE ID: STD CHK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDCHK3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 2:39:54 PM  
PRINTED: 11/18/2009 2:44:43 PM  
USER: System



## FPD Results

Name	Retention Time	Area
H2S	2.598	76096
Totals		76096

B104

**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

GC #6 FPD Analysis - Initial Calibration: November 18, 2009

H2S Standards Post Check-2 Result - November 18, 2009

(FPD@Range 2)


		Area Counts	Concentration
Sample ID No.	GC Run No.	H2S area counts (sq. root)	H2S
Initial Cal. Standard Avg.		274.5	1.46 ppmv
Calibration Post-Check Run	1	275.2	1.47
	2	274.7	1.46
	3	274.6	1.46
Average =			1.46
Standard Response Factor =			188.01
QA Response Factor =			188.23
Mean Response Factor =			188.12
Percent Difference, % =			0.12
Pass/Fail Criterion (<5%) =			Pass

Operator Shy Date 11-18-09

B105

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: STD POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDPOSTCHK1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 6:41:52 PM  
PRINTED: 11/18/2009 6:46:31 PM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.552	75717

Totals		75717
--------	--	-------

4106

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: STD POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDPOSTCHK2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 6:47:05 PM  
PRINTED: 11/18/2009 6:52:38 PM  
USER: System



FPD Results


Name	Retention Time	Area
H2S	2.552	75433

Totals		75433
--------	--	-------

B107

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: STD POST CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\STDPOSTCHK3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 6:36:36 PM  
PRINTED: 11/18/2009 6:56:30 PM  
USER: System



FPD Results

Name	Retention Time	Area
H2S	2.552	75429

Totals		75429
--------	--	-------

*D108*

**Sulfur Analysis - GC / FPD**  
Sample Line Loss Check  
(EPA method 15)



**ENTECH ENGINEERING INC.**

P.O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Sample Line Loss Check (EPA Method 15) - November 17, 2009**

Initial Calibration Date: Nov. 17, 2009 - GC#6 FPD (Range2), Chromosil-310 Column


ID	X-Coefficient	Standard		Sample Line Feed		Sample Line Loss Recovery (%)	QC Limit (< 20%)
		Area	Conc. (ppmv)	Area	Conc. (ppmv)		
H2S	9.50E-03	22909	1.44	21523	1.39	-3.07	Pass

Operator Shy Date 11-17-09

B110

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S CHK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\H2SCHK.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 2:26:33 PM  
PRINTED: 11/17/2009 2:32:03 PM  
USER: System



## FPD Results


Name	Retention Time	Area
H2S	2.515	22909

Totals		22909
--------	--	-------

Bill

# Entech Engineering Inc. League City, Texas

SAMPLE ID: line loss check  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\linechk2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 7:47:06 PM  
PRINTED: 11/17/2009 7:50:34 PM  
USER: System



## FPD Results

Name	Retention Time	Area
H2S	2.502	21523
Totals		21523

112

**ENTECH ENGINEERING INC.**

P.O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Sample Line Loss Check (EPA Method 15) - November 18, 2009**

Initial Calibration Date: Nov. 18, 2009 - GC#6 FPD (Range2), Chromosil-310 Column


ID	X-Coefficient	Standard		Sample Line Feed		Sample Line Loss Recovery (%)	QC Limit (< 20%)
		Area	Conc. (ppmv)	Area	Conc. (ppmv)		
H2S	5.33E-03	58602	1.29	50560	1.20	-7.11	Pass

Operator Ans Date 11-18-09

B113

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S CHK  
FILE: E:\GC DATA\2009\Calibration\GC#6\111809ical\H2S CHK2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 10:39:15 AM  
PRINTED: 11/18/2009 10:42:25 AM  
USER: System




FPD Results

Name	Retention Time	Area
H2S	2.517	58602
	2.735	873
Totals		59475

B114

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: LINE LOSS CHECK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\LINE\_LOSS\_CHK.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/18/2009 6:57:16 PM  
PRINTED: 11/18/2009 7:00:27 PM  
USER: System



FPD Results

Name	Retention Time	Area
H2S	2.547	50560


Totals		50560
--------	--	-------

B115

**Sulfur Analysis - GC / FPD**  
Test -1  
Sample Results & Raw Data  
(EPA method 15)

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: BLANK CHK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\BLANK.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_5 min.met  
ACQUIRED: 11/17/2009 1:45:52 PM  
PRINTED: 11/17/2009 1:51:47 PM  
USER: System



FPD Results


Name	Retention Time	Area
------	----------------	------

B 117



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN1  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 3:08:33 PM  
PRINTED: 11/17/2009 3:20:39 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B118

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN2  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 3:20:55 PM  
PRINTED: 11/17/2009 3:33:07 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B119

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN3  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 3:35:27 PM  
PRINTED: 11/17/2009 3:48:33 PM  
USER: System



FPD Results


Name	Retention Time	Area
	5.336	837
	7.361	928
	11.901	1497

Totals		3262
--------	--	------

B120

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN4  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R4.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 3:48:59 PM  
PRINTED: 11/17/2009 4:01:05 PM  
USER: System



FPD Results


Name	Retention Time	Area
	6.326	890

Totals		890
--------	--	-----

B121

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN5  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R5.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 4:01:36 PM  
PRINTED: 11/17/2009 4:13:41 PM  
USER: System



FPD Results


Name	Retention Time	Area
	1.600	809
	9.141	1801
	9.301	978

Totals		3588
--------	--	------

B122

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN6  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R6.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 4:14:07 PM  
PRINTED: 11/17/2009 4:26:13 PM  
USER: System



FPD Results


Name	Retention Time	Area
	10.323	1933
	10.468	1620
	10.838	2339
	11.118	1158

Totals		7050
--------	--	------

2123

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN7  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R7.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 4:26:43 PM  
PRINTED: 11/17/2009 4:38:54 PM  
USER: System



FPD Results


Name	Retention Time	Area
	0.523	1386
	0.808	1002
	9.793	914
	10.418	888
	11.311	830

Totals		5020
--------	--	------

B124

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 111709 TEST1 RUN8  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R8.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 4:39:17 PM  
PRINTED: 11/17/2009 4:51:30 PM  
USER: System



## FPD Results

Name	Retention Time	Area
	8.453	1206
	9.633	1025
	10.044	841
	10.168	1157
	11.026	1595


Totals		5824
--------	--	------

B125



# Entech Engineering Inc. League City, Texas

SAMPLE ID: 111709 TEST1 RUN9  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R9.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 4:51:55 PM  
PRINTED: 11/17/2009 5:04:09 PM  
USER: System




## FPD Results

Name	Retention Time	Area
	11.119	887
Totals		887

B126

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN10  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R10.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 5:04:34 PM  
PRINTED: 11/17/2009 5:16:39 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

5127

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN11  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R11.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 5:17:17 PM  
PRINTED: 11/17/2009 5:29:24 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B128

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN12  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R12.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 5:29:55 PM  
PRINTED: 11/17/2009 5:42:00 PM  
USER: System




FPD Results

Name	Retention Time	Area
	1.618	814
Totals		814

B129

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN13  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R13.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 5:43:47 PM  
PRINTED: 11/17/2009 5:55:53 PM  
USER: System



FPD Results


Name	Retention Time	Area
	1.088	816
	1.642	879
	5.768	859
	8.398	971

Totals		3525
--------	--	------

B130

# Entech Engineering Inc. League City, Texas

SAMPLE ID: 111709 TEST1 RUN14  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R14.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 5:57:03 PM  
PRINTED: 11/17/2009 6:09:07 PM  
USER: System



## FPD Results


Name	Retention Time	Area
	0.928	833
	10.238	1265
	10.628	930

Totals		3028
--------	--	------

B131

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN15  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R15.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 6:09:32 PM  
PRINTED: 11/17/2009 6:21:55 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B-132

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: 111709 TEST1 RUN16  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R16.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 6:22:21 PM  
PRINTED: 11/17/2009 6:33:57 PM  
USER: System



FPD Results


Name	Retention Time	Area
------	----------------	------

*B133*



# Entech Engineering Inc. League City, Texas

SAMPLE ID: 111709 TEST1 RUN17  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111709\T1R17.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/17/2009 6:34:26 PM  
PRINTED: 11/17/2009 6:44:22 PM  
USER: System



## FPD Results

Name	Retention Time	Area
	0.690	1590
	1.600	835
	5.583	839

Totals		3264
--------	--	------


B134

**Sulfur Analysis - GC / FPD**  
Test - 2  
Sample Results & Raw Data  
(EPA method 15)

*B-135*

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: BLANK  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\BLANK.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 10:45:39 AM  
PRINTED: 11/18/2009 10:49:47 AM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B-136

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN1  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 11:15:57 AM  
PRINTED: 11/18/2009 11:25:16 AM  
USER: System




## FPD Results

Name	Retention Time	Area
	3.321	6727
Totals		6727

B137

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN2  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 11:28:19 AM  
PRINTED: 11/18/2009 11:34:24 AM  
USER: System



## FPD Results


Name	Retention Time	Area
	0.703	3142
	1.048	3150
	3.716	5197

Totals		11489
--------	--	-------

B-138

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN3  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 11:40:22 AM  
PRINTED: 11/18/2009 11:46:38 AM  
USER: System




## FPD Results

Name	Retention Time	Area
	0.510	3248
	0.583	2576

Totals		5824
--------	--	------

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN4  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN4.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 11:52:25 AM  
PRINTED: 11/18/2009 12:07:31 PM  
USER: System



FPD Results


Name	Retention Time	Area
	3.381	4855
	3.585	3335

Totals		8190
--------	--	------

B140

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN5  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN5.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 12:07:49 PM  
PRINTED: 11/18/2009 12:14:30 PM  
USER: System



## FPD Results


Name	Retention Time	Area
	0.527	2864
	0.587	4519
	0.850	5412
Totals		12795

B-141



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN6  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN6.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 12:20:24 PM  
PRINTED: 11/18/2009 12:30:06 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B-142

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN7  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN7.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 12:32:29 PM  
PRINTED: 11/18/2009 12:39:28 PM  
USER: System



## FPD Results


Name	Retention Time	Area
	0.095	4215
	0.195	4743
	0.325	4452

Totals		13410
--------	--	-------

B143

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN8  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN8.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 12:44:07 PM  
PRINTED: 11/18/2009 12:50:42 PM  
USER: System



## FPD Results


Name	Retention Time	Area
	0.105	2613
	1.083	6282
	1.292	4084

Totals		12979
--------	--	-------

B144

# Entech Engineering Inc. League City, Texas

SAMPLE ID: H2S TEST2 RUN9  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN9.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 12:56:16 PM  
PRINTED: 11/18/2009 1:04:06 PM  
USER: System




## FPD Results

Name	Retention Time	Area
	3.760	6123
	3.855	3018
Totals		9141

B145

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN10  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN10.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 1:09:13 PM  
PRINTED: 11/18/2009 1:21:41 PM  
USER: System



FPD Results


Name	Retention Time	Area
	3.083	2973
	3.446	2931

Totals		5904
--------	--	------

B146

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN11  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN11.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 1:22:00 PM  
PRINTED: 11/18/2009 1:27:37 PM  
USER: System



FPD Results


Name	Retention Time	Area
	0.295	3369

Totals		3369
--------	--	------

B147

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN12  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN12.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 1:34:03 PM  
PRINTED: 11/18/2009 1:40:07 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B148

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN13  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN13.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 1:46:21 PM  
PRINTED: 11/18/2009 1:52:45 PM  
USER: System



FPD Results


Name	Retention Time	Area
------	----------------	------

B149



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN14  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN14.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 1:58:27 PM  
PRINTED: 11/18/2009 2:05:56 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B-150

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN15  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN15.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 2:10:30 PM  
PRINTED: 11/18/2009 2:15:50 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B151

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: H2S TEST2 RUN16  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\TEST2RUN16.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 2:22:34 PM  
PRINTED: 11/18/2009 2:28:10 PM  
USER: System



FPD Results

Name	Retention Time	Area
------	----------------	------


B152

**Sulfur Analysis - GC / FPD**  
Test - 3  
Sample Results & Raw Data  
(EPA method 15)

*B-153*

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN1  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN1.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 3:15:04 PM  
PRINTED: 11/18/2009 3:21:28 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B154

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN2  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN2.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 3:27:16 PM  
PRINTED: 11/18/2009 3:33:01 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B155

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN3  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN3.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 3:39:20 PM  
PRINTED: 11/18/2009 3:47:56 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B156

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN4  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN4.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 3:51:31 PM  
PRINTED: 11/18/2009 4:03:54 PM  
USER: System



FPD Results


Name	Retention Time	Area
------	----------------	------

B-157



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN5  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN5.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 4:04:11 PM  
PRINTED: 11/18/2009 4:10:08 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B158

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN6  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN6.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 4:16:15 PM  
PRINTED: 11/18/2009 4:23:02 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

2159

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN7  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN7.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 4:28:26 PM  
PRINTED: 11/18/2009 4:34:49 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B160

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN8  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN8.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 4:40:34 PM  
PRINTED: 11/18/2009 4:47:30 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

4161

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN9  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN9.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 4:52:40 PM  
PRINTED: 11/18/2009 4:59:55 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B162

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN10  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN10.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 5:04:44 PM  
PRINTED: 11/18/2009 5:10:46 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

*B163*

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN11  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN11.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 5:16:47 PM  
PRINTED: 11/18/2009 5:22:10 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

15164

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN12  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN12.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 5:28:50 PM  
PRINTED: 11/18/2009 5:34:45 PM  
USER: System



FPD Results


Name	Retention Time	Area
------	----------------	------

B165



**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN13  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN13.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 5:40:55 PM  
PRINTED: 11/18/2009 5:47:36 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B-166

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN14  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN14.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 5:53:00 PM  
PRINTED: 11/18/2009 6:02:12 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B167

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN15  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN15.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 6:05:04 PM  
PRINTED: 11/18/2009 6:11:06 PM  
USER: System




FPD Results

Name	Retention Time	Area
------	----------------	------

B168

**Entech Engineering Inc. League City, Texas**

SAMPLE ID: TEST-3 RUN16  
FILE: E:\GC DATA\2009\Sample\GC#6\ConocoPhillips\111809\T3RUN16.dat  
METHOD: E:\GC DATA\2009\Calibration\GC#6\FPD\_12MIN.met  
ACQUIRED: 11/18/2009 6:17:08 PM  
PRINTED: 11/18/2009 6:23:08 PM  
USER: System



FPD Results

Name	Retention Time	Area
------	----------------	------

B169

**ENTECH ENGINEERING INC.**

**P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118**

---

**APPENDIX C.**

**EXAMPLE CALCULATIONS AND QA/QC DATA**

**EPA METHOD 2 & 4  
FLOW AND MOISTURE DETERMINATION**

Company	Conoco Phillips
Location	Sweeny, TX
Unit Tested	Unit 39.1 TGI (EPN 39.1-95-118)
Test Type	RM 4
<b>FUEL PARAMETERS</b>	
Type of Fuel Used	N/A
<b>TEST CONDITIONS</b>	
Test Number	1
Test Date	11/17/09
Test Site	Stack
Unit Load	Full
Sampling Elevation (feet)	150.00
Barometric Pressure, Uncorrected (inches Hg)	30.06
Barometric Pressure, Corrected (inches Hg)	29.91
Stack Internal Diameter (inches)	78.000
Stack Area at Sampling Plane (ft <sup>2</sup> )	33.18307
<b>SAMPLING RESULTS</b>	
Nozzle Diameter (inches)	0.470
Nozzle Area (ft <sup>2</sup> )	0.00120
Pitot Tube Coefficient (C <sub>p</sub> )	0.84
Control Box Delta H Standard	1.89
Dry Gas Meter Correction Factor	1.006
Average Delta H (inches H <sub>2</sub> O)	1.514
Average Square Root Delta P (inches H <sub>2</sub> O <sup>1/2</sup> )	0.2558
Average Meter Temperature (degrees F)	67.8
Sampling Time (minutes)	60
Meter Volume, Uncorrected (ft <sup>3</sup> )	40.489
Meter Volume, Corrected (ft <sup>3</sup> )	40.732
Meter Volume at Standard Conditions (DSCF)	40.895
Moisture Measured by Weight (g)	77.4
<b>STACK GAS PARAMETERS</b>	
Excess Oxygen (vol% dry)	2.71
Carbon Dioxide (vol% dry)	4.30
Temperature (degrees F)	547.5
Molecular Weight, dry	28.80
Molecular Weight, wet	27.91
Moisture Content (vol%)	8.19
Static Head (inches H <sub>2</sub> O)	0.10
Stack Pressure (inches Hg)	29.91
Average Velocity (ft/sec)	20.17
Actual Volumetric Flow Rate (ACFM)	40163.19
Volumetric Flow Rate (DSCFM)	19326.95
Average Isokinetic Sampling Rate (%)	98.15

C1

**EPA METHOD 2 & 4  
FLOW AND MOISTURE DETERMINATION**

Company	Conoco Phillips
Location	Sweeny, TX
Unit Tested	Unit 39.1 TGI (EPN 39.1-95-118)
Test Type	RM 4
<b>FUEL PARAMETERS</b>	
Type of Fuel Used	N/A
<b>TEST CONDITIONS</b>	
Test Number	2
Test Date	11/18/09
Test Site	Stack
Unit Load	Full
Sampling Elevation (feet)	150.00
Barometric Pressure, Uncorrected (inches Hg)	30.04
Barometric Pressure, Corrected (inches Hg)	29.89
Stack Internal Diameter (inches)	78.000
Stack Area at Sampling Plane (ft <sup>2</sup> )	33.18307
<b>SAMPLING RESULTS</b>	
Nozzle Diameter (inches)	0.470
Nozzle Area (ft <sup>2</sup> )	0.00120
Pitot Tube Coefficient (C <sub>p</sub> )	0.84
Control Box Delta H Standard	1.89
Dry Gas Meter Correction Factor	1.006
Average Delta H (inches H <sub>2</sub> O)	1.428
Average Square Root Delta P (inches H <sub>2</sub> O <sup>1/2</sup> )	0.2579
Average Meter Temperature (degrees F)	71.9
Sampling Time (minutes)	60
Meter Volume, Uncorrected (ft <sup>3</sup> )	38.555
Meter Volume, Corrected (ft <sup>3</sup> )	38.786
Meter Volume at Standard Conditions (DSCF)	38.603
Moisture Measured by Weight (g)	92.5
<b>STACK GAS PARAMETERS</b>	
Excess Oxygen (vol% dry)	3.00
Carbon Dioxide (vol% dry)	4.43
Temperature (degrees F)	548.7
Molecular Weight, dry	28.83
Molecular Weight, wet	27.73
Moisture Content (vol%)	10.15
Static Head (inches H <sub>2</sub> O)	0.10
Stack Pressure (inches Hg)	29.89
Average Velocity (ft/sec)	20.43
Actual Volumetric Flow Rate (ACFM)	40670.48
Volumetric Flow Rate (DSCFM)	19118.60
Average Isokinetic Sampling Rate (%)	91.76

**EPA METHOD 2 & 4  
FLOW AND MOISTURE DETERMINATION**

Company	Conoco Phillips
Location	Sweeny, TX
Unit Tested	Unit 39.1 TGI (EPN 39.1-95-118)
Test Type	RM 4
<b>FUEL PARAMETERS</b>	
Type of Fuel Used	N/A
<b>TEST CONDITIONS</b>	
Test Number	3
Test Date	11/18/09
Test Site	Stack
Unit Load	Full
Sampling Elevation (feet)	150.00
Barometric Pressure, Uncorrected (inches Hg)	30.04
Barometric Pressure, Corrected (inches Hg)	29.89
Stack Internal Diameter (inches)	78.000
Stack Area at Sampling Plane (ft <sup>2</sup> )	33.18307
<b>SAMPLING RESULTS</b>	
Nozzle Diameter (inches)	0.470
Nozzle Area (ft <sup>2</sup> )	0.00120
Pitot Tube Coefficient (C <sub>p</sub> )	0.84
Control Box Delta H Standard	1.89
Dry Gas Meter Correction Factor	1.006
Average Delta H (inches H <sub>2</sub> O)	1.738
Average Square Root Delta P (inches H <sub>2</sub> O <sup>1/2</sup> )	0.2564
Average Meter Temperature (degrees F)	85.8
Sampling Time (minutes)	60
Meter Volume, Uncorrected (ft <sup>3</sup> )	41.965
Meter Volume, Corrected (ft <sup>3</sup> )	42.217
Meter Volume at Standard Conditions (DSCF)	40.982
Moisture Measured by Weight (g)	93.5
<b>STACK GAS PARAMETERS</b>	
Excess Oxygen (vol% dry)	1.85
Carbon Dioxide (vol% dry)	4.47
Temperature (degrees F)	546.3
Molecular Weight, dry	28.79
Molecular Weight, wet	27.74
Moisture Content (vol%)	9.71
Static Head (inches H <sub>2</sub> O)	0.10
Stack Pressure (inches Hg)	29.89
Average Velocity (ft/sec)	20.28
Actual Volumetric Flow Rate (ACFM)	40375.03
Volumetric Flow Rate (DSCFM)	19116.66
Average Isokinetic Sampling Rate (%)	97.42



FLUE GAS MOISTURE CONTENT  
(EPA Reference Method 4)  
Test 1

Nomenclature:

- $B_{ws}$  = Proportion of water vapor, by volume, in the gas stream.  
 $M_w$  = Molecular weight of water, 18.0 lb/lb mole.  
 $P_{bar}$  = Barometric pressure at dry gas meter, in Hg.  
 $\Delta H$  = Average pressure differential in dry gas meter, in H<sub>2</sub>O.  
 $P_m$  =  $P_{bar} + (H/13.6)$   
13.6 = Specific gravity of Hg.  
 $P_{std}$  = Standard absolute pressure, 29.92 in Hg.  
 $T_{mf}$  = Temperature at meter, F.  
 $T_m$  = Absolute temperature at meter, R.  
 $V_m$  = Dry gas volume measure by dry gas meter, dcf (dry cubic feet).  
 $V_{mstd}$  = Dry gas volume measured by dry gas meter corrected to standard conditions, dscf.  
 $V_{wcstd}$  = Volume of water vapor collected, volumetrically measured, corrected to standard conditions, scf.  
 $V_{wgstd}$  = Volume of water vapor collected, gravimetrically measured, corrected to standard conditions, scf.  
 $V_w$  = Volume of water collected, ml.  
 $W_g$  = Weight of water collected, g.  
 $Y$  = Dry gas meter calibration factor.  
 $K1$  = Constant, 0.04707 cubic feet/ml.  
 $K2$  = Constant, 0.04715 cubic feet/g.  
 $K3$  = Constant, 17.64 R/in Hg.

Variables:

$$\Delta H := 1.514 \cdot \text{in\_H}_2\text{O}$$

$$P_{bar} := 29.91 \cdot \text{in\_Hg}$$

$$T_{mf} := 67.8 \cdot \text{F}$$

$$V_m := 40.489 \cdot \text{ft}^3$$

$$V_w := 0 \cdot \text{ml}$$

$$W_g := 77.4 \cdot \text{g}$$

$$Y := 1.006$$

Constants:

$$M_w := 18.0 \cdot \frac{\text{lb}}{\text{lb} \cdot \text{mole}}$$

$$P_{std} := 29.92 \cdot \text{in\_Hg}$$

$$K1 := 0.04707 \cdot \frac{\text{ft}^3}{\text{ml}}$$

$$K2 := 0.04715 \cdot \frac{\text{ft}^3}{\text{g}}$$

$$K3 := 17.64 \cdot \frac{\text{R}}{\text{in\_Hg}}$$

Calculations:

$$P_m := P_{\text{bar}} + \left( \frac{\Delta H \cdot \text{in\_Hg}}{13.6 \text{ in\_H2O}} \right)$$

$$P_m = 30.021 \cdot \text{in\_Hg}$$

$$T_m := T_{\text{mf}} + 460$$

$$T_m = 527.8 \cdot \text{R}$$

$$V_{\text{wc\_std}} := K1 \cdot V_w \quad (\text{Equation 4.1})$$

$$V_{\text{wc\_std}} = 0.000 \cdot \text{ft}^3$$

$$V_{\text{wg\_std}} := K2 \cdot W_g \quad (\text{Equation 4.2})$$

$$V_{\text{wg\_std}} = 3.649 \cdot \text{ft}^3$$

$$V_{\text{m\_std}} := K3 \cdot Y \cdot \left( \frac{V_m \cdot P_m}{T_m} \right) \quad (\text{Equation 4.3})$$

$$V_{\text{m\_std}} = 40.869 \cdot \text{ft}^3$$

$$B_{\text{ws}} := \frac{V_{\text{wc\_std}} + V_{\text{wg\_std}}}{V_{\text{wc\_std}} + V_{\text{wg\_std}} + V_{\text{m\_std}}} \quad (\text{Equation 4.4})$$

$$B_{\text{ws}} = 0.0820$$

$$\text{Moisture} := B_{\text{ws}} \cdot 100 \cdot \%$$

$$\text{Moisture} = 8.20 \cdot \%$$

MOLECULAR WEIGHT OF FLUE GAS  
(EPA Reference Method 3)

Nomenclature:

$M_d$  = Dry molecular weight, lb/lb-mole.

$M_s$  = Wet molecular weight, lb/lb-mole.

$CO_2$  = Percent  $CO_2$  by volume, dry basis.

$O_2$  = Percent  $O_2$  by volume, dry basis.

$CO$  = Percent  $CO$  by volume, dry basis.

$N_2$  = Percent  $N_2$  by volume, dry basis.

Balance =  $N_2 + CO$ .

$B_{ws}$  = Flue gas moisture fraction, by volume.

Variables:

$$CO_2 := 4.30 \cdot \%$$

$$O_2 := 2.71 \cdot \%$$

$$\text{Balance} := 100 \cdot \% - CO_2 - O_2$$

$$\text{Balance} = 92.99 \cdot \%$$

$$B_{ws} = 0.0820$$

Constants:

$$44.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \quad \text{Molecular Weight of } CO_2.$$

$$32.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \quad \text{Molecular Weight of } O_2.$$

$$28.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \quad \text{Molecular Weight of } N_2.$$

Calculations:

$$M_d := \left( 44.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \cdot CO_2 \right) + \left( 32.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \cdot O_2 \right) + \left( 28.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \cdot \text{Balance} \right) \quad (\text{Equation 3-1})$$

$$M_d = 28.80 \cdot \frac{\text{lb}}{\text{lb-mole}}$$

$$M_s := M_d \cdot (1.0 - B_{ws}) + \left( 18.0 \cdot \frac{\text{lb}}{\text{lb-mole}} \cdot B_{ws} \right) \quad (\text{EPA Reference Method 2, Equation 2-5})$$

$$M_s = 27.91 \cdot \frac{\text{lb}}{\text{lb-mole}}$$

VELOCITY & VOLUMETRIC FLOW RATE OF FLUE GAS  
(EPA Reference Method 2)

Nomenclature:

Diameter = Diameter of stack, in.

Area = Cross-sectional area of stack, ft<sup>2</sup>.

B<sub>ws</sub> = Flue gas moisture fraction.

C<sub>p</sub> = Pitot tube coefficient.

K<sub>p</sub> = Pitot tube Constant.

M<sub>s</sub> = Molecular weight of stack gas, wet basis, lb/lb-mole.

P<sub>bar</sub> = Barometric pressure at measurement site, in\_Hg.

P<sub>g</sub> = Stack static pressure, in\_H2O.

P<sub>s</sub> = Absolute stack pressure, in\_Hg.

P<sub>std</sub> = Standard pressure, 29.92 in\_Hg.

Q<sub>std</sub> = Dry volumetric stack gas flow rate, corrected to standard conditions, dscf/min.

T<sub>sf</sub> = Stack temperature, F.

T<sub>s</sub> = Absolute stack temperature, R.

T<sub>std</sub> = Standard absolute temperature, 528 R.

V = Average stack gas velocity, ft/sec.

Δp = Average square root of delta p, in\_H2O<sup>0.5</sup>.

Variables:

$$\text{Area} := 33.18307 \cdot \text{ft}^2$$

$$C_p := 0.84$$

$$M_s = 27.91 \cdot \frac{\text{lb}}{\text{lb} \cdot \text{mole}}$$

$$P_{\text{bar}} = 29.91 \cdot \text{in}_\text{Hg}$$

$$P_g := 0.01 \cdot \text{in}_\text{H2O}$$

$$T_{\text{sf}} := 547.5 \cdot \text{F}$$

$$\Delta p := 0.2558 \cdot \text{in}_\text{H2O}^{0.5}$$

$$B_{\text{ws}} = 0.0820$$

Constants:

$$K_p := 85.49 \cdot \frac{\text{ft}}{\text{sec}} \cdot \left( \frac{\frac{\text{lb}}{\text{lb} \cdot \text{mole}} \cdot \text{in}_\text{Hg}}{\text{R} \cdot \text{in}_\text{H2O}} \right)^{0.5}$$

$$T_{\text{std}} := 528 \cdot \text{R}$$

$$P_{\text{std}} := 29.92 \cdot \text{in}_\text{Hg}$$

Calculations:

$$T_s := T_{sf} + 460 \cdot R$$

$$T_s = 1.008 \cdot 10^3 \cdot R$$

$$P_s := P_{bar} + \left( \frac{P_g \cdot \text{in\_Hg}}{13.6 \text{ in\_H2O}} \right)$$

$$P_s = 29.911 \cdot \text{in\_Hg}$$

$$V := K_p \cdot C_p \cdot \Delta p \cdot \sqrt{\frac{T_s}{P_s \cdot M_s}} \quad (\text{Equation 2-9})$$

$$V = 20.18 \cdot \frac{\text{ft}}{\text{sec}}$$

$$Q_{std} := 60 \cdot \frac{\text{sec}}{\text{min}} \cdot (1.0 - B_{ws}) \cdot V \cdot \text{Area} \cdot \left( \frac{T_{std}}{T_s} \right) \cdot \left( \frac{P_s}{P_{std}} \right) \quad (\text{Equation 2-10})$$

$$Q_{std} = 19323.69 \cdot \frac{\text{dscf}}{\text{min}}$$

# ISOKINETIC SAMPLING RATE (EPA Reference Method 5)

## Nomenclature:

$A_n$  = Cross-sectional area of nozzle tip, ft<sup>2</sup>.

$B_{ws}$  = Flue gas moisture content.

$I$  = Percent of isokinetic sampling, %.

$K4$  = Constant

$N_d$  = Diameter of nozzle tip, in.

$N_r$  = Radius of nozzle tip, ft.

$P_s$  = Absolute stack pressure

$P_{std}$  = Standard pressure, 29.92 in\_Hg

$T_{sf}$  = Stack temperature, F

$T_s$  = Absolute stack temperature, R

$Vm_{std}$  = Dry gas meter volume at standard conditions, dscf.

$V$  = Velocity of flue gas, ft/sec.

Time = Total test time, min.

## Variables:

$$N_d := 0.470 \cdot \text{in}$$

$$B_{ws} = 0.0820$$

$$P_s = 29.91 \cdot \text{in\_Hg}$$

$$T_{sf} = 547.5 \cdot \text{F}$$

$$T_s = 1.008 \cdot 10^3 \cdot \text{R}$$

$$Vm_{std} = 40.87 \cdot \text{dscf}$$

$$V = 20.18 \cdot \frac{\text{ft}}{\text{sec}}$$

$$\text{Time} := 60 \cdot \text{min}$$

## Constants:

$$K4 := 0.09450 \cdot \frac{\text{in\_Hg} \cdot \% \cdot \text{min}}{\text{sec} \cdot \text{R}}$$

$$P_{std} = 29.92 \cdot \text{in\_Hg}$$

$$T_{std} = 528 \cdot \text{R}$$

Calculations:

$$Nr := \frac{Nd}{2}$$

$$Nr = 0.235 \cdot \text{in}$$

$$An := \pi \cdot \left( \frac{Nr}{12 \cdot \frac{\text{in}}{\text{ft}}} \right)^2$$

$$An = 0.00120482 \cdot \text{ft}^2$$

$$I := \frac{K_4 \cdot T_s \cdot Vm_{std}}{P_s \cdot V \cdot An \cdot \text{Time} \cdot (1.0 - B_{ws})} \quad (\text{Equation 5-8})$$

$$I = 97.1 \cdot \%$$

CO Emissions in lb/hr  
EPA Reference Method 10 Test I

Definitions:

$CO_{conc}$  = CO concentration, dry ppmv.

$CO_{em}$  = CO emission rate, lb/hr.

$Q_{std}$  = Volumetric flow rate, DSCFM

Variables:

Constants:

$$CO_{conc} := 55.49 \cdot \text{ppmv}$$

$$MW_{CO} := 28.01 \cdot \frac{\text{lb}}{\text{lb} \cdot \text{mole}}$$

$$Q_{std} := 19326.95 \cdot \frac{\text{dscf}}{\text{min}}$$

Calculations:

$$CO_{em} := CO_{conc} \cdot \left( 60 \cdot \frac{\text{min}}{\text{hr}} \right) \cdot Q_{std} \cdot \left[ \frac{MW_{CO}}{385.33 \cdot \frac{\text{dscf}}{\text{lb} \cdot \text{mole}} \cdot (10^6 \cdot \text{ppmv})} \right]$$

$$CO_{em} = 4.68 \cdot \frac{\text{lb}}{\text{hr}}$$



NOx Emissions in lb/hr  
EPA Reference Method 7E Test 1

Definitions:

$\text{NOx}_{\text{conc}}$  = NOx concentration, dry ppmv.

$\text{NOx}_{\text{em}}$  = NOx emission rate, lb/hr.

$Q_{\text{std}}$  = volumetric flow rate, DSCFM

Variables:

$$\text{NOx}_{\text{conc}} := 13.98 \cdot \text{ppmv}$$

$$Q_{\text{std}} := 19326.95 \cdot \frac{\text{dscf}}{\text{min}}$$

Constants:

$$\text{MW}_{\text{NOx}} := 46.00 \cdot \frac{\text{lb}}{\text{lb-mole}}$$

Calculations:

$$\text{NOx}_{\text{em}} := \text{NOx}_{\text{conc}} \cdot \left( 60 \cdot \frac{\text{min}}{\text{hr}} \right) \cdot Q_{\text{std}} \cdot \left[ \frac{\text{MW}_{\text{NOx}}}{385.33 \cdot \frac{\text{dscf}}{\text{lb-mole}} \cdot (10^6 \cdot \text{ppmv})} \right]$$

$$\text{NOx}_{\text{em}} = 1.94 \cdot \frac{\text{lb}}{\text{hr}}$$

***ENTECH ENGINEERING INC.***

***P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118***

---

**APPENDIX D.**

**INSTRUMENT SPECIFICATIONS**

**Shimadzu  
Model 14A  
Gas Chromatograph  
Specifications**

**2. Specifications**

**GC1-C10893500574**

■ **Column oven**

Dimensions of column compartment: 230 (W) X 140 (D) X 360 (H) mm

Length of columns to be accommodated:	Stainless steel column	10m X 2
	Glass column	3m X 2
	Capillary column	100m X 1

Range of temperature setting:	Temperature;	-80~+399	1°C step
	Rate of temperature rise;	0~40°C/min	0.1°C step
	Constant temperature hold time;	0~655 min	0.1°C step
	Program stages	5 stages max.	

Range of temperature control (with power voltage of 100V)

Range of linear temperature increase:	30°C/min	150°C or less
	20°C/min	250°C or less
	10°C/min	330°C or less
	5°C/min	399°C or less

Lower-limit temperature: When INJ and DET temperature is 300°C,  
room temperature + 15°C

When INJ and DET temperature is 150°C,  
room temperature + 10°C

Additional cryogenic equipment is required for controlling at lower temperatures than those above.

Cooling speed: Approx. 9 min to reduce from 399°C to 100°C with room temperature of 25°C.

Approx. 14 min for reducing from 399°C to 50°C with room temperature of 25°C

■ **Detector oven**

Range of temperature setting: Room temperature ~ 399°C (in 1°C steps)

■ **TCD oven**

Range of temperature setting: Room temperature ~ 399°C (in 1°C steps)

■ **Sample injection port unit**

Range of temperature setting: Room temperature ~ 399°C

Sample injection port unit (Either one of the following units is provided)

Single injection port unit:	For packed glass column, combination type of glass insert and on-column injection.	1 flow line
Dual injection port unit:	For packed glass column, combination of glass insert and on column injection.	2 flow lines
Injection port unit for for capillary columns:	Exclusive injection port for capillary analysis	

■ Overheat protection circuit      3 circuits

1. 450°C fixed independent protection circuit.
2. Protection circuit for which the upper-limit temperature can be set via key operation.
3. Overheat protection circuit by CPU abnormality detection

■ Combination of Detectors


1. Four detectors at maximum from among TCD, FID, ECD, FPD, and FTD can be simultaneously installed to the detector oven of the GC main body. Two TCD detectors cannot be set simultaneously.
2. Only two types of detector controllers (except FTD) can be installed simultaneously to the control section of the GC. Another unit should be applied for installing three or more detector controllers, or FTD.

MODEL 721M

# SO<sub>2</sub>, O<sub>2</sub> Analyzer

## APPLICATIONS

- ✓ Sulfur recovery incinerators
- ✓ Coal fired power plant stack sources
- ✓ Cement kilns
- ✓ Mineral smelters
- ✓ Pulp and paper



## SPECIFICATIONS

Methodology	Non-dispersive ultraviolet
Speed of response	90% in less than 30 seconds
Accuracy	±0.5% of full-scale (typically)
Zero drift	±0.5% of full-scale (24 hours)
Calibration	SO <sub>2</sub> calibration gas in nitrogen
Sensitivity	±0.5% of full-scale
Outputs	100 mVDC, 1 VDC, 10 VDC field selectable as standard. Optional 4 to 20 mA (isolated) Self-powered standard. loop-power available upon request.
Electrical requirement	50 W: 120 VAC/50 to 60 Hz or 220 VAC/50 to 60 Hz
Electrical classification	General Purpose
Typical flow	1 to 2 L/min. (2 to 5 SCF/H)
Ambient temperature	10 to 35°C (50 to 95°F)
Temperature stability	0.25 ppm per Celsius degree
Electrical classification	General purpose (non-hazardous)
Physical dimensions	HxWxD 180x480x600 mm (7x19x23.5 in.) Weight: 12.3 kg (27 lb.)

## STANDARD FEATURES

- Outputs**
- 4 voltage outputs any combination of 0 to 100 mVDC, 0 to 1 VDC, 0 to 10 VDC

- Electrical requirements**
- 120 VAC/50 to 60 Hz

- Measurement ranges: SO<sub>2</sub>**
- 0 to 100 ppm minimum up to 0 to 5000 ppm maximum (see note 1)
  - Sample must be dry, i.e. dewpoint of the sample must be below the lowest ambient temperature at which the analyzer will be used.

Note 1: Standard cell configuration - others available

## OPTIONS

- Outputs**
- 4 to 20 mA self-powered
  - 4 current outputs (maximum 4)
  - 4 to 20 mA loop-powered (available on request in place of self-powered)

- Electrical requirements:**
- 240 VAC/50 to 60 Hz

- Optional configurations**
- Low range concentration cell for full-scale ranges less than 100 ppm
  - Mid-range concentration cell for applications requiring full-scale ranges between 5000 ppm and 5%
  - High concentration cell for applications requiring full-scale ranges between 5% and 10%
  - O<sub>2</sub> sensor installed in the Model 721M enclosure
  - O<sub>2</sub> compensation
  - Mass flow calculations

The information contained on this information sheet is subject to change without notice. BOVAR reserves the right to correct any errors.



CANADA • 8 Manning Close N.E., Calgary, Alberta T2E 7N5 Phone: (403) 235-8300 Fax: (403) 248-3550  
 UNITED STATES • Suite 150, 10200 Richmond Avenue, Houston, Texas, USA 77042 Phone: (713) 789-1084 Fax: (713) 784-1015  
 EUROPE • Voltastrasse 7, D-65795 Hattersheim 1, Germany Phone: (49) 6190-8591 Fax: (49) 6190-73660  
 ASIA / PACIFIC • Level 28, Menara Haw Par, Jalan Sultan Ismail, 50250 Kuala Lumpur, Malaysia Phone: 60-3-233-6121 Fax: 60-3-233-6222




---

# MODEL 42C

Chemiluminescence NO-NO<sub>2</sub>-NO<sub>x</sub> Analyzer

---



---


## "C" SERIES TECHNOLOGY PROVIDING:

---

- Sub Parts Per Billion (PPB) Detection
- Electronic Diagnostic Transducers
- Multi-Line Alpha Numeric Display
- Dedicated Communications Processor
- Remote Performance Diagnostics
- U.S. EPA Designated Method (RFNA-1289-074)

## FLEXIBLE COMMUNICATIONS

- Bidirectional Addressable RS-232 Communication Port
- Analog Data Outputs with Selectable Voltages
- Isolated Current Outputs (optional)
- Analog Status Outputs (optional)
- Instrument Diagnostics, local and remote



## MODEL 42C SPECIFICATIONS

Ranges	0-50, 100, 200, 500, 1000 ppb and 0.1, 2, 5, 10, 20, 50, 100 ppm
Noise	0.20 ppb RMS (60 second averaging time)
Lower Detectable Limit	0.40 ppb (60 second averaging time)
Zero Drift (24 hour)	<0.40 ppb
Span Drift (24 hour)	± 1% full scale
Response Time	40 sec (10 second averaging time) 80 sec (60 second averaging time) 300 sec (300 second averaging time)
Precision	± 0.4 ppb (500 ppb range)
Linearity	± 1% full scale
Sample Flow Rate	0.6 liters/min
Operating Temperature	15° - 35° C ( may be safely operated over the range of 5° - 40° C)
Power Requirements	105-125 VAC, 60 HZ    220-240 VAC, 50 HZ    300 WATTS
Physical Dimensions	16.75" (W) x 8.62" (H) x 23" (D)
Weight	53 lbs.
Outputs	Selectable voltages and RS-232 standard; 4-20 mA isolated current (optional)

**TE** **Thermo Environmental  
Instruments Inc.**

**Thermo Unicam**

P.O. Box 208, York Street, Cambridge CB1 2SR, England  
Telephone: 44/1223374234 • Fax: 44/1223374338

**Thermo Instrument Systems, GMBH**

Martenerstrasse 539, 44379 Dortmund 70, WEST GERMANY  
Telephone: 231 / 9613160 • Fax: 231/96131620

**ThIS Analytical B.V.**

Heerbaan 220, 4817 Breda, HOLLAND  
Telephone: 31 / 76 / 5713717 • Fax: 31/76/5812028

**8 West Forge Pky., Franklin, MA 02038 USA**  
**TEL: 508-520-0430 Fax: 508-520-1460**

**Thermo Instrument Systems N.V.**

Waaslandlaan 8 A3, 9160 Lokeren, Belgium  
Telephone: 32 / 93 / 485841 • Fax: 32/9/3484575

**Thermo Electron Nippon Co., LTD.**

151 Ichinotsubo, Makishimacho, Uji-Shi, Kyoto, 611, JAPAN  
Telephone: 81-774-21-2111 • Fax: 81-774-21-2240

**Thermo Instrument Australia PTY., LTD**

Unit 20, 38-46 South Street, Rydalmere, N.S.W. 2116, Australia  
Telephone: 61-2-898-1244 • Fax: 61-2-684-4244

DS

# SPECIFICATIONS

SPECIFICATION	NONDISPERSIVE INFRARED (NDIR)				OXYGEN	
	CO	CO <sub>2</sub>	CH <sub>4</sub>	SO <sub>2</sub>	O <sub>2</sub>	O <sub>2</sub>
DETECTOR	Microflow				Paramagnetic	Galvanic Fuel Cell
SAMPLE CONTACT MATERIAL	Stainless steel and Tygon <sup>†</sup> disposable gold plated cell liner. Window material CaF <sub>2</sub> .				Platinum, Glass, Stainless Steel, Viton and Tygon <sup>†</sup>	Stainless Steel and Tygon <sup>†</sup>
RANGES	See Table, Page 3				See Table, Page 3	See Table, Page 3
LINEARITY	Better than 1% Full Scale				Better than 1% Full Scale	Better than 1% Full Scale
REPEATABILITY	Better than 1% Full Scale				Better than 1% Full Scale	Better than 1% Full Scale
RESPONSE TIME	90% of Full Scale in less than 1 second*				90% of Full Scale in 2 seconds	90% of Full Scale in 5 seconds
SAMPLE FLOW RATE	0.5 to 2 liters/minute				1 liter/minute	1 liter/minute
NOISE	Less than 1% Full Scale				Less than 1% Full Scale	Less than 1% Full Scale
ZERO & SPAN DRIFT	Less than 1% of Full Scale in 24 hours				Less than 1% of Full Scale in 24 hours	Less than 1% of Full Scale in 24 hours
ZERO & SPAN ADJUSTMENT	Ten turn potentiometer				Ten turn potentiometer	Ten turn potentiometer
DISPLAY	Individual 4 1/2 digit panel meter				3 1/2 digit panel meter	3 1/2 digit panel meter
OUTPUTS	Select: 0-10VDC or 4-20 or 0-20mA (RS-232 optional)				Select: 0-10VDC and 4-20 or 0-20mA (RS-232 optional)	Select: 0-10VDC and 4-20 or 0-20mA (RS-232 optional)
ALARMS (optional)	2 each: Form C-10A SPDT and Form A-5A SPST (Dry Relay Contacts)				Single or Dual SPDT/5A (Dry Relay Contacts)	Single or Dual SPDT/5A (Dry Relay Contacts)
AMBIENT TEMPERATURE	-5 to 45°C				-5 to 45°C	-5 to 45°C
SAMPLE TEMPERATURE	0 to 50°C				0 to 50°C	0 to 50°C
SAMPLE CONDITION	Clean, non-condensing gas				Clean, non-condensing gas	Clean, non-condensing gas
FITTINGS	1/4" tube				1/4" tube	1/4" tube
POWER REQUIREMENTS	115/220/240 VAC, 50/60 Hz, 70 watts/channel				115/220/240 VAC, 50/60 Hz, 70 watts/channel	115/220/240 VAC, 50/60 Hz, 70 watts/channel
DIMENSIONS	5 1/4" H x 19" W x 22" D 133mm x 483mm x 508mm				5 1/4" H x 19" W x 15" D 133mm x 483mm x 381mm	5 1/4" H x 19" W x 15" D 133mm x 483mm x 381mm
RELATIVE HUMIDITY	Less than 90% R.H.**				Less than 90% R.H.**	Less than 90% R.H.**
WEIGHT (single unit)	24 lbs. (approximate) 10.8 Kg				15 lbs. (approximate) 6.8 Kg	10 lbs. (approximate) 4.8 Kg

\*Depending on cell length and flow rate \*\*Non-condensing

Specifications are subject to change without notice.



**ENTECH ENGINEERING INC.**

*P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

**APPENDIX E.**

**EQUIPMENT CALIBRATIONS**

# **ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## **METER BOX CALIBRATION WET TEST METER (SN# P2768)**

Date **10/23/09**

Barometric Pressure **30.00**

Meter box number

**1**

Sample Line Vacuum  Hg		Orifice manometer setting  $\Delta H$  in. H <sub>2</sub> O	Spirometer (wet meter) gas volume  Vw (ft <sup>3</sup> )	Dry gas meter volume  Vm (ft <sup>3</sup> )	Temperatures				Time  $\theta$ min	Spirometer (wet meter) Vacuum Pw in. H <sub>2</sub> O
					Spirometer (wet meter)	Dry Gas Meter				
						Inlet  ti °F	Outlet  to °F	Average  tm °F		
5.0	Start	0.5	7709.000	503.650	80		76	77.0	12.40	0.7
	End		7714.000	508.600	80		78			
	Volume		5.000	4.950	80					
5.0	Start	1.0	7714.000	508.600	80		78	79.0	8.70	1
	End		7719.000	513.565	80		80			
	Volume		5.000	4.965	80					
5.0	Start	2.0	7719.000	513.565	80		80	80.5	6.35	1.8
	End		7724.000	518.535	80		81			
	Volume		5.000	4.970	80					
5.0	Start	3.0	7724.000	518.535	80		81	81.5	5.30	2.8
	End		7729.000	523.490	80		82			
	Volume		5.000	4.955	80					
5.0	Start	4.0	7729.000	523.490	80		82	82.5	4.45	3.5
	End		7734.000	528.457	80		83			
	Volume		5.000	4.967	80					

### Calculations

$$\gamma = \frac{(V_w) (P_b - (P_w/13.6)) (t_m + 460)}{(V_m) (P_b + (H/13.6)) (t_w + 460)} \Delta H @ = \frac{(0.0317) (\Delta H)_x}{(P_b) (t_o + 460)} \left[ \frac{(t_w + 460) (\theta)}{(V_w)} \right]^2$$

$\Delta H$	$\Delta$	$\Delta H @$	Condition
0.5	1.002	1.77	OK
1	1.000	1.73	OK
2	0.998	1.84	OK
3	0.998	1.92	OK
4	0.993	1.80	OK
Average	<b>0.998</b>	<b>1.81</b>	

K-factor = ( Volume (ft)<sup>3</sup> ) X (  $\gamma$  ) X ( Meter Temp (F)+460 )

(B.P.) X (Time)

0

Thermocouple Readout Calibration							
NIST Tracable Simulator Make: Omega Model: CL540K S.N. 105823							
Type-K Simulator	Stack **	Probe **	Filter **	Last Imp. **	Aux **	Meter In **	Meter Out **
0	1	1	1	1	1	1	1
50	50	50	50	50	50	50	50
100	99	99	99	99	99	99	99
200	199	199	199	199	199	199	199
500	502	502	502				
1000	1001						
1900	1900						

\*\*Note: all temperatures are in °F

Sign: *[Signature]*

Date: **10-23-09**

*EL*

# ENTECH ENGINEERING INC.

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

## METER BOX CALIBRATION WET TEST METER (SN# P2768)

Date **12/08/09**

Barometric Pressure **30.15**

Meter box number

**1**

Sample Line Vacuum  Hg		Orifice manometer setting  Δ H  in. H2O	Spirometer (wet meter) gas volume  Vw (ft3)	Dry gas meter volume  Vm (ft3)	Temperatures				Time  θ min	Spirometer (wet meter)  Vacuum Pw in. H2O
					Spirometer  (wet meter)	Dry Gas Meter				
						Inlet	Outlet	Average		
5.0	Start	0.5	7761.000	799.884	68		68	69.0	12.40	0.7
	End		7766.000	804.850	69		70			
	Volume		5.000	4.966	68.5					
5.0	Start	1.0	7766.000	804.850	69		70	71.0	9.00	1
	End		7771.000	809.805	70		72			
	Volume		5.000	4.955	69.5					
5.0	Start	2.0	7771.000	809.805	70		72	72.5	6.50	1.8
	End		7776.000	814.756	69		73			
	Volume		5.000	4.951	69.5					
5.0	Start	3.0	7776.000	814.756	69		73	73.5	5.30	2.8
	End		7781.000	819.670	69		74			
	Volume		5.000	4.914	69					
5.0	Start	4.0	7781.000	819.670	69		74	74.5	4.50	3.5
	End		7786.000	824.594	70		75			
	Volume		5.000	4.924	69.5					

### Calculations

$$\gamma = \frac{(Vw) (Pb - (Pw/13.6)) (tm + 460)}{(Vm) (Pb + (H/13.6)) (tw + 460)} \quad \Delta H@ = \frac{(0.0317) (\Delta H_x)}{(Pb) (to + 460)} \left[ \frac{(tw + 460) (\theta)}{(Vw)} \right]^2$$

$\Delta H$	$\Delta$	$\Delta H@$	Condition
0.5	1.005	1.71	OK
1	1.007	1.80	OK
2	1.006	1.87	OK
3	1.012	1.86	OK
4	1.006	1.79	OK
Average	1.007	1.80	

$$K\text{-factor} = \frac{(\text{Volume (ft}^3\text{)}) \times (\gamma) \times (\text{Meter Temp (F)} + 460)}{(\text{B.P.}) \times (\text{Time})}$$


Thermocouple Readout Calibration							
NIST Tracable Simulator Make: Omega Model: CL540K S.N. 105823							
Type-K Simulator	Stack **	Probe **	Filter **	Last Imp. **	Aux **	Meter In **	Meter Out **
0	1	1	1	1	1	1	1
50	50	50	50	50	50	50	50
100	99	99	99	99	99	99	99
200	199	199	199	199	199	199	199
500	502	502	502				
1000	1001						
1900	1900						

\*\*Note: all temperatures are in °F

Sign: 

Date: 12.8.09

E2



# Certificate Of Calibration

for

ENTECH ENGINEERING

Cust. P.O. #: 52257

Model #: CL540K

Omega W.O. #: 809948892

Serial #: 105823

Report #: 081810

Recal Date: 11-Sep-09


## CAL-3

Omega Engineering, Inc. certifies that the above instrumentation has been calibrated and tested to meet or exceed the published specifications. This calibration and testing was performed using instrumentation and standards that are traceable to the National Institute of Standards and Technology. This calibration is in compliance with ISO 10012-1, ISO 9001 Section 4.11 and ANSI/NCSL Z540-1-1994. This Certificate/Report shall not be reproduced, except in full, without written consent of Omega Engineering, Inc.

Test Conditions: 22 C

Relative Humidity: 40%

Test Equipment	NIST Traceable Test Number(s)
Hewlett Packard 34401A	775585-5985307: 1168883856
Omega TRC-III	OM-80310864

  
Metrology Technician

  
Quality Assurance Inspector

Calibration Date: 11-Sep-08

o:\cal certs\cal-3cal.doc  
Page 1 of 2



Model No: CL540K

Serial No: 105823

Report No: 081810

Maximum Calibration System Uncertainty: 9ppm Rdg+3ppm Rng(mVDC)

Procedure: QAP-2100

Note: OOT denotes out of tolerance condition otherwise, UUT has passed published specifications.

Test Equipment	S/N	Asset	Calibration Due
Hewlett Packard 34401A	US36030996	01109	17-Dec-08
Omega TRC-III	I-0006	01158	17-Sep-08

# ENTECH ENGINEERING INC.

P. O. Box 890746, Houston, Texas 77289-0746, (281) 332-3118

## Temperature Measurement Device Calibration

Device ID: T36-F3  
Device Type: Type K Thermocouple  
Person Performing Calibration: J. Vawter  
Calibration Date: 01/16/09  
Reference Standard Type: Mercury Thermometer  
Reference Standard Calibration Date: 08/06/08  
NIST Traceable Standard Manufacturer: ICL Calibration Laboratories Inc.  
NIST Traceable Standard Serial No. C92043

Calibration Point	Measurement Range	Thermocouple		Reference Standard		Percent Deviation	Pass / Fail Specification
		°F	°R	°F	°R		
1	Cold	32.4	492.4	32.0	492.0	-0.08%	Pass
2	Ambient	74.5	534.5	76.0	536.0	0.28%	Pass
3	Elevated 1	229.0	689.0	227.0	687.0	-0.29%	Pass
4	Elevated 2	644.0	1104.0	641.0	1101.0	-0.27%	Pass

Specification: Percent deviation  $\leq 1.5\%$  of reference temperature.

I hereby certify the temperature measurement device as being accurate to within the acceptable specification:

Signature:

*James Vawter*

Date:

1-16-09

# ICL CALIBRATION LABORATORIES, INC.



CERT. 526.91 CALIBRATION

**ISO/IEC 17025 and ANSI/NCSL Z540-1 accredited**  
*The specialists in ASTM and laboratory thermometers & hydrometers*  
Members: ASTM API NCSLI ASQ NCWM

1501 Decker Avenue Suite 118 Stuart, FL 34994 USA  
Tel: 772 286 7710 1-800-713-6647  
Fax: 772 286 8737 E-mail: sales@icllabs.com  
Internet: www.icllabs.com

Setting new standards in calibration excellence!

## CALIBRATION REPORT FOR THERMOMETER

Report No. S168354 Page 1 of 3

THE INSTRUMENT DESCRIBED BELOW WAS EXAMINED AND TESTED IN ICL'S ISO/IEC 17025 ACCREDITED CALIBRATION LABORATORY, USING NIST TRACEABLE REFERENCE STANDARDS, IN ACCORDANCE WITH ICL'S ISO/IEC 17025 CALIBRATION PROCEDURE REFERENCED BELOW. THIS CALIBRATION MEETS THE REQUIREMENTS OF ISO/IEC 17025, ANSI/NCSL Z540-1-1994, (WHICH SUPERCEDED AND REPLACED MIL-STD 45662A), AND THE ISO-9000 AND QS-9000 SERIES OF QUALITY STANDARDS.

### CUSTOMER INFORMATION

ENTECH ENGINEERING, INC.  
408 EAST MAIN STREET  
LEAGUE CITY, TX 77573

PURCHASE ORDER NUMBER: NOT AVAILABLE

SUBMITTED BY: ENTECH ENGINEERING, INC.

### INSTRUMENT INFORMATION

DATE REPORT ISSUED: 09-11-2009

THERMOMETER ASTM 3F-99 MODEL NUMBER: 10003F-C

SERIAL NUMBER: 1675 INSCRIPTION: LSW

ENGINEERING UNITS: degrees Fahrenheit RANGE: 20/760F DIVISIONS: 2 °F

IMMERSION: 76MM

ACCURACY TOLERANCE (maximum scale error permitted by ASTM E 1): +/- 2.0F TO 574F, +/- 3.0F ABOVE 574F

### RESULTS OF PHYSICAL EXAMINATION

THIS INSTRUMENT WAS EXAMINED UNDER A POLARIZED LENS AND STRAINS IN THE GLASS, IF ANY, WERE JUDGED TO BE MINIMAL AND OF NO DETRIMENT TO THE FUNCTION OF THE INSTRUMENT.

THE CAPILLARY OF THIS THERMOMETER WAS EXAMINED UNDER MAGNIFICATION WITH RESULTS AS FOLLOWS:  
NO FOREIGN MATERIAL, MOISTURE, OR OTHER EVIDENCE OF CONTAMINATION WERE DISCOVERED. NO DISCERNABLE CAPILLARY IRREGULARITIES WERE NOTED.

IT WAS DETERMINED THAT THIS INSTRUMENT IS IN GOOD WORKING ORDER AND IS THEREFORE SUITABLE FOR CALIBRATION.

**CALIBRATION PROCEDURE USED:** ICL Procedure 01, which is based upon ASTM E 77, NBS Monograph 150 & NIST SP 250-23

### RESULTS OF CALIBRATION

NOTE: The indications of this instrument cannot be adjusted or modified by ordinary means; accordingly, the readings given in the table below should be considered, in effect, to be both 'As Found' and 'As Left' readings.

The values and corrections presented herein are derived, or calculated, data as required by ASTM method E 77 for the calibration of this ASTM partial immersion thermometer. Raw data (actual observations) obtained in the performance of this calibration and an explanation of the calculations are presented on Page 3 of this report.

TEST TEMP	READING	CORRECTION	ACCEPT LIMIT* (+ or -)	P/M/F	ASTM SPECIFIED STEM TEMP	UNCERTAINTY
32.00°F	30.2°F	+1.8°F	2.00°F	PASS	70°F	0.83°F
200.00°F	200.3°F	-0.3°F	2.00°F	PASS	90°F	0.83°F
370.00°F	370.9°F	-0.9°F	2.00°F	PASS	101°F	0.83°F
540.00°F	540.9°F	-0.9°F	2.00°F	PASS	110°F	0.83°F
700.00°F	700.9°F	-0.9°F	3.00°F	PASS	129°F	0.84°F

\*ACCEPT LIMIT(s) The acceptance limit(s) shown above represent a statistical evaluation of the instrument's tolerance relative to the uncertainty of the measurement. If required, the acceptance limit is set to a value smaller than the tolerance. The difference between the tolerance and the acceptance limit is the 'guard band'. The guard band is imposed to reduce the probability of a false acceptance (PFA), or a false failure, to 2% or less.

P/M/F Accordingly, there are three possible calibration outcomes:

1. **PASS** The calibration result falls within the interval described by the test point + or - (the tolerance MINUS the guard band).
2. **MARG\*\*** (marginal) The calibration result is 'borderline', or indeterminate; it is therefore statistically and metrologically imprudent to declare that the instrument is definitively either 'in-tolerance' or 'out-of-tolerance'.
3. **FAIL** The calibration result falls outside the interval described by the test point + or - (the tolerance PLUS the guard band).

The methodology and equations used for determination of guard bands and acceptance limits comply with the requirements of ANSI/NCSL Z540.3

THE TEST POINTS LISTED IN THE ABOVE TABLE ARE THOSE SPECIFIED IN ASTM E 1 (CURRENT REVISION).  
Our best measurement capabilities are: at Liquid Nitrogen (approximately -196C), +/- 0.0062C; from -80 to 0C, +/- 0.0089C; at 0C,

EL

+/- 0.0039C; at 0.01C (FPW), +/- 0.0019C; from 0.01 to 100C, +/- 0.0085C; from 100 to 200C, +/- 0.0094C; from 200 to 300C, +/- 0.0098C; from 300 to 420C, +/- 0.014C; from 420 to 500C, +/- 0.034C; from 500 to 700C, +/- 0.26C; from 700 to 1000C, +/- 0.86C. These uncertainties have been calculated utilizing the methods recommended in NIST Technical Note 1297 and the ANSI-NCCL document Z-540-2 entitled 'Guide to the Expression of Uncertainty in Measurement'. A coverage factor of 2 sigma (k=2) has been applied to the standard uncertainty in order to express the expanded uncertainty at (approximately) a 95% confidence level.

THE UNCERTAINTIES PRESENTED ABOVE IN THE 'RESULTS' TABLE ARE LARGER THAN OUR BEST MEASUREMENT CAPABILITIES, AS THE RESOLUTION OF THIS INSTRUMENT, ESTIMATED TO BE 0.2°F, AND OTHER CONTRIBUTIONS HAVE BEEN FACTORED INTO THE CALCULATION.

BECAUSE THIS IS A PARTIAL IMMERSION THERMOMETER, AN ADDITIONAL 0.6°F HAS BEEN FACTORED INTO THE UNCERTAINTY CALCULATION AS SUGGESTED BY NIST IR 5341.

THE EXPANDED UNCERTAINTIES (K=2) REPORTED HERE DO NOT CONTAIN ESTIMATES FOR (1) ANY EFFECTS THAT MAY BE INTRODUCED BY TRANSPORTATION OF THE INSTRUMENT BETWEEN ICL AND THE USER'S LABORATORY, (2) DRIFT OF THE INSTRUMENT, (3) HYSTERESIS OF THE INSTRUMENT, OR (4) ANY MEASUREMENT UNCERTAINTIES INTRODUCED BY THE USER.

FOR A DISCUSSION OF ACCURACIES ATTAINABLE WITH THERMOMETERS SUCH AS THIS INSTRUMENT SEE NIST SPECIAL PUBLICATION 250-23, NIST PUBLICATION IR-5341, ASTM E 1 AND ASTM E 77.

LABORATORY ENVIRONMENTAL CONDITIONS: TEMPERATURE: 23°C +/- 2°C RELATIVE HUMIDITY: BETWEEN 40% AND 65%

ALL TEMPERATURES GIVEN IN THIS REPORT ARE THOSE DEFINED BY THE INTERNATIONAL TEMPERATURE SCALE OF 1990 (ITS-90)

\*\* IMPORTANT NOTE: THE READINGS AND CORRECTIONS NOTED ABOVE APPLY FOR THE CONDITION OF IMMERSION INDICATED PROVIDED THE ICE POINT READING, TAKEN AFTER EXPOSURE FOR NOT FEWER THAN THREE DAYS TO A TEMPERATURE OF ABOUT 23 DEGREES CELSIUS (73°F), IS 30.2°F. IF THE ICE POINT READING IS FOUND TO BE HIGHER (OR LOWER) THAN STATED, ALL OTHER READINGS WILL BE HIGHER (OR LOWER) BY THE SAME AMOUNT.

THIS CALIBRATION WAS PERFORMED BY: DEBORAH M. WEBER

THE CALIBRATION PERFORMED AND DOCUMENTED BY THIS REPORT OF TEST IS A FULL SCALE CALIBRATION AND NO LIMITATIONS OF USE ARE IMPOSED ON THIS INSTRUMENT.

### TRACEABILITY INFORMATION

This calibration is traceable to NIST through an unbroken chain of comparisons. The reference standard is used to calibrate the transfer standard, which in turn is used to calibrate the client's instrument. Each step in the chain is fully documented, and measurement uncertainty at each step has been calculated.

Our NIST primary reference thermometer from -196 to 420C is a Rosemount model 162CE 25.5 Ohm SPRT, serial no. 5206, calibrated by NIST on April 1, 2009. NIST GMP-11 recommends a 36 month calibration cycle for SPRTs. PRT transfer standards and ASTM liquid-in-glass transfer standards are calibrated annually against this SPRT, per NIST GMP-11 recommendations.

Our primary reference thermometer for temperatures from 500 to 1000C is a Hart Scientific model 5624 PRT sensor, serial #0105, calibrated by Hart Scientific. PRT and noble metal thermocouple transfer standards are calibrated annually against this reference sensor, per NIST GMP-11 recommendations.

Test Point	Comparator	MTE#	Manufacturer	Transfer Standard	MTE#	Manufacturer	Next Due
32.00°F	Ice bath	000	Lab Glass	Ice bath	222	Lab Glass	10/06/09
200.00°F	7310 oil bath	005	PolyScience	5628-15 PRT 1211	271	Hart Scientific	06/02/10
370.00°F	6022 oil bath	021	Hart Scientific	5628-15 PRT 1111	286	Hart Scientific	06/02/10
540.00°F	6045 salt bath	004	Hart Scientific	5628 PRT 0523	228	Hart Scientific	10/07/09
700.00°F	6050H salt bath	016	Hart Scientific	5628 PRT 0523	228	Hart Scientific	10/07/09

### ICL CALIBRATION LABORATORIES, INC.

In ISO/IEC 17025 & ANSI/NCSL Z-540-1 accredited laboratory - American Association for Laboratory Accreditation Certificate #526.01

J. JEFF KELLY, TECHNICAL DIRECTOR

DEBORAH M. WEBER, A.S.C.P. ACCREDITED TECHNOLOGIST This document prepared by LORI PARR and reviewed by KAREN DOMINGUEZ

DATE REPORT ISSUED: 09-11-2009

RECALIBRATION DATE SPECIFIED BY CLIENT: September 11, 2010

NIST GMP-11 (Mar '03), 'Good Measurement Practice for Assignment and Adjustment of Calibration Intervals for Standards' states that, 'Temperature standards are dynamic with use. Shock, contamination and other factors can cause drift from accepted values'. Table 4 of GMP-11 recommends recalibration of liquid-in-glass thermometers, standard thermistors and PRTs at 12 month intervals. Liquid-in-glass thermometers used for 'Temperature Critical Parameters' should be recalibrated at 6 month intervals. NIST GMP-11 is available for download in Adobe .pdf format on our website at [www.icllabs.com](http://www.icllabs.com) Follow the link for 'Downloads'.

The API 'Manual of Petroleum Measurement Standards', Chapter 7, June, 2001, specifies a 12 month recalibration interval for liquid-in-glass thermometers (see section 8.3) and for portable electronic thermometers (PETs). See section 8.2

The user should be aware that any number of factors may cause this instrument to drift out of calibration before the specified calibration interval has expired.

This Report of Test may not be reproduced except in full without the express written permission of ICL Calibration Laboratories, Inc.

This calibration report applies only to the item calibrated. This calibration report shall not be used to claim product endorsement by the A2LA.

REPORT NUMBER: S168354 Page 2 of 3

E7



## RAW DATA

The actual raw data obtained during testing of this instrument is as follows:

Test temperature	Reading	Emergent column temp (To)	degs of emergent stem (n)
32.00°F	30.2°F	71°F	89
200.00°F	200.2°F	86°F	257
370.00°F	370.2°F	84°F	427
540.00°F	539.8°F	89°F	597
700.00°F	698.8°F	98°F	757

THE ABOVE READINGS WERE MADE UNDER MAGNIFICATION AND RESOLVED TO ONE TENTH OF ONE SCALE DIVISION.

Because the temperature of the emergent stem (the portion of the mercury column above the bath, exposed to room temperatures) realized during testing differs from the ASTM E 1 specified emergent stem temperatures (assumed to be the conditions of actual or intended use of the instrument), the method requires that an emergent stem temperature correction be calculated and applied. This is explained fully in ASTM E 77, section 7, Treatment of Data.

The calculation utilized is:

$$\text{Emergent stem correction} = kn(T_s - T_o)$$

where:

$k$  = differential expansion coefficient of the liquid and the glass of which the thermometer is made; for Celsius mercury-in-glass thermometers the value of  $k$  is 0.00016, for Fahrenheit mercury-in-glass thermometers  $k=0.00009$  For Celsius organic liquid-in-glass thermometers  $k=0.001$ , and for Fahrenheit organic liquid-in-glass thermometers  $k=0.0006$

$n$  = number of (scale) degrees of liquid column emergent from the bath, as measured from the immersion mark to the top of the mercury column. The ungraduated portion of the stem between the immersion mark and the beginning of the scale is evaluated and included into the value of  $n$

$T_s$  = the emergent stem temperature specified in ASTM E 1, table 4

$T_o$  = observed average temperature of the emergent liquid column of  $n$  degrees

This calculation has been performed for each temperature tested, and, as required by ASTM E 77, the results have been incorporated into the readings and corrections which appear under RESULTS OF CALIBRATION on page 1 of this report.

Report Number: S168354 Page 3 of 3

E8



Calibration  
Certificate No. 1750.01

Calibration complies with  
ISO/IEC 17025 AND ANSI/NCSL Z540-1



Cert. No.: 4015-1806859

# Traceable® Certificate of Calibration for Total-Range Thermometer

Instrument Identification:

**I-6**

Model: 4015

S/N: 80148561

Manufacturer : Control Company

Standards/Equipment:

Description  
Thermocouple Calibrator

Serial Number  
00633877

Due Date  
11/29/08

NIST Traceable Reference  
1000235086

Certificate Information:

Technician: 76

Procedure: CAL-4015-06

Cal Date: 2/04/08

Cal Due: 2/03/10

Test Conditions: 23.0°C 52.0 %RH 1012 mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±uc	TUR
°C		N.A.		0.000	0.3	Y	-1.0	1.0	0.360	2.8:1
°C		N.A.		1,000.000	999.8	Y	998.0	1,002.0	0.760	2.6:1

This Instrument was calibrated using Instruments Traceable to National Institute of Standards and Technology.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ±uc=Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=±(Max-Min)/2; Date=MM/DD/YY

Wallace Berry  
Wallace Berry, Technical Manager

Maintaining Accuracy:

In our opinion once calibrated your Total-Range Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Total-Range Thermometers change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

CONTROL COMPANY 4455 Rex Road Friendswood, TX 77546 USA  
Phone 281 482-1714 Fax 281 482-9448 service@control3.com www.control3.com

Control Company is an ISO 17025 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01  
Control Company is ISO 9001 Quality Certified by (DNV) Det Norske Veritas, Certificate No. CERT-01805-AQ-11011  
Control Company is a member of the International Laboratory Accreditation Cooperation (ILAC) - Multilateral Recognition Arrangement (MLA)

Page 1 of 1

Printed on 2/4/08 at 10:00 AM

2/4/08 10:00 AM

E9

**ENTECH ENGINEERING INC.**

P. O. Box 890746, Houston, Texas 77289-0746, (281) 332-3118

**Toploader Balance Calibration**

Manufacturer: OHAUS

Model No: SC6010

Serial No: BJ346242

Entech ID: TL - 3

NIST Traceable Standard Manufacturer: TROEMNER

NIST Traceable Standard Serial No. 25143, 25144, 82048

Calibration Temperature (°F):

Calibration Date: 10/26/09

Operator: JMM

Next Calibration:

Standard ID	NIST Traceable Standard  (grams)	Measured Value  (grams)	Difference  (grams)	Percent Error  (%)
1	100.0	100.0	0.0	0.0
2	200.0	200.0	0.0	0.0
3	300.0	300.0	0.0	0.0

Average Percentage Error =

0.0%

E10

# ENTECH ENGINEERING INC.

P. O. Box 890746, Houston, Texas 77289-0746, (281) 332-3118

## Portable Barometer Calibration

Barometer ID: Barometer #10 (I-3)

### Presample Calibration

Date: 11/12/09

Time: 10:35 AM

Entech Laboratory Mercury Barometer Standard (Actual)

30.04 in Hg

Portable Barometer (Actual)

30.03 in Hg

Temperature at Barometer Reading Site

83.3 °F

Portable Barometer Reading Error

0.01 in Hg

Portable Barometer Reading Acceptable? (Yes/No)

Yes

\* (Tolerance +/- 0.10 in Hg)

I certify that I have check the portable barometer indentified above and that it agrees with the laboratory barometer standard within +/- 0.10 in Hg.

  
Signature

11/12/09  
Date

If Unacceptable, Input Correction  
Corrected Portable Barometer Reading

in Hg

in Hg

I certify that I have adjusted the portable barometer indentified above to the laboratory barometer standard.

\_\_\_\_\_  
Signature

\_\_\_\_\_  
Date

### Postsample Calibration Check

Date: 11/23/09

Time: 08:00 AM

Entech Laboratory Mercury Barometer Standard (Actual)

30.15 in Hg

Portable Barometer (Actual)

30.15 in Hg

Temperature at Barometer Reading Site

73 °F

Portable Barometer Reading Error

0.00 in Hg

Portable Barometer Reading Acceptable? (Yes/No)

Yes

\* (Tolerance +/- 0.10 in Hg)

I certify that I have check the portable barometer indentified above and that it agrees with the laboratory barometer standard within +/- 0.10 in Hg.

  
Signature

11/23/09  
Date



**ENTECH ENGINEERING INC.**

P. O. Box 890746 . Houston, Texas 77289-0746 . (281)332-3118

**Caliper Calibration**

**Entech Caliper #1**

**02/10/09**

NIST Gauge Block Size	NIST Gauge Block Iden. No. / Tracable Number	Caliper Reading
0.400	011133 (NIST No. 821/263174-00)	0.399
0.200	010690 (NIST No. 821/263174-00)	0.199

Technician Marlon Jeffery Bowen

Sign. *M. J. Bowen*

Date 02/10/09

***ENTECH ENGINEERING INC.***

***P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118***

---

**APPENDIX F.**

**CALIBRATION GAS CERTIFICATIONS**

235

**AIR LIQUIDE**Air Liquide America  
Specialty Gases LLC**Scott™****RATA CLASS***Dual-Analyzed Calibration Standard*

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800

Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: EPA Protocol Gas**Assay Laboratory

P.O. No.: ALAS-44414/33645229  
 AIR LIQUIDE AMERICA SPECIALTY GASES LLC Project No.: 04-74415-001  
 9810 BAY AREA BLVD  
 PASADENA, TX 77507

Customer

ENTECH ENGINEERING INC  
 PO#52447  
 408 E MAIN  
 ATTN: JARAD VAWTER  
 LEAGUE CITY TX 77573

**ANALYTICAL INFORMATION**

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards; Procedure G-1; September, 1997.

Cylinder Number: CC57866 Certification Date: 17Jun2009 Exp. Date: 17Jun2011  
 Cylinder Pressure\*\*\*: 1935 PSIG

COMPONENT

SULFUR DIOXIDE \*  
 NITROGEN

CERTIFIED CONCENTRATION (Moles)

47.3 PPM  
 BALANCE

ANALYTICALACCURACY\*\*

+/- 1%

TRACEABILITY

Direct NIST and NMI

\*\*\* Do not use when cylinder pressure is below 150 psig.

\*\* Analytical accuracy is based on the requirements of EPA Protocol Procedure G1, September 1997.

\* This Protocol has been certified using corrected NIST SO2 standard values, per EPA guidance dated 7/24/96 and will not correlate with uncorrected Pro

**REFERENCE STANDARD**

TYPE/SRM NO.	EXPIRATION DATE	CYLINDER NUMBER	CONCENTRATION	COMPONENT
NTRM 1693	15Aug2009	ALM024377	50.79 PPM	SULFUR DIOXIDE

**INSTRUMENTATION**INSTRUMENT/MODEL/SERIAL#

FTIR/000929060

DATE LAST CALIBRATED

02Jun2009

ANALYTICAL PRINCIPLE

FTIR

**ANALYZER READINGS**

(Z = Zero Gas R = Reference Gas T = Test Gas r = Correlation Coefficient)

First Triad Analysis      Second Triad Analysis      Calibration Curve

**SULFUR DIOXIDE \***

Date: 09Jun2009 Response Unit: PPM

Z1=0.00240	R1=50.82262	T1=47.37937
R2=50.87813	Z2=0.01212	T2=47.38182
Z3=0.01269	T3=47.40593	R3=50.89103
Avg. Concentration: 47.32 PPM		

Date: 17Jun2009 Response Unit: PPM

Z1=-0.00449	R1=50.87268	T1=47.30392
R2=50.87285	Z2=0.00996	T2=47.31434
Z3=0.01586	T3=47.33644	R3=50.90194
Avg. Concentration: 47.23 PPM		

Concentration = A + Bx + Cx<sup>2</sup> + Dx<sup>3</sup> + Ex<sup>4</sup>  
 r = 9.99995E-1

Constants: A = 0.00000E+0  
 B = 9.94627E-1 C = 0.00000E+0  
 D = 0.00000E+0 E = 0.00000E+0

Special Notes: LOT# PAS00595

APPROVED BY:

Ramien JR



**AIR LIQUIDE**Air Liquide America  
Specialty Gases LLC

Scott™

**RATA CLASS***Dual-Analyzed Calibration Standard*

234

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800

Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: EPA Protocol Gas**Assay LaboratoryAIR LIQUIDE AMERICA SPECIALTY GASES LLC  
9810 BAY AREA BLVD  
PASADENA, TX 77507P.O. No.: ALAS-44414/33645193  
Project No.: 04-74414-001CustomerENTECH ENGINEERING INC  
PO#52447  
408 E MAIN  
ATTN: JARAD VAWTER  
LEAGUE CITY TX 77573**ANALYTICAL INFORMATION**This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards;  
Procedure G-1; September, 1997.Cylinder Number:

CC144163

Certification Date:

25Jun2009

Exp. Date: 24Dec2009Cylinder Pressure\*\*\*:

1848 PSIG

**COMPONENT**SULFUR DIOXIDE \*  
NITROGEN**CERTIFIED CONCENTRATION (Moles)**

22.2

PPM

BALANCE

**ANALYTICAL****ACCURACY\*\***

+/- 1%

**TRACEABILITY**

Direct NIST and NMI

\*\*\* Do not use when cylinder pressure is below 150 psig.

\*\* Analytical accuracy is based on the requirements of EPA Protocol Procedure G1, September 1997.

\* This Protocol has been certified using corrected NIST SO2 standard values, per EPA guidance dated 7/24/96 and will not correlate with uncorrected Pro

**REFERENCE STANDARD**TYPE/SRM NO.

NTRM 1693

EXPIRATION DATE

15Aug2009

CYLINDER NUMBER

ALM024377

CONCENTRATION

50.79 PPM

COMPONENT

SULFUR DIOXIDE

**INSTRUMENTATION**INSTRUMENT/MODEL/SERIAL#

FTIR/000929060

DATE LAST CALIBRATED

03Jun2009

ANALYTICAL PRINCIPLE

FTIR

**ANALYZER READINGS**First Triad Analysis

(Z=Zero Gas R=Reference Gas T=Test Gas r=Correlation Coefficient)

Second Triad AnalysisCalibration Curve**SULFUR DIOXIDE \***

Date:	18Jun2009	Response Unit:	PPM
Z1=	0.00975	R1=	50.85954
T1=	22.18030		
R2=	50.87841	Z2=	0.02603
T2=	22.24813		
Z3=	0.03306	T3=	22.25493
R3=	50.91176		
Avg. Concentration:	22.17	PPM	

Date:	25Jun2009	Response Unit:	PPM
Z1=	-0.00794	R1=	50.84770
T1=	22.31024		
R2=	50.85970	Z2=	0.03687
T2=	22.31388		
Z3=	0.04870	T3=	22.33800
R3=	50.87003		
Avg. Concentration:	22.27	PPM	

Concentration = A + Bx + Cx2 + Dx3 + Ex4	
r = 9.99995E-1	
Constants:	A = 0.00000E+0
B = 1.00387E+0	C = 0.00000E+0
D = 0.00000E+0	E = 0.00000E+0

**Special Notes:**

LOT # PAS00615

APPROVED BY:

Peter Brandon



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Cyl. Number: CC236179

Customer: Entech Engineering  
P.O. Number: 51831  
Item Number:  
Notes:

Shipping Order #: 25000332  
Transfer #: 25000332  
LOT #: LPX220346  
Valve: CGA350  
Cyl. Pressure: 1900psig

Assay Date: 3-May-07

Expiration Date: 2-May-10

\*Cylinder should not be used when gas pressure is below 150 psig

Component	Requested Concentration	Assay Concentration
Propane	50 ppm	52.6 ±0.7 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS281	GMIS281	49.9	ppm	0.184	C3H8	N2	CC28041	5/3/2008	N.A.

#### Analysis Information:

Component 1: Propane		First Triad Analysis On: 5/3/2007					Second Triad Analysis On:			
Analyzer Information		Zero	Trial 1	Trial 2	Trial 3	Units	Zero	Trial 1	Trial 2	Trial 3
Manufacturer:	HP		0.00	0.00	0.00					
Model Number:	6890									
Serial Number:	8295									
Analytical Principle:	GC-FID									
MPC Calibrated:	04/05/07									
		Reference	47.99	47.96	47.93		Reference			
		Candidate	50.34	50.78	50.63		Candidate			
		Result	52.37	52.84	52.68	ppm	Result			
		Mean Result:		52.63	ppm		Mean Result:			

Analyst Signature: Warren Pereira Warren Pereira

Calculated by: Warren Pereira Warren Pereira

13



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: Entech Engineering  
P.O. Number: 51831  
Item Number:  
Notes:

Cyl. Number: CC161053

Shipping Order #: 25000357  
Transfer #: 25000357  
LOT #: LPX220344  
Valve: CGA350  
Cyl. Pressure:\* 1900psig

\*Cylinder should not be used when gas pressure is below 150 psig

Assay Date: 3-May-07

Expiration Date: 2-May-10

Component	Requested Concentration	Assay Concentration
Propane	30 ppm	31.0 ±0.5 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS281	GMIS281	49.9	ppm	0.184	C3H8	N2	CC28041	5/3/2008	N.A.

#### Analysis Information:

Component 1: Propane		First Triad Analysis On: 5/3/2007				Second Triad Analysis On:			
Analyzer Information		Zero		Reference		Candidate		Result	
Manufacturer:	HP	Trial 1	Trial 2	Trial 3	Units	Trial 1	Trial 2	Trial 3	Units
Model Number:	6890	0.00	0.00	0.00					
Serial Number:	8295	47.99	47.96	47.93					
Analytical Principle:	GC-FID	29.79	29.77	29.84					
MPC Calibrated:	04/05/07	31.00	30.98	31.05	ppm				
		Mean Result:		31.01	ppm	Mean Result:			

Analyst Signature: \_\_\_\_\_

Warren Pereira


Calculated by: \_\_\_\_\_

Warren Pereira

F4



Air Liquide America  
Specialty Gases LLC



## COMPLIANCE CLASS

*Dual-Analyzed Calibration Standard*

11426 FAIRMONT PKWY, LA PORTE, TX 77571

Phone: 800-248-1427

Fax: 281-474-8419

### CERTIFICATE OF ACCURACY: EPA Protocol Gas

#### Assay Laboratory

AIR LIQUIDE AMERICA SPECIALTY GASES LLC  
11426 FAIRMONT PKWY  
LA PORTE, TX 77571

P.O. No.: 52579

Project No.: 04-78124-001

#### Customer

ENTECH ENGINEERING, INC  
TERRY HUDY  
PO# 52579  
408 E MAIN  
LEAGUE CITY TX 77573

#### ANALYTICAL INFORMATION

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards;  
Procedure G-1; September, 1997.

Cylinder Number: CC151954

Certification Date:

03Nov2009

Exp. Date: 02Nov2014

Cylinder Pressure\*\*\*: 1500 PSIG

Batch No: LAP0003953

#### COMPONENT

#### CERTIFIED CONCENTRATION (Moles)

#### ANALYTICAL

#### ACCURACY\*\*

#### TRACEABILITY

PROPANE

88.4

PPM

+/- 2%

NIST and VSL

NITROGEN

BALANCE

\*\*\* Do not use when cylinder pressure is below 150 psig.

\*\* Analytical accuracy is based on the requirements of EPA Protocol procedures, September 1997.

#### REFERENCE STANDARD

#### TYPE/SRM NO.

#### EXPIRATION DATE

#### CYLINDER NUMBER

#### CONCENTRATION

#### COMPONENT

NTRM 1668

02Oct2012

ALM020070

98.80 PPM

PROPANE

#### INSTRUMENTATION

#### INSTRUMENT/MODEL/SERIAL#

#### DATE LAST CALIBRATED

#### ANALYTICAL PRINCIPLE

HP-Y/HP 6890/US00000974

16Oct2009

GAS CHROMATOGRAPHY

#### Special Notes:

CERTS AND TAGS: ENTECH ENGINEERING

PO# 52579 ATTN: TERRY HUDY

APPROVED BY:

ROGER NGUYEN

**AIR LIQUIDE****RECERTIFICATION OF ANALYSIS****Interference Free Multi-Component EPA Protocol Gases**

Note: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121  
Section 2.2, Procedure G-1

Customer: ENTECH ENGINEERING, INC.  
Location: LA PORTE, TX

Cylinder S/N: EB0008543

Shipping Order Number: 30709722  
Transfer Number: 30709716  
Lot Number: SFS124927  
Valve: CGA 660  
Cylinder Pressure\*: 2000 PSIG  
\*Cylinder should not be used when  
gas pressure is below 150 psig

P.O. Number: 52251

Assay Date: 4-Sep-2008

Expiration Date: 4-Sep-2010

Components	Requested Concentration	Assay Concentration
Air	Balance	Balance
Nitrogen Dioxide	50 ppm	50.0 ± 1.2 ppm

**Reference Standard(s) Employed For Analysis**

Certified Concentration and Uncertainty	Component	Balance	Cyl. No.	SRM/PRM/Mix No.	Exp. Date	Sample No.	Type
51.4 ± 1.2 ppm	Nitrogen Dioxide	Nitrogen	EB0009973	SFS119543	12-Jul-2010	BI	GMIS

**Analytical Data**

Component: Nitrogen Dioxide		FIRST TRIAD ANALYSIS 26-Feb-2008				SECOND TRIAD ANALYSIS 4-Sep-2008						
Analyzer Information						Trial 1			Trial 2	Trial 3	Units	
Analyzer Type:	Fourier Transform IR					Zero	0.032	0.055	0.041		ppm	
Manufacturer:	MKS Instruments	INITIAL				Reference	51.323	51.299	51.384		ppm	
Model Number:	2031	ASSAY				Candidate	50.084	49.935	49.931		ppm	
Serial Number:	10387278					Result	50.14	50.03	49.95		ppm	
MPR Last Calibrated:	4-Sep-2008	Evaluation:	Valid	Valid	Valid	Evaluation	Valid	Valid	Valid			
Analytical Principle:	FTIR					Mean Analytical Result:				50.00	ppm	
						Mean Analytical Result:				50.04	ppm	

Analyst: 

Tan Ngo

Approved by: 

Tuan Tran

F6

AIR LIQUIDE AMERICA, L.P.

8832 Dice Road, Santa Fe Springs, CA 90670-2516  
Phone: (562) 945-1383 • Fax: (562) 693-1156



# CERTIFICATE of ANALYSIS

## Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: Entech  
P.O. Number: 52300  
Item Number:  
Notes:  
File Name: 313971401B  
Assay Date: 30-Nov-08

Cyl. Number: CC61285

Shipping Order #: 31397140  
Transfer #: 31397140  
LOT #: LPX245265  
Valve: CGA660  
Cyl. Pressure:\* 1900psig

\*Cylinder should not be used when gas pressure is below 150 psig

Expiration Date: 30-Nov-10

Component	Requested Concentration	Assay Concentration
Nitric Oxide	52.5 ppm	53.3 ±0.7 ppm
Total NOX		53.7 ppm
Nitrogen	Balance	Balance

### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS426	GMIS426	50.57	ppm	0.48	NO	N2	CC217252	6/26/2010	N.A.

### Analysis Information:

Component 1: Nitric Oxide									
Analyzer Information		First Triad Analysis On: 11/20/2008				Second Triad Analysis On: 11/30/2008			
Manufacturer:	KVB/Analect	Trial 1	Trial 2	Trial 3	Units	Trial 1	Trial 2	Trial 3	Units
Model Number:	EN3024	Zero	-0.50	0.41	-0.08	Zero	0.32	0.15	-0.12
Serial Number:	3024	Reference	51.66	50.61	50.89	Reference	50.59	51.30	50.71
Analytical Principle:	FTIR	Candidate	53.62	52.83	53.81	Candidate	53.93	54.02	54.20
MPC Calibrated:	11/13/08	Result	53.11	52.33	53.30	Result	53.63	53.71	53.89
		Mean Result:				Mean Result:			
		52.91 ppm				53.74 ppm			

Analyst Signature:  Ron Stitt

Calculated by:  Ron Stitt

F7



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: Entech  
P.O. Number: 52300  
Item Number:  
Notes:  
File Name: 313974211A  
Assay Date: 30-Nov-08

Cyl. Number: CC241846

Shipping Order #: 31397421  
Transfer #: 31397421  
LOT #: LPX245227  
Valve: CGA660  
Cyl. Pressure\*: 1900psig

Expiration Date: 30-Nov-10

\*Cylinder should not be used when gas pressure is below 150 psig

Component	Requested Concentration	Assay Concentration
Nitric Oxide	22 ppm	24.3 ±0.4 ppm
Total NOX		24.6 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS426	GMIS426	50.57	ppm	0.48	NO	N2	CC217252	6/26/2010	N.A.

#### Analysis Information:

Component 1: Nitric Oxide		First Triad Analysis On: 11/20/2008				Second Triad Analysis On: 11/30/2008			
Analyzer Information		Trial 1	Trial 2	Trial 3	Units	Trial 1	Trial 2	Trial 3	Units
Manufacturer:	KVB/Analect	Zero	-0.50	0.41	-0.08	Zero	0.32	0.15	-0.12
Model Number:	EN3024	Reference	51.66	50.61	50.89	Reference	50.59	51.30	50.71
Serial Number:	3024	Candidate	24.07	23.78	23.93	Candidate	25.72	24.59	24.92
Analytical Principle:	FTIR	Result	23.87	23.59	23.73	Result	25.52	24.39	24.71
MPC Calibrated:	11/13/08	Mean Result: 23.73				Mean Result: 24.87			
		ppm				ppm			

Analyst Signature:  Ron Stitt

Calculated by:  Ron Stitt

F8



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

**Customer:** ENTECH ENGINEERING  
**P.O. Number:** 51721  
**Item Number:**  
**Notes:**

**Cyl. Number:** CC111686

**Shipping Order #:** 23921644  
**Transfer #:** 23921644  
**LOT #:** LPX215846  
**Valve:** CGA350  
**Cyl. Pressure:\*** 1900psig

**Assay Date:** 1-Feb-07  
**Expiration Date:** 31-Jan-10

\*Cylinder should not be used when gas pressure is below 150 psig

Component	Requested Concentration	Assay Concentration
Carbon Monoxide	300 ppm	297 ±2 ppm
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS339	GMIS339	99.44	ppm	0.48	CO	N2	CC82421	11/14/2008	N.A.

#### Analysis Information:

Component 1: Carbon Monoxide		First Triad Analysis On: 1/24/2007					Second Triad Analysis On: 2/1/2007				
Analyzer Information		Trial 1	Trial 2	Trial 3	Units		Trial 1	Trial 2	Trial 3	Units	
Manufacturer:	KVB/Analect	Zero	-0.39	0.10	-0.14		Zero	-0.28	0.08	0.04	
Model Number:	EN3024	Reference	92.40	92.32	92.55		Reference	93.91	94.49	93.92	
Serial Number:	3024	Candidate	276.06	275.47	275.87		Candidate	282.27	283.37	282.12	
Analytical Principle:	FTIR	Result	296.72	296.08	296.52	ppm	Result	298.16	299.32	298.00	ppm
MPC Calibrated:	01/18/07	Mean Result:				296.44 ppm	Mean Result:				298.49 ppm

Analyst Signature:  Bryan Leger

Calculated by:  M. Adnane

F9



227



Air Liquide America  
Specialty Gases LLC



**RATA CLASS**

*Dual-Analyzed Calibration Standard*

9810 BAY AREA BLVD, PASADENA, TX 77507

Phone: 281-474-5800

Fax: 281-474-5857

**CERTIFICATE OF ACCURACY: EPA Protocol Gas**

Assay Laboratory

AIR LIQUIDE AMERICA SPECIALTY GASES LLC  
9810 BAY AREA BLVD  
PASADENA, TX 77507

P.O. No.: ALAS-44414/33518189  
Project No.: 04-74021-001

Customer

ENTECH ENGINEERING  
PO#52438  
ATTN: TERRY HUDY  
408 E MAIN  
LEAGUE CITY TX

**ANALYTICAL INFORMATION**

This certification was performed according to EPA Traceability Protocol For Assay & Certification of Gaseous Calibration Standards; Procedure G-1; September, 1997.

Cylinder Number: ALM014290      Certification Date: 04Jun2009      Exp. Date: 04Jun2012  
Cylinder Pressure\*\*\*: 1917 PSIG

**COMPONENT**

CARBON MONOXIDE  
NITROGEN

**CERTIFIED CONCENTRATION (Moles)**

132      PPM  
BALANCE

**ANALYTICAL**

**ACCURACY\*\***  
+/- 1%

**TRACEABILITY**

Direct NIST and NMI

\*\*\* Do not use when cylinder pressure is below 150 psig.

\*\* Analytical accuracy is based on the requirements of EPA Protocol Procedure G1, September 1997.

**REFERENCE STANDARD**

TYPE/SRM NO.	EXPIRATION DATE	CYLINDER NUMBER	CONCENTRATION	COMPONENT
NTRM 2636	02Oct2011	KAL003907	240.8 PPM	CARBON MONOXIDE

**INSTRUMENTATION**

INSTRUMENT/MODEL/SERIAL#

FTIR//000929060

DATE LAST CALIBRATED

30May2009

ANALYTICAL PRINCIPLE

FTIR

**ANALYZER READINGS**

(Z = Zero Gas    R = Reference Gas    T = Test Gas    r = Correlation Coefficient)

First Triad Analysis      Second Triad Analysis      Calibration Curve

**CARBON MONOXIDE**

Date: 28May2009      Response Unit: PPM

Z1 = -0.09635	R1 = 240.6386	T1 = 131.4899
R2 = 240.7436	Z2 = 0.14992	T2 = 131.5533
Z3 = 0.20132	T3 = 131.6579	R3 = 240.8057
Avg. Concentration: 131.6 PPM		

Date: 04Jun2009      Response Unit: PPM

Z1 = -0.05429	R1 = 240.4354	T1 = 131.5436
R2 = 240.7585	Z2 = 0.17570	T2 = 131.5554
Z3 = 0.25626	T3 = 131.5985	R3 = 240.8407
Avg. Concentration: 131.6 PPM		

Concentration = A + Bx + Cx<sup>2</sup> + Dx<sup>3</sup> + Ex<sup>4</sup>  
r = 9.99999E-1

Constants:      A = 0.00000E+0  
B = 5.08875E-1      C = 1.87000E-4  
D = 0.00000E+0      E = 0.00000E+0

**Special Notes:**

CERTS AND TAGS:      LOT# PAS00481    ENTECH ENGINEERING  
PO#52438    ATTN: TERRY HUDY

**APPROVED BY:**

Ramien JR

FID



## CERTIFICATE of ANALYSIS

### Interference-Free Multi-Component EPA Protocol Gases

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: ENTECH  
P.O. Number: 52300  
Item Number:  
Notes:

Cyl. Number: CC55212

Shipping Order #: 31397843  
Transfer #: 31397843  
LOT #: LPX245230  
Valve: CGA590  
Cyl. Pressure\*: 1900psig

Assay Date: 24-Nov-08

Expiration Date: 24-Nov-11

\*Cylinder should not be used when gas pressure is below 150 psig

Component	Requested Concentration	Assay Concentration
Oxygen	9.5 %	9.49 $\pm$ 0.10 %
Nitrogen	Balance	Balance

#### Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS413	GMIS413	10.0	%	0.1	O2	N2	CC204761	2/21/2010	N.A.

#### Analysis Information:

Component 1: Oxygen		First Triad Analysis On: 11/24/2008					Second Triad Analysis On:			
Analyzer Information		Trial 1		Trial 2		Trial 3	Units	Trial 1		Trial 2
Manufacturer:	Servomex	Zero		0.00		0.01	0.00	Zero		Trial 3
Model Number:	4605C	Reference		9.87		9.89	9.89	Reference		Units
Serial Number:	1101	Candidate		9.38		9.39	9.38	Candidate		
Analytical Principle:	Paramag	Result		9.49		9.50	9.49	Result		
MPC Calibrated:	11/17/08									
				Mean Result:		9.49				
						%		Mean Result:		

Analyst Signature: 

Debra Jackson

Calculated by: 

Debra Jackson

F11

Mix Assayed At: Air Liquide America

11426 Fairmont Pkwy, La Porte, TX, 77571

Phone: (281) 474-8400 Fax: (281) 474-8419

**AIR LIQUIDE****CERTIFICATE of ANALYSIS****Interference-Free Multi-Component EPA Protocol Gases**

NOTE: Analytical uncertainty and NIST traceability are in compliance with EPA-600/R-97/121

Section 2.2

Procedure: G-1

Customer: ENTECH  
 P.O. Number: 52300  
 Item Number:  
 Notes:

Cyl. Number: CC9824

Shipping Order #: 31397611  
 Transfer #: 31397611  
 LOT #: LPX245229  
 Valve CGA580  
 Cyl. Pressure:\* 1900psig

Assay Date: 24-Nov-08

Expiration Date: 24-Nov-11

\*Cylinder should not be used when gas pressure is below 150 psig

Component	Requested Concentration	Assay Concentration
Oxygen	5 %	4.94 ±0.05 %
Nitrogen	Balance	Balance

## Reference Standard(s) Employed For Analysis:

Std name	Std #	Conc.	Units	Std. Error	Comp.	Balance	Cyl. No.	Exp. Date	Sample No.
GMIS413	GMIS413	10.0	%	0.1	O2	N2	CC204761	2/21/2010	N.A.

## Analysis Information:

Component 1: Oxygen		First Triad Analysis On: 11/24/2008				Second Triad Analysis On: -			
Analyzer Information		Trial 1	Trial 2	Trial 3	Units	Trial 1	Trial 2	Trial 3	Units
Manufacturer:	Servomex	0.00	0.01	0.00					
Model Number:	4605C	9.87	9.89	9.89					
Serial Number:	1101	4.88	4.88	4.88					
Analytical Principle:	Paramag.	4.94	4.94	4.94	%				
MPC Calibrated:	11/17/08	Mean Result: 4.94			%	Mean Result:			

Analyst Signature: \_\_\_\_\_

Warren Pereira

Calculated by: \_\_\_\_\_

Warren Pereira

*F12*

***ENTECH ENGINEERING INC.***

***P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118***

---

**APPENDIX G.**

**PROCESS DATA**

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**PROCESS DATA**

Process data concerning the Unit 39.1 Tail Gas Incinerator initial demonstration of compliance test will be provided by ConocoPhillips Company.

***ENTECH ENGINEERING INC.***

*P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

**APPENDIX H.**

**RESUMES**

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

Steven L. Yin

2002-present	<i>Laboratory Manager</i> Entech Engineering Inc. <ul style="list-style-type: none"><li>- Laboratory Analysis</li><li>- Instrument maintenance</li><li>- Method development</li><li>- Analytical QA/QC assurance</li></ul>
1990-2002	<i>Supervisor/Senior Environmental Chemist</i> Pace Analytical Services, Inc.
1998-1990	<i>Technical Support Chemist</i> Analytical Sensors, Inc.
1985-1998	<i>M.S. Chemist</i> Texas Southern University Houston, Texas
1978-1982	<i>B.S. Material Science</i> Shanghai University of Science and Technology Shanghai, China

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

Michael Brocato

2008-present	Entech Engineering Inc. <i>Environmental Scientist II</i>
2006-2007	Dexter Field Services, Inc. <i>Project Manager</i> -Region 1 & 2 Cooling Towers -BWON Compliance Sampling - Assist with LDAR stream speciation
2001-2006	Lamar University Beaumont, TX B.S. Biology -minor in Chemistry



Gustavo Vargas

2009- Present	Senior Environmental Specialist Entech Engineering, Inc. <ul style="list-style-type: none"><li>- Source Compliance Testing</li><li>- Emission Abatement</li><li>- CEMS Certification</li><li>- Laboratory Analysis</li></ul>
2004-2009	Environmental Field Supervisor Golden Specialty Consulting, Ltd. <ul style="list-style-type: none"><li>- Source Compliance Testing</li><li>- Emission Abatement</li><li>- CEMS Certification</li><li>- Laboratory Analysis</li></ul>
2001-2004	Environmental Field Technician Airtech Environmental Services <ul style="list-style-type: none"><li>-Source Compliance Testing</li><li>- Emission Abatement</li><li>- CEMS Certification</li><li>- Laboratory Analysis</li></ul>
1994-1999	Law Degree U.A.N.L. Monterrey, Mexico

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

Mark A Lester

- |              |   |
|--------------|---|
| 2002-present | <i>Principal Environmental Technician</i><br>Entech Engineering Inc. <ul style="list-style-type: none"><li>- Source Compliance Testing</li><li>- Equipment calibration and maintenance</li><li>- Source Compliance Testing</li><li>- Emissions Abatement</li><li>- CEMS Certification</li><li>- Laboratory Analysis</li></ul>   |
| 1983-2001    | <i>Senior Environmental Tech.- Field Supervisor</i><br>Tetra tech NUS formally Haliburton NUS <ul style="list-style-type: none"><li>- Source Compliance Testing</li><li>- Emissions Abatement</li><li>- CEMS Certification</li><li>- Laboratory Analysis</li><li>- Fugitive emission sampling</li><li>- Ambient air sampling</li><li>- Under ground storage tank removal</li><li>- Soil Analysis</li><li>- Soil remediation</li><li>- Heavy equipment operator</li><li>- Ground water sampling</li><li>- Ground water remediation</li><li>- Ground water analysis</li><li>- Hazardous waste clean up (Super Fund Sites)</li></ul> |
| 1978-1983    | <i>Environmental Technician</i><br>Energy Impact Associates <ul style="list-style-type: none"><li>- Ambient air sampling</li><li>- Meteorological equipment set-up and calibration</li><li>- Ambient air station maintenance and calibration</li></ul>  |

**RESUMES**

Joseph Muir

- |              |  |
|--------------|--|
| 2007-Present | <i>Senior Environmental Scientist</i><br>Entech Engineering Inc.<br>-Source Compliance Testing<br>-Equipment Calibration and Maintenance                         |
| 2006-2007    | <i>Order Filler</i><br>UTI<br>-Distribute items to various regions<br>-Process trailers for loading and logistical transactions<br>-Heavy equipment cerification |
| 2002-2006    | <i>Contract worker and entertainment specialist</i><br>Staff Mart/Target<br>-Loading and processing cargo<br>-Supervised and monitored subordinate employees     |
| 2000-2005    | <i>B.S. Wildlife and Fisheries Science</i><br>Texas A&M University<br>College Station, Texas   |

***ENTECH ENGINEERING INC.***

*P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

**APPENDIX I.**

**CHAIN OF CUSTODY**

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**CHAIN OF CUSTODY**

Mr. Joseph Muir, Senior Environmental Scientist of Entech Engineering Inc., was in charge of obtaining the emission samples and/or data and transporting them from ConocoPhillips Company's Sweeny Refinery in Old Ocean, Texas to Entech Engineering Inc. in League City, Texas. The emission samples and/or data were delivered to and analyzed at the Entech Engineering office in League City, Texas.

P. O. Box 890748, Houston, Texas 77289-0748, (281)332-3178

## Chain of Custody

[illegible]

Original

10

I 2

**ENTECH ENGINEERING INC.**  
P. O. Box 890746, Houston, Texas 77289-0746, (281) 332-3718

P. O. Box 890746 • Houston, Texas 77289-0746 • (281)332-3118

## Chain of Custody

[illegible]

original

ORIGINAL

P. O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3778

## Chain of Custody

Original

I 4



***ENTECH ENGINEERING INC.***

*P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118*

---

**APPENDIX J.**

**PERSONNEL INFORMATION**

**ENTECH ENGINEERING INC.**

P.O. Box 890746 • Houston, Texas 77289-0746 • (281) 332-3118

---

**PERSONNEL INFORMATION**

The ConocoPhillips Company contact for additional information is Mr. Sean O'Brien. The testing organization is Entech Engineering Inc. and the contacts at Entech Engineering are Mr. Billy T.Y. Lee and Mr. Edward J. Pasternak. The names and addresses of these contacts are as follows:

Mr. Sean O'Brien  
Environmental Department  
ConocoPhillips Company  
Sweeny Refinery  
Old Ocean, Texas 77463  
P.O. Box 866  
Sweeny, Texas 77480  
(979) 491-2705

Mr. Billy T.Y. Lee, P.E.  
President  
Entech Engineering Inc.  
408 E. Main  
League City, Texas 77573  
(281) 332-3118

Mr. Edward J. Pasternak  
Technical Manager  
Entech Engineering Inc.  
408 E. Main  
League City, Texas 77573  
(281) 332-3118